

Article

Spectral Signatures of X-ray Scatter Using Energy-Resolving Photon-Counting Detectors

Cale E. Lewis and Mini Das *

Department of Physics, University of Houston, Houston, TX 77204, USA; celewis2@central.uh.edu

* Correspondence: mdas@central.uh.edu

Received: 9 October 2019; Accepted: 1 November 2019; Published: 18 November 2019



Abstract: Energy-resolving photon-counting detectors (PCDs) separate photons from a polychromatic X-ray source into a number of separate energy bins. This spectral information from PCDs would allow advancements in X-ray imaging, such as improving image contrast, quantitative imaging, and material identification and characterization. However, aspects like detector spectral distortions and scattered photons from the object can impede these advantages if left unaccounted for. Scattered X-ray photons act as noise in an image and reduce image contrast, thereby significantly hindering PCD utility. In this paper, we explore and outline several important characteristics of spectral X-ray scatter with examples of soft-material imaging (such as cancer imaging in mammography or explosives detection in airport security). Our results showed critical spectral signatures of scattered photons that depend on a few adjustable experimental factors. Additionally, energy bins over a large portion of the spectrum exhibit lower scatter-to-primary ratio in comparison to what would be expected when using a conventional energy-integrating detector. These important findings allow flexible choice of scatter-correction methods and energy-bin utilization when using PCDs. Our findings also propel the development of efficient spectral X-ray scatter correction methods for a wide range of PCD-based applications.

Keywords: X-ray scatter; photon-counting detectors; imaging; Compton scatter

1. Introduction

Image contrast in transmission X-ray imaging results from attenuation variations of material types within the object. While these attenuation variations are spectrally dependent, conventional energy-integrating detectors (EIDs), widely used in clinical and industrial X-ray imaging, do not capture this important information. Significant developments in the area of energy-resolving photon-counting detectors (PCDs) are happening globally. PCDs developed by CERN Medipix collaboration [1,2] are some of the most advanced in this category. Exploiting spectral information allows quantitative material identification and characterization. Examples of these techniques include K-edge imaging [3–5], material decomposition [6–8] and phase-contrast imaging [9–11]. Their other advantages include zero dark noise, and the ability for flexible energy weighting [12–14].

Energy-resolving PCDs are direct conversion detectors. Impinging X-ray photons are converted into charge packets in the semiconducting wafer bump-bonded to a virtually pixellated electronic readout. The application-specific integrated circuit (ASIC) consists of an individual readout for each pixel. This includes an electronic discriminator to exclude photons above a given threshold energy value selected by the experimenter. Using efficient energy-calibration methods, the experimenter can accurately link these electronic thresholds to the corresponding photon energy [15].

Detector performance is highly influenced by the semiconductor sensor material. Charge-sharing limitations, such as charge diffusion and X-ray fluorescence, reduce the spectral resolution by falsely attributing a single-incident photon as one or more lower-energy photons. These distortions restrict

detector design since influence increases for thicker sensors and smaller pixel sizes. High-Z materials (such as CdTe and CdZnTe [16]) have desirable X-ray absorption efficiency, but are often prone to spectral distortions resulting from K-fluorescence and polarization. These issues are diminished for low-Z material (such as silicon) at the cost of reduced absorption efficiency. In addition, charge-sharing correction techniques are being developed to improve the spectral response of PCDs [17,18]. Advanced PCDs, such as Medipix3RX (a recent version developed by the Medipix collaboration), benefit from interpixel communication known as charge-summing mode [1] while maintaining one of the highest spatial resolutions (55 μm) and multiple energy-threshold capabilities.

Inaccurate spectral-intensity measurements also occur due to photons scattered from the object. The spectral characteristics of scattered X-rays differ from the primary due to the energy transfer and angular deflection of the collision. Maintaining low scatter-to-primary ratios (SPRs) is crucial for spectral imaging where scatter degrades spectral fidelity and hence material identification [19,20], including material decomposition [21] and iodine K-edge imaging [22]. Thus, these photons generally increase the total apparent counts and decrease the contrast-to-noise ratio [23].

Here, we report on the spectral characteristics of incoherent X-ray scatter in transmission X-ray imaging using an energy-resolving PCD. We used both Monte Carlo simulations and bench-top experiments to identify key features. As a specific application example, we considered spectral and object parameters relevant to applications such as mammography. Identifying spectral and detector features that vary the SPR can help develop better PCD imaging design and scatter-correction techniques. We use both simulations and experiments to explore these questions.

2. Materials and Methods

2.1. Monte Carlo Simulations

The BEAMnrc software package [24] was used for Monte Carlo simulations to emulate our bench-top experiment (Figure 1). A tungsten anode X-ray tube was used to generate a polychromatic beam from a point source with 5×10^7 X-rays per exposure. The spectral distribution of the cone-beam source was modeled by simulating the bremsstrahlung radiation of an electron-beam incidence on a tungsten target. Biological samples (and other low-Z materials like explosives) are considerably susceptible to object scatter, as demonstrated by linear-attenuation coefficients shown in Figure 2a; certain plastics like polymethyl methacrylate (PMMA) mimic their X-ray absorption and scattering properties (see Figure 2a).

In our study, X-ray photons were incident on a homogeneous PMMA slab with lateral dimensions of $12.4 \times 12.4 \text{ cm}^2$, and thickness range was 2.0–8.0 cm to approximate the typical breast area and thickness observed in mammography [25]. Photoelectric absorption, Rayleigh and Compton scattering (with cross-sections for bound electrons obtained from the NIST database [26,27]) were considered as predominant interactions. Photons that interacted with the PMMA were ‘tagged’ such that scatter and primary intensities could be differentiated at the detector. Figure 2b shows the simulated spectral distributions of total, primary, and scattered photons through a 6.0 cm thick PMMA slab at a 0.5 cm air gap. Under these conditions, primary and scattered photons contribute in nearly equal amounts to total measured intensity. The scattered radiation also shifted the total distribution to slightly lower energies, reflecting a spectrally dependent SPR.

Detection was performed by recording the energy of each X-ray photon that crossed the scoring region. This region was centered on the transmission axis 70 cm from the source with an area of $12.4 \times 12.4 \text{ cm}^2$. The proportion of scattered photons reaching the detector was altered by changing the object-to-detector distance from 0.5 to 30 cm. Spectral distributions of the detected photons were obtained with histograms ranging from 10 keV to the source peak voltage (kVp) divided into 200 bins. The resultant spectra were then modified by the absorption function of a 300 μm thick silicon detector.




Figure 1. Schematic of experiment configuration.




Figure 2. (a) Linear attenuation coefficients of breast tissue (blue) and polymethyl methacrylate (PMMA; red). Total linear-attenuation coefficient shown (solid curves) along with photoelectric absorption (dotted curves) and incoherent scatter components (dashed curves). (b) Simulated spectral distribution of total, primary, and scattered intensities for a 6.0 cm thick PMMA slab at a 0.5 cm air gap with 60 kVp incidence. Each distribution is normalized by the unattenuated source intensity.

Quantitative accuracy due to changing scatter counts was evaluated by estimating the linear-attenuation coefficient of the PMMA slabs under varying conditions:

$$\mu(E) = -\ln \left(\frac{I(E)}{I_0(E)} \right) / t, \quad (1)$$

where $I(E)$ and $I_0(E)$ are measured intensities with and without the object, respectively, and t is slab thickness. This was compared against the ideal (obtained from the NIST XCOM database [28]) or known attenuation values to obtain the percent error:

$$\%Error = \frac{\mu(E) - \mu(E)^{ideal}}{\mu(E)^{ideal}}. \quad (2)$$

Finally, an EID detection scheme was approximated by summing photon counts across the spectral distributions with weights corresponding to their respective energies,

$$I_{EID} = \int_{TH}^{kVp} E \cdot I(E) D(E) dE, \quad (3)$$

where $I(E)$ is photon-intensity distribution, and summation ranged from the lowest photon energy at TH up to the peak photon energy (kVp). Factor $D(E) = 1 - e^{-\mu_d(E)t_d}$ represents detector absorption function that depends on detector-material attenuation (μ_d) and thickness (t_d). Lower energy limit for the threshold in Equation (3) was set to 10 keV, which includes the full width of the simulated spectral distribution of the source.

2.2. Bench-Top Experiments

Experiments were conducted on a bench-top X-ray imaging system to validate the simulation results and to gain additional insights. X-rays were generated using a Hamamatsu microfocus X-ray tube unit (L8122-01) consisting of a tungsten anode target and a 200 μm thick beryllium output window. The tube was operated at 60 kVp with a 450 μA current and 50 μm focal spot to produce the source X-ray distribution. X-rays were propagated through rectangular PMMA slabs and recorded with a Medipix3RX PCD operated in charge-summing mode at a fixed distance (70 cm) from the source, as in the case of simulations described above.

3. Results

Contrast loss due to increased scatter is demonstrated with experimentally obtained projections of a breast phantom in Figure 3. Both lesion sharpness and microcalcification contrast improved by reducing the scatter through increasing the air gap between object and detector from 1.2 to 10.0 cm.




Figure 3. Projections of a CIRS breast phantom with embedded microcalcifications obtained at air gaps of (top) 1.2 cm and (bottom) 10.0 cm. Data for 20–30 keV energy bin.

3.1. Energy-Integrated SPR

As shown in Equation (3), photons were weighed by their energy in an EID, and all spectral information was lost in the detection process. Scattered and primary photons could easily be separated in a Monte Carlo simulation. Figure 4 shows the simulated SPR values obtained with energy-integrating detection for different kVps and object thicknesses. SPR values are provided for both a 300 μm thick silicon detector and a fully absorbing ideal detector to illustrate dependence on the detection-absorption efficiency. In both cases, the EID SPR increased linearly with object thickness and showed only slight dependence on kVp, which is consistent with previous investigations [25,29–31].

In comparison to an ideal detector with 100% absorption across the entire spectrum, detectors with Si sensors have poor and significantly decreasing absorption efficiency with increasing photon energies. This results in SPR deviation when using a detector with a thin Si sensor (here, 300 μm thickness). This can be described when considering the detector absorption function $D(E)$ in SPR estimation as well as the energy-integration process,

$$SPR = \frac{\int_{TH}^{kVp} E \cdot I_s(E) D(E) dE}{\int_{TH}^{kVp} E \cdot I_p(E) D(E) dE}, \quad (4)$$

where $I_s(E)$ and $I_p(E)$ represent scatter and primary intensities, respectively. Deviation between ideal- and finite-thickness Si detectors increases with increasing spectral kVp and object thickness (see

Figure 4). This is the result of a higher proportion of lower-energy photons in the absorbed spectrum when using an Si detector that has very poor high-energy-absorption efficiency.




Figure 4. Simulated scatter-to-primary ratio (SPR) values using (solid) a 300 μm thick silicon energy-integrating detector (EID) and (dashed) an ideal EID as a function of kVp for different object thicknesses at a fixed 0.5 cm air gap.

Distinction from the above can be made when considering spectrally resolved imaging with sufficiently narrow energy windows. Referring to Equation (4), in a spectroscopic measurement (as in a PCD), the integration limit shrinks to a very narrow energy bin or even simply single energy, resulting in the SPR to be merely a ratio of $I_S(E)$ and $I_P(E)$, thereby removing the $D(E)$ dependence in the SPR estimates. Therefore, the spectral SPR due to object scatter is expected to be identical for all detector sensor materials and thicknesses, making our following findings on spectral SPR more general.

3.2. Energy-Resolved SPR

Figure 5 shows energy-resolved SPR values for different kVps and object thicknesses. The SPR remains low near the kVp and increases approximately linearly as the energy is reduced. The elevated SPR at low energies is due to a combination of residual scatter and preferential absorption of the low-energy primary photons. This is highlighted by a rise in SPR with increasing object thickness with decreasing photon energies. Additionally, this causes a low-energy limit where the SPR becomes very large and estimation becomes unreliable due to the eventual high absorption rate and the resulting low photon counts at low energies.

Additionally, Figure 5 includes the spectral SPR for varying kVps. For a given object thickness, increasing the kVp has little effect on the maximum SPR value. This reflects minor kVp dependence on the EID SPRs shown previously in Figure 4. Rather, increasing the kVp reduces the rate at which spectral SPR decreases with energy. For a given energy, the number of incident photons that can contribute to scatter intensity increases with the kVp, which in turn increases the SPR. Convergence of the spectral SPR at low energies may suggest an ‘upper limit’ for the given sample thickness and air gap. However, this assumption may not be suitable for conditions where increased multiple scatter is expected, such as imaging thicker samples. This is partially illustrated by the higher degree of convergence of the 2.0 cm thick sample relative to the 8.0 cm thick sample.

Figure 6 shows the EID and spectral SPR for varying object thickness for a small air gap. These SPR values have been normalized by the object thickness to emphasize the relative variations of these two quantities. In each case, the EID SPR lies within the range of spectral SPR values. A small range of lower energies have greater SPR than the EID SPR, while a larger range of energies have lower SPR in comparison to the EID SPR. This implies that quantitative inaccuracies and image quality can be improved over a conventional EID by carefully selecting energy bins. A larger range of energies with SPR lower than the EID SPR can be obtained by increasing the kVp. This is also considering the previous observation that the EID SPR varied only slightly, and that the spectral SPR values were extended across a wider energy range.




Figure 5. Spectrally-resolved SPRs from simulation using a 0.5 cm air gap for different kVps and object thicknesses of (a–d) 2, 4, 6, and 8 cm, respectively.




Figure 6. Simulated spectral SPR (solid) and EID SPR (dashed) for varying object thicknesses at a 0.5 cm air gap for a 60 kVp X-ray spectrum. SPR values were normalized by the object thickness to emphasize the linear dependence. EID SPR values lack energy resolution and are instead shown by horizontal lines across the full energy range for comparison with spectral SPR.

3.3. SPR Influence on Estimated Linear Attenuation

X-ray scatter reduces apparent sample attenuation by increasing photon counts across the detector. The percent error of the estimated attenuation coefficient can be expressed as

$$\%Error(E) = \frac{1}{\mu(E)^{ideal}t} \ln \left(\frac{1}{1 + \frac{I_S(E)}{I_P(E)}} \right). \quad (5)$$

(see Appendix A for further detail). This expression of the percent error highlights the influence of ideal object attenuation and thickness. The percent error is reduced when μ^{ideal} is large, which occurs at lower energies. Similarly, the percent error reduces as the object thickness increases. This inverse dependence of these terms reflects the difficulty of identifying objects that have smaller μt values.

Figure 7 shows percent errors of estimated attenuation coefficients for different PMMA thicknesses and air gaps. The error was quite large ($>20\%$) for most energy bins at the small (0.5 cm) air gap. Similar to the spectral SPR (see Figure 6), the error diminished for energies near the kVp. However, the error crested at an intermediate energy bin and began to reduce until a rapid increase at a low energy limit. The error improved at the intermediate energies considering that object attenuation increased as energy was reduced. The rapid increase of errors at low energies was due to a combination of increased scatter and the starvation of primary photons due to higher absorption of the object (PMMA) at these energies.




Figure 7. Percent error of estimated attenuation coefficient via simulations (left column) and experiments (right column) for varying energies. Shown for different object thicknesses and air gaps of (a,b) 0.5 and (c,d) 30 cm, respectively.

It was also observed that attenuation errors due to the presence of scatter (see Figure 7a,b) were higher for thinner objects. This can be due to the percent error for estimated attenuation being inversely proportional to the product of object thickness and ideal attenuation. This reflects the concept of a thinner, less attenuating object being less likely to be identified than thicker objects as scatter increases. Examples of these are detecting microcalcifications or small low-contrast cancerous masses in mammograms as well as the detection of explosives among plastics.

Increasing the air gap between object and detector is an effective technique to reduce the number of scattered photons from reaching the detector, and to improve the estimated attenuation accuracy (compare top and bottom Figure 7 rows). However, even in this case, if the detector's spectral response has some fluence dependence, one can observe a residual error in the attenuation estimation (Figure 7d). This is due to the increasing difference between object intensity and flat-field intensity (measured without the object) in attenuation estimation (Equation (1)), when object thickness increases. These fluence-dependent detector distortions were not accounted for in simulations where these thickness-dependent errors were not observed (Figure 7c). Increasing the air gap from 0.5 to 30 cm reduced the maximum error in attenuation estimation of the 2 cm thick PMMA slab from $\sim 60\%$ to $\sim 10\%$. However, increasing the air gap did not improve the errors at energies <20 keV where the primary was severely reduced by increased object absorption. Additionally, increasing the air gap is also not a practical possibility in many applications, such as in mammography. In these cases, efficient scatter correction is the only solution.

4. Discussion and Conclusions

In conclusion, we illustrated several key spectral characteristics of scattered X-ray photons when using an ideal detector and a silicon detector. These general characteristics can be applied to any PCD. For high-Z sensors like CdTe and CdZnTe, there are additional detector distortions to consider and for which to correct. Our finding that a large number of high-energy bins in PCD measurement exhibit lower SPR than a conventional EID X-ray detector, used clinically and industrially, can offer the means to effectively use these detectors without physical antiscatter grids. The spectral characteristics we reported could also be used for developing design and algorithmic strategies for energy-weighting, and material-decomposition and -characterization methods for which these detectors are being developed.

Author Contributions: The authors contributed equally to this work.

Funding: This work was partially supported by funding from the US Department of Defence (DOD) Congressionally Directed Medical Research Program (CDMRP) Breakthrough Award BC151607 and the National Science Foundation CAREER Award 1652892.

Acknowledgments: We are thankful for the discussions with several Medipix3 collaboration members at CERN.

Conflicts of Interest: The authors declare no conflict of interest.

Appendix A

The following information has been included to provide additional detail to the percent errors described in Section 3.3. Although the SPR continuously increased as energy was reduced (see Figure 6), the percent error of the attenuation may not follow a similar trend due to dependence on object attenuation and thickness. The measured attenuation coefficient could be described by [23]:

$$\begin{aligned}\mu(E)t &= \ln \left(\frac{I_0(E)}{I_P(E) + I_S(E)} \right) \\ &= \ln \left(\frac{I_0(E)}{I_P(E)} \cdot \frac{1}{1 + \frac{I_S(E)}{I_P(E)}} \right) \\ &= \mu(E)^{ideal}t + \ln \left(\frac{1}{1 + \frac{I_S(E)}{I_P(E)}} \right).\end{aligned}\tag{A1}$$

By combining Equations (2) and (A1), the percent error of the attenuation is described by

$$\%Error(E) = \frac{1}{\mu(E)^{ideal}t} \ln \left(\frac{1}{1 + \frac{I_S(E)}{I_P(E)}} \right).$$

The inverse relationship on object properties suggests that accuracy improves at lower energies (where $\mu(E)$ is elevated) and as thickness increases. Figure 7a,b reflects these considerations.

References

1. Ballabriga, R.; Alozy, J.; Blaj, G.; Campbell, M.; Fiederle, M.; Frojdh, E.; Heijne, E.; Llopart, X.; Pichotka, M.; Procz, S.; et al. The Medipix3RX: A high resolution, zero dead-time pixel detector readout chip allowing spectroscopic imaging. *J. Instrum.* **2013**, *8*, C02016. [\[CrossRef\]](#)
2. Taguchi, K.; Iwanczyk, J.S. Vision 20/20: Single photon counting X-ray detectors in medical imaging. *Med. Phys.* **2013**, *40*. [\[CrossRef\]](#)
3. Roessl, E.; Proksa, R. K-edge imaging in X-ray computed tomography using multi-bin photon counting detectors. *Phys. Med. Biol.* **2007**, *52*, 4679. [\[CrossRef\]](#)
4. Schlomka, J.; Roessl, E.; Dorscheid, R.; Dill, S.; Martens, G.; Istel, T.; Bäumer, C.; Herrmann, C.; Steadman, R.; Zeitler, G.; et al. Experimental feasibility of multi-energy photon-counting K-edge imaging in pre-clinical computed tomography. *Phys. Med. Biol.* **2008**, *53*, 4031. [\[CrossRef\]](#)

5. Roessl, E.; Brendel, B.; Engel, K.J.; Schlomka, J.P.; Thran, A.; Proksa, R. Sensitivity of Photon-Counting Based K-Edge Imaging in X-ray Computed Tomography. *IEEE Trans. Med. Imaging* **2011**, *30*, 1678–1690. [\[CrossRef\]](#)
6. Shikhaliev, P.M.; Fritz, S.G. Photon counting spectral CT versus conventional CT: Comparative evaluation for breast imaging application. *Phys. Med. Biol.* **2011**, *56*, 1905. [\[CrossRef\]](#)
7. Fredette, N.R.; Vespucchi, S.; Das, M. Experimental validation of a multi-step material decomposition method for spectral CT. In Proceedings of the SPIE Medical Imaging 2019: Physics of Medical Imaging, San Diego, CA, USA, 16–21 February 2019; p. 109482T.
8. Fredette, N.R.; Kavuri, A.; Das, M. Multi-step material decomposition for spectral computed tomography. *Phys. Med. Biol.* **2019**, *64*, 145001. [\[CrossRef\]](#)
9. Gürsoy, D.; Das, M. Single-step absorption and phase retrieval with polychromatic X rays using a spectral detector. *Opt. Lett.* **2013**, *38*, 1461–1463. [\[CrossRef\]](#)
10. Das, M.; Liang, Z. Spectral X-ray phase contrast imaging for single-shot retrieval of absorption, phase, and differential-phase imagery. *Opt. Lett.* **2014**, *39*, 6343–6346. [\[CrossRef\]](#)
11. Vazquez, I.; Fredette, N.R.; Das, M. Quantitative phase retrieval of heterogeneous samples from spectral X-ray measurements. In Proceedings of the SPIE Medical Imaging 2019: Physics of Medical Imaging, San Diego, CA, USA, 16–21 February 2019; p. 109484R.
12. Shikhaliev, P.M. Computed tomography with energy-resolved detection: A feasibility study. *Phys. Med. Biol.* **2008**, *53*, 1475. [\[CrossRef\]](#)
13. Schmidt, T.G. Optimal “image-based” weighting for energy-resolved CT. *Med. Phys.* **2009**, *36*, 3018–3027. [\[CrossRef\]](#) [\[PubMed\]](#)
14. Schmidt, T.G. CT energy weighting in the presence of scatter and limited energy resolution. *Med. Phys.* **2010**, *37*, 1056–1067. [\[CrossRef\]](#) [\[PubMed\]](#)
15. Vespucchi, S.; Park, C.S.; Torrico, R.; Das, M. Robust energy calibration technique for photon counting spectral detectors. *IEEE Trans. Med. Imaging* **2018**, *38*, 968–978. [\[CrossRef\]](#) [\[PubMed\]](#)
16. Barber, W.; Wessel, J.; Nygard, E.; Iwanczyk, J. Energy dispersive CdTe and CdZnTe detectors for spectral clinical CT and NDT applications. *Nucl. Instrum. Methods Phys. Res. Sect. A Accel. Spectrom. Detect. Assoc. Equip.* **2015**, *784*, 531–537. [\[CrossRef\]](#) [\[PubMed\]](#)
17. Brambilla, A.; Ouvrier-Buffet, P.; Rinkel, J.; Gonon, G.; Boudou, C.; Verger, L. CdTe linear pixel X-ray detector with enhanced spectrometric performance for high flux X-ray imaging. *IEEE Trans. Nucl. Sci.* **2012**, *59*, 1552–1558. [\[CrossRef\]](#)
18. Abbene, L.; Principato, F.; Gerardi, G.; Bettelli, M.; Seller, P.; Veale, M.C.; Zambelli, N.; Benassi, G.; Zappettini, A. Digital fast pulse shape and height analysis on cadmium–zinc–telluride arrays for high-flux energy-resolved X-ray imaging. *J. Synchrotron Radiat.* **2018**, *25*, 257–271. [\[CrossRef\]](#)
19. Wiegert, J.; Engel, K.J.; Herrmann, C. Impact of scattered radiation on spectral CT. In Proceedings of the SPIE Medical Imaging 2009: Physics of Medical Imaging, Lake Buena Vista, FL, USA, 14 March 2009; Volume 7258, p. 72583X.
20. Lewis, C.; Park, C.S.; Fredette, N.R.; Das, M. Impact of Compton scatter on material decomposition using a photon counting spectral detector. In Proceedings of the SPIE Medical Imaging 2017: Physics of Medical Imaging, Orlando, FL, USA, 11–16 February 2017; p. 101323T.
21. Sossin, A.; Rebuffel, V.; Tabary, J.; Létang, J.; Freud, N.; Verger, L. Influence of scattering on material quantification using multi-energy X-ray imaging. In Proceedings of the 2014 IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC), Seattle, WA, USA, 8–15 November 2014; pp. 1–5.
22. Sossin, A.; Rokni, M.; Brendel, B.; Daerr, H.; Thran, A.; Erhard, K. Experimental evaluation of the influence of scattered radiation on quantitative spectral CT imaging. In Proceedings of the SPIE Medical Imaging 2018: Physics of Medical Imaging, Houston, TX, USA, 10–15 February 2018; p. 105731B.
23. Siewerdsen, J.H.; Jaffray, D.A. Cone-beam computed tomography with a flat-panel imager: Magnitude and effects of X-ray scatter. *Med. Phys.* **2001**, *28*, 220–231. [\[CrossRef\]](#)
24. Kawrakow, I. *The EGSnrc Code System, Monte Carlo Simulation of Electron And Photon Transport*; NRCC Report Pirs-701; NRCC: Ottawa, ON, Canada, 2001.
25. Boone, J.M.; Lindfors, K.K.; Cooper, V.N., III; Seibert, J.A. Scatter/primary in mammography: Comprehensive results. *Med. Phys.* **2000**, *27*, 2408–2416. [\[CrossRef\]](#)

26. Hubbell, J.; Veigele, W.J.; Briggs, E.; Brown, R.; Cromer, D.; Howerton, d.R. Atomic form factors, incoherent scattering functions, and photon scattering cross sections. *J. Phys. Chem. Ref. Data* **1975**, *4*, 471–538. [[CrossRef](#)]
27. Hubbell, J.H.; Seltzer, S.M. *Tables of X-ray Mass Attenuation Coefficients and Mass Energy-Absorption Coefficients 1 keV to 20 MeV for Elements Z = 1 to 92 and 48 Additional Substances of Dosimetric Interest*; Technical Report; National Institute of Standards and Technology-PL: Gaithersburg, MD, USA, 1995.
28. Berger, M.; Hubbell, J.; Seltzer, S.; Chang, J.; Coursey, J.; Sukumar, R.; Zucker, D.; Olsen, K. XCOM: Photon Cross Sections Database. 2010. Available online: <http://www.nist.gov/pml/data/xcom> (accessed on 22 May 2018).
29. Fritz, S.L.; Chang, C.; Livingston, W.H. Scatter/primary ratios for X-ray spectra modified to enhance iodine contrast in screen–film mammography. *Med. Phys.* **1983**, *10*, 866–870. [[CrossRef](#)]
30. Jarry, G.; Graham, S.A.; Moseley, D.J.; Jaffray, D.J.; Siewerdsen, J.H.; Verhaegen, F. Characterization of scattered radiation in kV CBCT images using Monte Carlo simulations. *Med. Phys.* **2006**, *33*, 4320–4329. [[CrossRef](#)] [[PubMed](#)]
31. Kwan, A.L.; Boone, J.M.; Shah, N. Evaluation of X-ray scatter properties in a dedicated cone-beam breast CT scanner. *Med. Phys.* **2005**, *32*, 2967–2975. [[CrossRef](#)] [[PubMed](#)]



© 2019 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).