

Performance of a new co-axial ion-molecule reaction region for low-pressure chemical ionization mass spectrometry with reduced instrument wall interactions

Brett B. Palm¹, Xiaoxi Liu², Jose L. Jimenez², and Joel A. Thornton¹

¹Department of Atmospheric Sciences, University of Washington, Seattle, WA, USA

²Department of Chemistry and Cooperative Institute for Research in Environmental Sciences, University of Colorado, Boulder, CO, USA

Correspondence: Joel A. Thornton (joelt@uw.edu)

Abstract

Chemical ionization mass spectrometry (CIMS) techniques have become prominent methods for sampling trace gases of relatively low volatility. Such gases are often referred to as being “sticky,” i.e. having measurement artifacts due to interactions between analyte molecules and instrument walls, given their tendency to interact with wall surfaces via absorption or adsorption processes. These surface interactions can impact the precision, accuracy, and detection limits of the measurements. We introduce a low-pressure ion-molecule reaction (IMR) region primarily built for performing iodide-adduct ionization, though other adduct ionization schemes could be employed. The design goals were to improve upon previous low-pressure IMR versions by reducing impacts of wall interactions at low pressure while maintaining sufficient ion-molecule reaction times. Chamber measurements demonstrate that the IMR delay times (i.e., magnitude of wall interactions) for a range of organic molecules spanning five orders of magnitude in volatility are 3 to 10 times lower in the new IMR compared to previous versions. Despite these improvements, wall interactions are still present and need to be understood. To that end, we also introduce a conceptual framework for considering instrument wall interactions and a measurement protocol to accurately capture the time-dependence of analyte concentrations. This protocol uses short-duration, high-frequency measurements of the total background (i.e., fast zeros) during ambient measurements as well as during calibration factor determinations. This framework and associated terminology applies to any instrument and ionization technique that samples compounds susceptible to wall interactions.

27 **1 Introduction**

28 Trace gases in the atmosphere are drivers of the chemistry that determines air quality and climate
29 effects (Seinfeld and Pandis, 2006), as well as oxidant budgets and oxidation pathways (e.g., Crutzen, 1979; Di
30 Carlo et al., 2004) and secondary organic aerosol (SOA) formation (Shrivastava et al., 2017). Trace organic
31 compounds are particularly complex, spanning more than 15 orders of magnitude in volatility (Donahue et al.,
32 2012; Hunter et al., 2017; Isaacman-VanWertz et al., 2018). Large gaps remain in our knowledge of the
33 chemistry and impacts of trace organic gases, in particular the lower volatility compounds (Goldstein and
34 Galbally, 2007). The ability to measure and quantify such lower volatility gases is an evolving analytical
35 measurement challenge, but remains a limiting factor in our ability to test important theories governing, e.g.,
36 organic gas-particle partitioning, oxidation mechanisms, SOA formation, vertical distributions, and dry
37 deposition.

38 Many of the recent advances in knowledge of atmospheric trace gases, particularly the lower volatility
39 compounds in the gas phase, have been due to the application and development of advanced instrumentation
40 (Mohr et al., 2013; Ehn et al., 2014; Isaacman et al., 2014; Krechmer et al., 2016a; Peng et al., 2016; Yuan et al.,
41 2017). One of these major advances has been the development of field-deployable mass spectrometers,
42 combined with the development of specialized inlets allowing the application of various chemical ionization
43 methods to atmospheric compounds (Ehn et al., 2014; Lee et al., 2014; Krechmer et al., 2016a). In chemical
44 ionization, analyte molecules are imparted an electrical charge either by charge transfer from or clustering with
45 a reagent ion, processes which are relatively low energy and typically induce little fragmentation of the analyte
46 molecules. A variety of reagent ions with the ability to ionize different subsets of analyte molecules have been
47 used, including H_3O^+ , acetate, iodide (I^-), nitrate (NO_3^-), ammonium (NH_4^+), and others (e.g., Jokinen et al., 2012;
48 Yatavelli et al., 2014; Zaytsev et al., 2019). Iodide-adduct ionization in particular has been used for both gas and
49 particle composition measurements, and is sensitive to a wide range of inorganic and organic molecules (e.g.,
50 Huey et al., 1995; Le Breton et al., 2012; Lopez-Hilfiker et al., 2013; Mohr et al., 2013; Lee et al., 2014; Veres et
51 al., 2015; Gaston et al., 2016; Lee et al., 2016).

52 One impediment to the measurement of lower volatility gases is the influence of inlet tubing and other
53 experimental apparatus surfaces. Several recent experiments have probed the effects of Teflon chamber walls
54 on experimental processes (Matsunaga and Ziemann, 2010; Krechmer et al., 2016b; Krechmer et al., 2017;

55 Huang et al., 2018). A variety of organic and inorganic gases have been shown to reversibly absorb into Teflon
56 (and other polymer) tubing or reversibly adsorb onto the surface of a variety of solid materials including
57 stainless steel (Pagonis et al., 2017; Deming et al., 2019; Liu et al., 2019). Current CIMS instrumentation typically
58 requires the use of such materials in the design of the inlet tubing as well as the ion-molecule reaction (IMR)
59 region where chemical ionization occurs, which allows for wall interactions to occur. The rates of flux of analyte
60 molecules to and from these wall surfaces can depend on complex factors of water vapor concentration, co-
61 analyte concentrations, etc. (Pagonis et al., 2017; Deming et al., 2019; Liu et al., 2019), leading to difficult
62 interpretations of data that is often not consistent across different studies.

63 Past CIMS IMR versions have employed different designs, typically constructed from varying fractions of
64 stainless steel and several types of Teflon (Eisele and Tanner, 1993; Jokinen et al., 2012; Lee et al., 2014; Zhao et
65 al., 2017; Lee et al., 2018). Some IMR designs, such as the NO_3^- CIMS, can operate at ambient pressure with an
66 IMR design that essentially eliminates wall interactions (e.g., Krechmer et al., 2015). However, the NO_3^- reagent
67 ion is sensitive only to a narrow subset of highly-oxidized molecules with which it has clustering strengths
68 greater than its cluster with HNO_3 . The I^- CIMS technique is sensitive to a much broader range of analyte
69 molecules, making it a powerful technique for studying atmospheric chemistry. But, I^- can also cluster with one
70 or more water molecules, causing the sensitivity of I^- toward other analyte molecules to be dependent on water
71 vapor concentrations in the IMR. To reduce this water vapor dependence, the IMR is typically operated at low
72 pressure (~2-200 Torr) to reduce the partial pressure of water vapor. For aircraft I^- CIMS measurements, a low-
73 pressure IMR has also been desired in order to allow pressure control systems to maintain constant pressure in
74 the ionization region with changing pressure/altitude, thus maintaining constant sensitivity to clustering
75 (Neuman et al., 2002; Crounse et al., 2006; Le Breton et al., 2012; Lee et al., 2018). In order to operate at low
76 pressure, the I^- CIMS must sample through an orifice, necessitating wall interactions in the IMR. Accounting for
77 the flux of analyte from the IMR walls is a challenge of particular importance to aircraft measurements, where
78 ambient concentrations can change rapidly on the edges of spatially narrow plumes from point or regional
79 sources such as power plants, biomass burning, or urban areas. Background measurement and subtraction from
80 the total observed signal is typical (Neuman et al., 2002; e.g., Crounse et al., 2006; Veres et al., 2008; Lee et al.,
81 2018), however a uniform standard method for background subtraction does not exist, and methods applied by
82 different research groups vary widely.

In this work, we present a new design of a co-axial, low pressure IMR to minimize wall interactions, incorporating knowledge acquired in the operation and analysis of past IMR designs. A detailed consideration of the process of sampling through an instrument inlet is presented, explaining how the measured signal is influenced by wall interactions. We suggest practices for accounting for wall interactions, both in experimental measurements and when performing calibration measurements that will be later applied to experiments. Finally, this new IMR design was characterized by measuring the magnitude of wall interactions of several organic compounds spanning a wide range of volatility. Both the new IMR design considerations and the broader discussion of wall interactions will be applicable to a broader community of analytical atmospheric chemistry.

2 Co-axial low-pressure IMR design

2.1 I⁻ CIMS method

Iodide-adduct chemical ionization has been described in detail in previous studies (Huey et al., 1995; Lee et al., 2014; Lee et al., 2018). Briefly, I⁻ anions are produced by passing methyl iodide (CH₃I) in nitrogen through alpha particles from a polonium-210 radioactive sealed source. The anions form adducts by colliding with neutral analytes inside an IMR, and the clusters are subsequently sampled by a high-resolution time-of-flight mass spectrometer (HR-ToF-MS; ToFwerk AG, Thun, Switzerland). This spectrometer provides a nominal mass resolving power ($m/\Delta m$) of approximately 5000 with pptv level detection limits for most compounds.

2.2 IMR description

Many different IMR designs have been employed in past CIMS measurements, each with advantages and disadvantages. The primary function of any IMR region in a CIMS is to facilitate the process of imparting an electrical charge onto analyte molecules in the sample air, whereupon they can be manipulated and analyzed inside the mass spectrometer. Depending on which reagent ion is chosen and which analyte molecules are targeted, the IMR will have different design requirements. Recent interest in identifying and quantifying a broad range of reactive and/or low volatility compounds presents substantial challenges for CIMS instruments with low-pressure ionization regions, including but certainly not limited to the I⁻ CIMS used in this work. The effects of IMR wall interactions can be a substantial impediment to making accurate and easily interpretable measurements of compounds that react on or reversibly partition to reactor walls.

Herein we describe the design of a new co-axial IMR, illustrated in the schematic in Fig. 1. This design aimed to improve upon that most recently employed by the Thornton research group during the WINTER 2015 research flights, which has been described in detail in Lee et al. (2018). That version was itself a design built to improve upon the characteristics of previous versions of the IMR including the model available commercially from Aerodyne Research Inc. with the mass spectrometer (Kercher et al., 2009; Bertram et al., 2011; Lee et al., 2014). In the commercially available low-pressure IMR, the analyte flow and ion flow are mixed via turbulence inside a region constructed out of stainless steel. In addition to the increased wall interactions that result from turbulence, stainless steel has been shown to suffer from enhanced wall effects for many compounds (Deming et al., 2019; Liu et al., 2019). The WINTER IMR made improvements by decreasing the wall surface area and residence time of the turbulent region, and also by constructing two of the three walls of the cylindrical IMR region out of machined PTFE Teflon (Lee et al., 2018). However, the third wall remained stainless steel, and turbulence remained an issue.

The main goals of our improved IMR design were to reduce wall effects while maintaining sufficient residence time for clustering (i.e., maintain sufficient sensitivity). The initial strategies were to remove as many wall surfaces as possible, and have any necessary wall surfaces be constructed from materials such as perfluoroalkoxy (PFA) Teflon which have been shown to have the weakest interactions with many analytes (Pagonis et al., 2017; Deming et al., 2019; Liu et al., 2019). To minimize wall effects further, we aimed to inject the sample flow into a co-axial sheath of ion flow, creating a larger distance between analyte and surfaces. This design feature was similar to what has been used in some previous IMR designs, in particular for the NO_3^- reagent ion (Eisele and Tanner, 1993; Jokinen et al., 2012; Massoli et al., 2018). Furthermore, we aimed to pump the flow out of the IMR in a similar manner to how it was injected, pumping a sheath flow radially outside of a sample flow. Any analyte that desorbed from a wall surface would be more likely pumped out in the sheath flow and not sampled into the MS.

The final design requirement was that the IMR was capable of operating at a constant IMR pressure on an aircraft platform, where ambient pressure can span the range from ~200-760 Torr. Ion-molecule reaction rates scale with total analyte number density, and ion-molecule cluster stability will depend on total pressure as well as H_2O partial pressure (Lee et al., 2014; Iyer et al., 2016; Lopez-Hilfiker et al., 2016). Thus, maintaining constant pressure (and temperature) can minimize changes in instrument response with large changes in

altitude. This feature was added to the WINTER version of the IMR by incorporating a variable orifice on the upstream side of the IMR (Lee et al., 2018), and it is also included in this new co-axial IMR. As long as the pressure downstream of the orifice remains roughly less than half of the pressure of ambient air upstream, critical flow is achieved in the orifice (i.e., the speed of the air through the orifice is approximately the speed of sound). The mass flow through the orifice is then only a function of upstream pressure. As upstream pressure changes with altitude, the variable orifice can be opened or closed via computer control to maintain constant mass flow into the IMR. As the pumps maintain constant mass flow out of the IMR, the pressure inside the IMR remains constant at ~70 Torr downstream inside the IMR where I⁺ is introduced and ionization occurs.

The benefits of constant reduced pressure, e.g. stable instrument response and reduced effects of water vapor on ionization efficiency, come with enhanced wall interactions which can contribute potentially large and often poorly understood artifacts to the measurement. The pressure drop between ambient pressure and ~70 Torr leads to a high velocity jet expansion, which induces turbulent mixing. The jet-induced turbulence ensured mixing of reagent ions and sample flows in previous IMR designs, but also enhanced contact of the sample flow with IMR surfaces. Moreover, the low pressure leads to an order of magnitude larger diffusivity compared to ambient pressure, such that even in the absence of jet induced turbulence, gases in the sample flow will randomly reach the walls of the IMR more efficiently than at typical ambient pressures. Consistent with these ideas, it has been previously shown that the low pressure IMR is the main source of instrument memory and reactive trace gas losses, not the ~0.5 m long sampling inlet at ambient pressure with fast (~10-20 slpm) flow rates typically used (Lee et al., 2018).

Given the above considerations, the first design challenge was to slow the sample flow rate down by expanding the flow cross section while limiting turbulent mixing of analyte molecules to wall surfaces. In order to expand the flow without causing turbulence, an expansion cone/diffuser with an angle of less than approximately 5–7 degrees could be used. Fluid dynamics simulations have shown that this method can prevent flow separation that leads to turbulence in expansions, though possibly not for the Reynold's numbers of less than 2000 in this IMR (Sparrow et al., 2009, and references therein). This cone angle would require a length of more than 13 cm. Diffusion calculations suggest that one third of the analyte would contact the diffuser wall surface under laminar conditions, which still requires getting the flow laminar after the orifice. Given these considerations, as well as time constraints prior to a field campaign, we opted not to test a conical diffuser at

166 this time. Instead, the jet of air exiting the orifice was allowed to expand immediately into a fluorinated
167 ethylene propylene (FEP) Teflon-lined cylinder with 1.2 cm diameter and 1 cm length, after which it passed
168 through a parallel cluster of 3.175 mm OD, 1.5875 mm ID (0.125 inch OD, 0.0625 inch ID) PFA Teflon tubes with
169 a length of 1.5 cm. Turbulence was limited to the 1.2 cm diameter cylinder, and then the subsequent tubing
170 cluster acted to develop laminar flow. As a rough approximation, turbulent flow can be converted to laminar
171 flow by passing through a tube with an entrance length that is 10 times its diameter (Çengel and Cimbala,
172 2014). This concept guided our design specifications. When the sample air exits the laminizer element, the flow
173 has slowed down and become much less turbulent, mitigating the effects of walls downstream of that point.
174 Since having an orifice upstream of the IMR effectively necessitates having some region of turbulence in contact
175 with walls, this design strategy was aimed at limiting the residence time and amount of wall surface area in the
176 region of the IMR that encountered turbulent sample gas. Future low-pressure IMR designs could aim to further
177 minimize wall effects in this region directly downstream of the variable orifice.

178 While the sample gas enters the IMR through the orifice, the I^- anions are concurrently injected into a
179 region of the IMR concentric with and outside of the sample flow laminizing element. The anions are produced
180 by flowing dry N_2 over a permeation tube containing methyl iodide and then through a Po-210 radioactive
181 sealed source, producing I^- . The ion flow experiences some turbulence when injected into the IMR, and then
182 passes through a parallel cluster of 6.35 mm OD, 3.175 mm ID (0.25 inch OD, 0.125 inch ID) PFA Teflon tubes
183 with a length of 1.27 cm (0.5 inch) that act as a laminizer element. The flow coming out of both the sample flow
184 laminizer and ion flow laminizer exit in the same plane and can be arranged to have approximately the same
185 velocity in the axial direction. In this work, this was achieved by maintaining a constant 2 slpm sample flow and
186 3 slpm ionizer flow. As part of the process of designing the IMR with laminizers, fluid modeling simulations were
187 performed to visualize the effects of turbulent vs. laminar flows. Two example cases are depicted in Fig. S1.

188 The exit of the laminizers marks the start of the drift region in the IMR where interactions of analyte
189 with I^- anions occur. Within the 3.49 cm (1.375 inch) ID, 3.81 cm (1.5 inch) long drift region, the I^- anions and
190 analyte flows mix together via diffusion, possibly aided by some residual turbulence. The design also includes
191 some exposed stainless steel surfaces on the drift region wall and at the exit of the sample flow laminizer and
192 entrance of sample pump flow tube, as far from the main sample flow as possible to limit wall interactions.
193 These surfaces can be used to apply an electric field inside the IMR to attempt to enhance the mixing of ions

194 into the sample flow. However, only modest total detected ion enhancements were measured when applying
195 such electric fields. We hypothesize that the relatively high diffusivity at 70 Torr, as well as any residual
196 turbulence, were dominating the flow mixing instead of the electric field forces in this particular design.
197 Because of the only modest gains and in the interest of simplicity, all exposed metal surfaces were grounded
198 together and electric fields were not employed during the measurements discussed herein.

199 Because the analyte molecules enter the drift region in the center, they would have to diffuse all the
200 way across the ion flow to reach a wall surface. In order to be sampled after encountering a wall, they would
201 also have to diffuse all the way back across the ion flow to the capillary into the mass spectrometer. To prevent
202 any molecules coming from the drift tube wall being sampled, half of the drift tube flow was pumped out along
203 the drift tube wall and away from the MS capillary. According to diffusion calculations, only 4% of the analyte
204 are predicted to encounter a wall in the drift region under laminar flow conditions, and a small fraction of those
205 molecules would diffuse back to the center to be sampled, essentially removing the effects of the drift region
206 walls. The other half of the drift region flow was pumped through an FEP Teflon-lined sample tube with ID of
207 2.18 cm (0.86 inch) and length of 5.08 cm (2.0 inch) and past the MS capillary, where it was sub-sampled into
208 the mass spectrometer.

209 Limiting the interaction between analyte and wall surfaces also limits the possibility of the analyte
210 undergoing chemical reactions on surfaces. To examine and quantify the improvements made in this design, we
211 start with a comprehensive discussion of the origin and meaning of wall effects. Although wall interactions are
212 not the only source of instrumental background signals, for semi-volatile and low volatility compounds they are
213 often the dominant source of residual non-ambient signal. The concept of background signal will be examined
214 using laboratory measurements, and further discussed in the context of ambient measurements and instrument
215 response calibrations. The improvements will be assessed by comparing laboratory measurements made with
216 this IMR to previous measurements from other IMRs and instruments.

217 **3 The effects of instrument wall surfaces**

218 **3.1 Measuring and subtracting instrument background signal**

219 In order to properly evaluate the new IMR design, we must first introduce a common framework that
220 can be used to describe how inlet tube and IMR wall interactions originate, what their effects are, and how they
221 can be understood. The CIMS experimental setup will be defined here as comprised of two parts: the sampling

222 tube (i.e., inlet) which transports the analyte from the sampling location (outside of aircraft, inside chamber,
223 etc.) to the IMR; and the IMR, where ionization occurs prior to entering the MS. The IMR is defined as part of
224 the instrument. The background signal is typically measured by flooding the sampling tube and/or IMR with
225 clean air or ultra-high-purity nitrogen (UHP N₂). Subtracting the resulting “background signal” from the total
226 signal measured while sampling ambient air is a common practice in atmospheric mass spectrometry. However,
227 the exact definition and quantification procedure of the ‘background’ can vary across different experimental
228 configurations and analysis goals. The processes that lead to the background signal can also be dynamic and
229 controlled by multiple factors.

230 The background signal can originate from molecules coming from either the sampling tube or the IMR.
231 In many cases, the sampling tube can be designed such that its background effects are small relative to the IMR
232 effects, e.g., by pulling a large flow through the inlet and subsampling into the IMR, thus minimizing inlet
233 residence time and also diluting the flux from the walls into a large flow volume. Sampling at ambient pressure
234 in the sample tube also minimizes diffusivity to and from the walls. The IMR walls have been shown to be the
235 dominant source of background signal in previous field measurement setups (Lee et al., 2018), so this discussion
236 will focus mainly on IMR background signal. The details and concepts discussed here of background signal
237 sources and how to quantify them are not specific to the I⁻ CIMS IMR, but can be adapted to other IMRs and
238 ionization types as well as for sampling tubes. The concepts involved are illustrated in Fig. 2a and demonstrated
239 using laboratory measurements in Fig. 2b, where a constant gas-phase concentration of nitric acid (HNO₃) was
240 injected into a short polytetrafluoroethylene (PTFE) Teflon inlet tube (~20 cm length, 0.75” diameter, 20 slpm
241 flow rate) and subsampled into the IMR in the sample flow for a specified amount of time. The effects of wall
242 interactions in such an inlet are minor relative to the effects of wall interactions inside the IMR (as
243 demonstrated in Fig. 2). The schematic in Fig. 2a and the following discussion applies mainly to analyte
244 molecules that partition reversibly to the walls (or to thin films of water adsorbed on the walls, as is the case for
245 HNO₃; Liu et al., 2019), and for wall surfaces that allow for absorption such as Teflon varieties. Adsorbing
246 surfaces such as stainless steel, and irreversible loss of analytes such as many radical species, will be discussed
247 as exceptions.

248 At the theoretical time $t=t_0$ in Fig. 2, consider an IMR that has never previously sampled a specific
249 analyte molecule in the sample flow. Prior to t_0 , there will be no signal at all from this analyte entering in the

sample flow, and the only signal corresponding to that analyte will be defined here as the persistent background, due to electronic noise and other baseline signal sources such as the ion source or carrier flows. In the specific case of HNO_3 in the Fig. 2b example, a substantial persistent background exists due to a source in the ion flow from the Po-210 ionizer. Most analytes will not have such a persistent background. At $t=t_0$, the analyte has entered the IMR and experienced one of the following two fates: 1) traveled directly from outside of the IMR to the detector in the gas phase without interacting with a wall surface (which may include bouncing off of a wall surface without interacting), or 2) absorbing in (or adsorbing on) a wall surface, where it remains for some amount of time longer than the residence time of the IMR before desorbing and being sampled to the detector. The fraction of analyte that follows each of these two paths will be a function of instrument design (i.e., what fraction of sampled air collides with a wall surface through turbulence or diffusion) and as a function of the uptake and partitioning coefficients of each analyte on each wall surface type. The uptake coefficients themselves will be a function of the exact environmental conditions of the wall surfaces at the time of collision. These environmental conditions can modify the wall surfaces and change how gases are taken up into/on surfaces or change how they desorb from the surfaces.

The most influential surface modifier is often water. The analyte can behave differently depending on whether it encounters a bare Teflon or stainless steel surface under completely dry conditions, a surface coated in a monolayer of water under low RH conditions, or a surface coated with a thick layer of water that causes an aqueous diffusion limitation to the analyte interacting with the actual surface. Liu et al. (2019) demonstrated that some polar compounds partition to walls as a function of their Henry's Law constants during humidified conditions. This IMR design has the ability to add water vapor directly downstream of the variable orifice as in Lee et al. (2018). This maintains a relatively narrow range of water vapor concentrations in the IMR regardless of the sample air humidity, keeping the environmental conditions (and uptake/partitioning coefficients) in the IMR roughly constant. Surfaces can also be modified by other analyte molecules, which essentially act in competition for surface sites. This behavior has been observed for materials such as stainless steel that are dominated by adsorption to a limited number of surface sites (Deming et al., 2019). While absorbing materials such as Teflon have been shown to be modified by water, they appear to be insensitive to the amount of other analytes absorbed in the surface (Pagonis et al., 2017; Deming et al., 2019; Liu et al., 2019), at least at analyte concentrations relevant to the atmosphere and typical laboratory chamber experiments.

As soon as there are analyte molecules ad/absorbed on surfaces, there will be a flux of that analyte from the surface back into the sample/ion flow. The flux from the surface will be a function of the amount of analyte on the surface, as well as the environmental conditions such as temperature, humidity, and history (see above). Therefore, in the moments just after $t=t_0$, e.g. $t=t_1$ in Fig. 2, there will be a flux of analyte from the walls. We define this flux as the source for the dynamic background signal, which is separate from the persistent background signal. Any analyte that is entering the IMR at this time will continue to split between reaching the detector directly or absorbing into the walls first, at the same fractional rates. These fractional rates will be constant as long as the environmental conditions remain constant, and the rates will not be a function of the flux of that analyte coming off of the wall. At $t=t_1$, the total flux into the IMR is greater than the total flux to the detector, and there is a net flux to the wall surfaces. As more analyte continues to enter the inlet and ad/absorb on the walls, the flux of analyte from the wall will continue to grow until a time comparable to $t=t_2$ in Fig. 2. Any analyte that partitions irreversibly to the walls or desorbs as a different compound due to surface reaction would appear to have no flux from the walls and no dynamic background signal. Only the fraction of such an analyte that did not interact with the walls would be detected, potentially at much lower sensitivities than expected from ionization efficiency considerations (Lopez-Hilfiker et al., 2016).

At times equivalent to $t=t_2$ in Fig. 2, the flux of reversibly-partitioning analyte from the wall has grown to be equal to the rate of ad/absorption of the analyte to the wall. The wall system is now in steady state. The amount of analyte arriving at the detector is now equal to the sum of the analyte that did not interact with walls and the analyte that entered the IMR at some earlier time, interacted with a wall, and then desorbed to reach the detector. Because the flux from the walls is equal to the flux to the walls, the total flux to the detector is equal to the total flux of analyte that is entering the IMR at that time. That is, the total signal is the same as it would be if the analyte were introduced into an IMR completely absent of wall interactions. This condition is only true when the incoming analyte concentration and environmental conditions have remained constant for long enough to establish wall steady state. As shown in Fig. 2b, the only signal that stays constant during a constant concentration injection with wall interactions is the background-subtracted signal. The background signal and thus also the total detected signal change over time and are both non-deterministically related to the analyte concentration entering the inlet. This concept is critical for the time-dependent quantification of analyte

305 in the sampled air, and is also important for the determination and interpretation of calibration factors, as
306 discussed later in Sect. 3.2.2.

307 The ratio of the background-subtracted signal to the background signal will remain constant after time
308 $t=t_2$, as long as the environmental conditions in the IMR remain constant. However, the ratio will not be the
309 same for all analytes. For analytes which are more volatile (or less soluble in water), interact with the wall
310 surfaces less strongly, and desorb more rapidly, the background signal may be negligible relative to the
311 background-subtracted signal (and the background-subtracted signal will be essentially equal to total signal).
312 For analytes which are less volatile (more soluble), interact strongly with surfaces, and desorb slowly, the
313 background signal may become a large majority of the total signal at the detector and the background-
314 subtracted signal may reach a detection limit. The IMR geometry and design will largely determine which
315 compounds qualify as 'more' and 'less' volatile on this relative scale. For instance, the NO_3^- CIMS and the cross-
316 flow ion source (Zhao et al., 2010; Zhao et al., 2017) which operate with laminar flows at atmospheric pressure
317 in the IMR thereby minimizing turbulence and diffusion to walls, both employ geometries that prevent sampling
318 of any analyte that encountered a wall surface in the ionization region, leading to an IMR background signal flux
319 that is essentially negligible. Atmospheric pressure sampling, which as noted above is ultimately the source of
320 such benefits, may not be suitable for an aircraft platform as discussed above.

321 Continuing our description of the evolution of wall interactions, consider that the source of the analyte
322 into the IMR is completely removed immediately following $t=t_2$. For instance, this could represent the injection
323 of analyte-free air during a background measurement, or a scenario where the sampled air transitions rapidly
324 from high concentrations of the analyte in a plume to very low concentrations outside of a plume. There will be
325 a short transition period, corresponding to the residence time distribution of air in the IMR downstream of
326 where analyte-free air is injected (approximately 100 ms on average in the IMR described herein) plus any time
327 for switching flows outside the ionization region (potentially several seconds), when the analyte-laden air is
328 replaced with analyte-free air in the IMR. The flux of analyte to the detector without wall interactions and the
329 flux of analyte to the walls both drop to zero at this point, which is specified as $t=t_3$. There remains essentially
330 the same amount of analyte adsorbed on the walls at $t=t_3$ as at $t=t_2$ immediately prior, so the flux from the
331 wall to the detector continues to provide the same dynamic background signal.

After more time passes and $t=t_4$ has been reached, the amount of analyte on the walls has been partially depleted since the wall system is now out of steady state. There is still a flux from the wall without a complementary flux to the wall to replenish the analyte. The flux from the wall is also lower at this time than at $t=t_3$ because the concentration of analyte on the wall is lower. At a subsequent time long after $t=t_4$, all of the analyte would eventually desorb from the walls, and the dynamic background signal from the inlet walls would reach zero, equivalent to a time $t<t_0$. As discussed further in Sect. 3.3, the amount of time required for the dynamic background signal to decay to 10% of the original signal (i.e., to near zero) can range from less than 1 s to tens of min or more, depending on the volatility of the analyte as well as environmental conditions and surface types. For some analytes (including HNO_3 in the iodide anion source discussed here), there can be other persistent sources of background signal coming from the tubing and carrier gas related to the ion source. The persistent background is present at all times from $t<t_0$ to $t>t_4$, and can be quantified by injecting analyte-free air for sufficient time to completely deplete the dynamic background signal of interest here. The persistent background is also included in the signal measured during clean-air injections at $t=t_3$.

The main goal of measuring and subtracting the background signal in an instrument is to ascertain the concentration of the analyte present in sampled air at the time of sampling with high temporal/spatial resolution, removing the instrumental artifacts related to the background signal caused by wall interactions. As illustrated in Fig. 2a, this task is often made complicated by the fact that the ratio of the background signal to the background-subtracted signal can vary widely during measurements. The entire signal could be due to background signal (as at $t=t_3$), due to gas phase signal ($t=t_0$), or some dynamic mix of the two ($t=t_1$ and $t=t_2$). Even when all signal is coming from the background, the magnitude of the background can also change ($t=t_4$).

Given these fluctuating factors combined with a potentially rapidly changing sampling environment due to a moving aircraft platform or rapidly shifting air masses with different source characteristics, the ideal way to determine the true concentration of the analyte in sampled air is to measure the amount of signal coming from the background sources at all points in time and subtract it from the total signal. But akin to Heisenberg's uncertainty principle, one cannot precisely measure both the total signal and the background signal at the same time. Instead, a practical method for determination of background signal is to measure the instantaneous flux of analyte off the walls using high-frequency, short-duration injections of analyte-free gas (typically UHP N_2) interspersed among the normal measurement of total signal, and then interpolate between these background

360 measurements. This method is has been referred to as performing ‘fast zeros’. Upon injection of analyte-free
361 gas, the measurement transitions from representing the total signal (equivalent to $t=t_1$, t_2 , or t_3 , depending on
362 whether wall steady state has been achieved) to a measurement of just the sum of dynamic and persistent
363 background signals (equivalent to $t=t_3$).

364 As seen in the inset of Fig. 2b, the decay of the analyte signal occurs in two parts (or more). The first
365 part is the rapid exponential decay as the volume of the inlet is cleared out of any remaining gas-phase analyte,
366 and stability of flows is achieved, etc. The next part, which applies when the analyte is of relatively lower
367 volatility or higher Henry’s Law constant into wall-adsorbed water, is the typically slower exponential decay that
368 accompanies desorption of the analyte from the walls. There may be multiple decay constants with varying time
369 scales (e.g., as illustrated in Krechmer et al., 2018) if there are multiple types of wall surfaces (e.g., both Teflon
370 and stainless steel in the same IMR) or voids with different residence times. To know the background value at
371 the time when the background measurement was initiated ($t=t_2$), one needs to know the magnitude at which
372 the slower exponential decay begins, i.e. the signal value at $t=t_3$ shown in the inset of Fig. 2b. The background
373 determined at successive times of $t=t_3$ are then interpolated to estimate the background at all points in time.
374 Such periodic background determinations would also inherently account for any changes in the environmental
375 conditions that would change the analyte uptake coefficient and thus the ratio of the flux to the walls vs the flux
376 to the detector without wall interaction, such as an aircraft platform flying through varying ambient H_2O
377 concentrations. In other words, as long as the background signal can be determined at a given time, it does not
378 matter when those particular analyte molecules that led to background signal entered the IMR.

379 Any background measurement value taken at a later time, e.g., at $t=t_4$ or at $t>t_4$ (a measure of the
380 persistent background), would no longer represent the magnitude of the background at $t=t_2$ and would
381 underestimate the contribution of background signal to the total at the time the background measurement was
382 initiated. This aspect is critical to the determination of so-called tails of measurements, e.g., when an aircraft
383 platform is measuring in an analyte plume and then abruptly exits the plume to analyte-free air. The signal
384 appears to decay as between $t=t_3$ and $t=t_4$ (and beyond) in Fig. 2b. The entirety of this signal is often due to
385 background signal. If this background signal is not subtracted as described herein, the data would be falsely
386 reporting a non-zero concentration (i.e., tail) of the analyte after exiting the plume, which could lead to large
387 errors in measurement-model comparisons that would not be captured by simple uncertainties estimated by

388 replicate calibrations. Note that when calculating the integral of signal across a plume pass, the same integrated
389 concentration can be found whether or not the background signal subtraction method is used, provided that a
390 self-consistent calibration value is applied (see Sect. 3.2).

391 With the IMR described in this work as well as previous versions, it was found that a ‘fast zero’
392 background measurement of 6 s duration was sufficient to pass through the fast exponential decay (which
393 typically lasts ~ 2 s) and capture $t=t_3$ at the start of the slow exponential decay of analyte. The frequency at
394 which the background measurement needs to be taken depends on the application. For measurement of rapidly
395 changing analyte concentrations, the background needs to be determined as rapidly as possible to minimize
396 errors in interpolation of the background. For recent aircraft measurements using this IMR, these 6 s
397 background measurements were performed once per minute, striking a balance between minimizing
398 background interpolation errors while maximizing the duty cycle of taking ambient measurements. One could
399 imagine taking a 6 s background (or shorter, e.g., 4 s) as fast as every 20-30 seconds to capture extremely rapid
400 changes in some specific circumstances, but information about the same temporal changes in background-
401 subtracted signal would be lost. Conversely if the analyte concentrations are known to be relatively constant,
402 e.g., in a laboratory experiment, then the intervals between background determinations could stretch much
403 longer without leading to substantial interpolation errors. Linear interpolation can be the simplest method,
404 however other methods could be used depending on specific circumstances. For instance, a relative-
405 concentration-dependent interpolation may better describe the background signal for a case where a plume
406 with large concentration gradient was entered and/or exited between background determinations.

407 **3.2 Wall Interactions and Calibration of Instrument Response**

408 The previous section discussed accounting for dynamic background signals in the context of
409 determining accurate gas-phase concentrations in laboratory or field experiments. Also important is to account
410 for the background signal during instrument response calibrations. When calibrating, a known amount of an
411 analyte is injected into the instrument, and the amount of raw signal measured per unit analyte is determined.
412 This raw signal has to be normalized to a constant number of reagent ions, given that the total number of ions
413 created by an ion source (and thus clusters formed and signal measured) can change with time. Therefore, a
414 calibration value for this I^- CIMS typically has units of counts per second per 1×10^6 total reagent ion count
415 (TRIC) per ppt of analyte, also called normalized counts per second (ncps) per ppt of analyte. When the raw

416 signal in units of ncps is divided by the calibration value, a concentration in units of ppt is derived. The
417 calibration value for each analyte must also be determined as a function of the amount of water vapor in the
418 IMR.

419 However, in light of the earlier discussion of background signal, considering the signal as units of ncps is
420 not enough information. The distinction between background-subtracted ncps and background (including
421 dynamic and persistent background) ncps, which add to total ncps, is necessary. As illustrated in Fig. 2b, when a
422 constant concentration of HNO_3 (approx. 2 ppbv) from a permeation tube was added into the inlet, neither the
423 background counts per second nor the total counts per second were constant functions of the amount of HNO_3
424 injected. The background-subtracted counts per second were constant, making that value the only properly
425 deterministic calibration constant that can be applied regardless of the relative amounts of background vs.
426 background-subtracted signals. Therefore, it is also recommended that the same background subtraction be
427 performed on both calibration data and field/laboratory measurement data. Also, care should be taken to
428 ensure that wall steady state is achieved in any tubing that is used to transfer a calibration gas from its source
429 to the IMR, such as in the PFA Teflon tubing between the HNO_3 permeation tube and the IMR used in this work.
430 This ensures that the flux of HNO_3 coming out of that transfer line is the same as the calibrated flux out of the
431 permeation device.

432 If background measurements are not performed or are not possible in a certain configuration, an
433 alternative method may be used in specific circumstances. The calibration constant could be measured as the
434 total ncps ppt^{-1} during a time equivalent to $t=t_2$, when wall steady state has been achieved. This calibration
435 constant represents the total (background-subtracted plus background) amount of signal that a given incoming
436 gas-phase concentration will generate. It applies only at wall steady state, only when environmental conditions
437 (e.g, RH) are the same as during calibration, and only in a given inlet configuration. Therefore, the calibration
438 can only be applied to data that has not been background subtracted, and it will only be accurate when wall
439 steady state has been achieved and environmental conditions are the same as during calibration. For laboratory
440 measurements, these conditions may be achieved if special care is taken (e.g., flow tubes, oxidation flow
441 reactors, or chambers operated in reproducible steady-state modes). However, dynamic conditions in field
442 studies likely preclude this calibration method from being a routinely viable option for analytes with substantial
443 background signal. The integral of signal across a plume would still be accurate (i.e., mass balance is achieved in

the IMR), but the real concentration would be underestimated at the start of the plume, and overestimated in the tail of the plume, provided that there are no signal tails from previous plumes still desorbing from surfaces. If sufficient background measurements were taken during a measurement period, but the calibration constant applied to that data was calculated using the total signal, the calibration constant can be retroactively converted to units of background-subtracted ncps ppt⁻¹ by finding a suitable time when wall steady state was achieved during the measurement period. The ratio of background-subtracted signal to total signal during wall steady state can be derived and multiplied by the total signal calibration constant to obtain the background-subtracted calibration constant, using the following equation:

$$\frac{\text{Background-subtracted ncps}}{\text{ppt}} = \frac{\text{Total ncps}}{\text{ppt}} \times \frac{(\text{Background-subtracted ncps})_{ss}}{(\text{Total ncps})_{ss}}, \quad (1)$$

where the subscript ss implies the value at wall steady state. Lastly, it will be important to keep this relationship between background-subtracted and total calibration constants in mind when comparing experimentally derived sensitivities to theoretically calculated sensitivities (as in, e.g., Iyer et al., 2016; Lopez-Hilfiker et al., 2016; Sekimoto et al., 2017). The theoretical calculations will be estimating the total signal per amount of analyte, without regard to wall effects.

In summary, the accuracy of a calibration constant will depend on how the wall interactions for an analyte are quantified during calibration and ambient measurements. For sticky compounds with substantial wall interactions, systematic biases in instrument response, and thus reported concentrations, can easily approach a factor of 2 (or much more in signal ‘tails’) without a self-consistent accounting of wall induced backgrounds during calibrations and measurements.

4 Quantifying IMR delay times

4.1 Chamber measurement methods

The wall interactions in the IMR designed in this work were characterized through a series of experiments, including extensive tests performed in the University of Colorado Environmental Chamber Facility in Boulder, CO. The chamber contained a 20 m³ FEP Teflon bag operated in batch mode. The experimental method used in this work has been described in more detail in similar experiments designed to characterize wall interactions in various types of tubing, Teflon bags, and other instrument inlets (Krechmer et al., 2017; Liu et al., 2019). Briefly, a series of 1-alkanol compounds (C₆, C₈, C₉, C₁₀, and C₁₂) were injected into the dark chamber

along with methyl nitrite and NO at room temperature. UV blacklights were turned on for 10 s to photolyze methyl nitrite, producing OH radicals through subsequent chemistry (Atkinson et al., 1981). Rapid oxidation of the 1-alkanol compounds until the OH radicals were depleted led to quasi-stable ppt-level concentrations of a range of oxidation products, including hydroxynitrates (HN), dihydroxynitrates (DHN), and dihydroxycarbonyls (DHC) as listed in Table S1. The volatilities of these compounds were estimated using the SIMPOL method (Pankow and Asher, 2008) as in Liu et al., (2019). Chamber air was sampled into the IMR through a 0.75" OD, approximately 8" long PTFE tube. The sample flow through this inlet and into the IMR was 2 slpm, and the ion flow into the IMR was 3 slpm, for a total flow of 5slpm at a constant ~70 Torr in the IMR.

4.2 IMR delay times vs. previous designs

The main goal of updating the IMR design as described in Sect. 3.1 was to reduce the measurement artifacts due interactions between analytes and IMR wall surfaces. As described in detail in Sect. 3.2 above, a reduction in wall-induced artifacts leads to improved spatial/temporal accuracy of the measurements, reduced impacts of possible surface chemistry artifacts, and more easily interpretable data. In this section, we describe the measurements used to quantify the improvement achieved in the new design.

In past experiments, wall interactions occurring in lengths of tubing or in IMRs have been quantified using the amount of time required for a signal to decay to 10% of the maximum total signal after wall steady state had been achieved and the signal source was removed (Neuman et al., 1999; Veres et al., 2008; Pagonis et al., 2017; Deming et al., 2019; Liu et al., 2019). When the ambient source of the compound is removed, the background-subtracted signal rapidly decays and all of the remaining signal is due to molecules evaporating or desorbing from the wall surfaces.

To systematically test delay times in the updated IMR design, we employed the recently developed method of sampling a range of HN, DHN, and DHC oxidation products spanning more than five orders of magnitude in volatility. Further details of the experimental setup can be found in related work (Krechmer et al., 2017; Liu et al., 2019). These compounds were allowed to equilibrate with the chamber walls, and sampling from the chamber then provided a constant source of these compounds. Chamber air was sampled through the co-axial IMR into the CIMS until IMR wall steady state was achieved for all compounds. At this point, UHP N₂ was injected into the variable orifice upstream of the IMR, removing the source of analyte and starting the measurement of delay times. While the chamber air was dry for all experiments, measurements were

performed with and without adding an estimated $1\text{--}2 \times 10^{16}$ molec cm^{-3} water vapor directly to the IMR. This way, the effects of water vapor on the IMR surfaces were probed.

The delay time measurement for one compound, a $\text{C}_9\text{H}_{19}\text{NO}_5$ DHN with an estimated C^* of $14.6 \mu\text{g m}^{-3}$ (which would typically be categorized as a semivolatile organic compound or SVOC), is shown in Fig. 3. Fast zero measurements (6 s every 1 min) of the background signal were taken prior to the start of the delay time measurement, illustrating that wall steady state was reached and that approximately 48% of the total signal was due to the background in the IMR. In other words, half of those analyte molecules that entered the IMR had interactions with a wall surface prior to desorbing and being sampled at the detector. Once the delay time measurement started, the signal due to molecules that did not interact with walls rapidly decayed (within several seconds) followed by the slower decay of the background signal. The amount of time required for the total signal to drop to within 10% of the persistent background level (which for this compound was essentially equal to the baseline noise) was measured to be 356 s, or 5.9 min. This DHN is an example of a compound that would require the fast zero method of background determination in order to achieve temporal/spatial resolution when sampling variable concentrations such as plumes. Delay times were also determined for the range of other compounds present in the chamber for both a dry and humidified IMR.

Liu et al. 2019 compiled delay times for the IMRs of several instruments, including a quadrupole proton transfer reaction MS (q-PTRMS; Pagonis et al., 2017), a Vocus inlet coupled with a time-of-flight MS (Krechmer et al., 2018), an I^- CIMS using the commercially available IMR (Aerodyne, Inc.) operated at dry conditions by the Jimenez group, and a custom design similar to the commercially available IMR operated under humidified conditions by the Ziemann group. The I^- CIMS instruments were tested using the same method and analytes as in this work, while the delay times for the q-PTRMS and Vocus instruments were measured using a similar method involving a series of ketones at equilibrium with the walls in a chamber (Pagonis et al., 2017; Deming et al., 2019). Figure 4 illustrates the delay times measured here in context with the previous results.

In general, the delay times for the co-axial IMR described herein were approximately an order of magnitude shorter than for the stainless steel IMR under dry conditions, and approximately 5 times shorter than the similar but humidified stainless steel IMR. The effects of humidity in an IMR appear to depend both on the material of the IMR as well as the type of analyte. In stainless steel IMRs, increased humidity led to uniformly shorter delay times for all analytes. However, in our new IMR, humidity led to slightly longer delay

527 times for DHN and no change for HN. These results illustrate how the interaction between an analyte and a
528 surface can be determined by a complex combination of factors, including the surface type, surface
529 modifications, and functional groups and properties of the analyte.

530 For both the dry and humidified stainless steel IMRs, results indicated that delay times started trending
531 back towards shorter values at the lowest measured C^* values. This trend is in contrast to the results from the
532 co-axial IMR. Liu et al. (2019) attributed this to irreversible loss of the analyte to the walls, which would
533 decrease the background signal relative to background-subtracted signal. It may be the case that this
534 irreversible loss for species of $C^* < 100 \mu\text{g m}^{-3}$ is unique to those stainless steel IMR surfaces and doesn't occur
535 on the PFA and FEP Teflon surfaces in the co-axial IMR. However, it may also be the case that those lowest
536 volatility compounds had not yet achieved wall steady state with the inlet tubes and stainless steel IMR walls.
537 This would have led to an artificially low amount of background signal relative to background-subtracted signal,
538 causing underestimates of delay times. Successively lower C^* compounds would be further away from steady
539 state for the sampling time prior to the start of the delay measurement, leading to successively more
540 underestimated delay times. If one assumes the linear relationship (in log-log space) observed in the co-axial
541 IMR and for $C^* > 100 \mu\text{g m}^{-3}$ in the stainless steel IMRs would hold for the lower C^* compounds, the delay times
542 in the stainless steel IMRs would reach on order of ~ 1000 minutes at most, which would become an implausible
543 amount of time to wait for wall steady state to be reached (and for all of the background to decay during the
544 delay measurement) during a batch mode chamber experiment. Also, it would be extremely difficult to
545 ascertain when wall steady state was achieved due to the slow rate of increase of the background signal.

546 At first glance, extrapolation of results would indicate that the Vocus and q-PTRMS instruments would
547 have one or several orders of magnitude longer respective delay times for the same HN, DHN, and DHC
548 compounds compared with our new IMR. The Vocus and q-PTRMS instruments are designed primarily for H_3O^+
549 ionization chemistry, typically to target a much more volatile set of analyte compounds compared with I^-
550 ionization. They also typically operate with an IMR pressure in the range of 2 Torr, which will greatly enhance
551 the rates of diffusion to the walls compared with the ~ 70 Torr I^- CIMS IMRs. Both our new IMR and the Vocus
552 have delay times spanning from a second to greater than several minutes over their respective volatility ranges
553 of interest. However, these results indicate that a Vocus-type design would not perform as well for I^- ionization
554 without modifications.

The IMR used in Lee et al. (2018), which employed the same variable orifice with an H₂O vapor addition port but with turbulent mixing of ions and analyte, was not tested on the CU chamber. However, laboratory experiments indicate that the delay time for HNO₃ under similar humidified conditions in the Lee et al. (2018) IMR was approximately a factor of three longer than in this new IMR (see Fig. S2), providing a measure of the improvements between that design and the one presented herein.

5 Conclusions

The effects of wall interactions in mass spectrometer inlets and IMRs have been a persistent but sometimes nebulous concern for as long as researchers have been sampling gases, particularly the lower volatility and soluble ones often referred to as “sticky” gases. As the importance of such gases to atmospheric processes like new particle formation/growth and SOA formation continues to be discovered, so does the need for higher precision and accuracy of the measurements. Recent research has begun to focus on analyte-surface interactions, including absorption and adsorption processes and how they can affect measurements in IMRs and in sample tubing. In this work, we introduced a new IMR design with the goal of reducing IMR wall interactions. This design was informed by the concepts in this and prior research. It sought to minimize wall interactions by limiting both turbulent and diffusive mixing to the walls, and by choosing wall surfaces that interact least with the analyte molecules. The new IMR was shown to have delay times that were 3–10 times shorter than previous IMR versions. This translates to higher signal-to-noise of the background-subtracted signal (i.e., the signal that did not interact with walls), less influence from possible surface reactions, and easier interpretation of measured time series.

Since there are a large number of factors affecting wall interactions, many of which are poorly understood, there has been little ability for researchers across different platforms to apply a uniform treatment to wall effects. Here, we aimed to provide a common framework of concepts with which the wall interactions in all instrumental systems could be described and treated. In this framework, the total signal measured at the detector for a given analyte can be described as originating from the sum of the following two pathways: 1) some fraction of the molecules do not interact with IMR wall surfaces, and are sampled with a time response equal to the average residence time in the IMR, and 2) the remaining fraction of molecules interact with the IMR walls via adsorption/absorption, and are sampled with a delayed time response longer than the average IMR residence time. We demonstrated a method of using fast zeroing to separate the signal into these parts,

583 namely the background-subtracted signal and the dynamic plus persistent background signals. The background-
584 subtracted signal is the only part that is a constant function of, and deterministic of, the concentration of
585 analyte entering the IMR as a function of time, and is thus an essential quantity for accurately capturing time-
586 dependence of analyte concentrations. This framework could be adapted to other inlet and instrument
587 configurations. A consistent manner of calibration was also presented.

588 This IMR design and the characterization of wall interactions represents an improvement over previous
589 low-pressure CIMS techniques used in atmospheric chemistry. Future work could build upon this design, for
590 instance by further decreasing wall interactions. One could also imagine a case where the walls are
591 modified/treated with a method similar to that in Roscioli et al. (2016), but in such a way as to make the walls
592 an irreversible sink for a particular analyte, thereby eliminating the background signal and making the total
593 signal equal to the background-subtracted signal. However, finding a modification technique that would work
594 for the entire range of diverse analyte molecules to which iodide-adduct ionization is sensitive could prove
595 challenging.

596 To facilitate comparisons and merging of data sets from different instruments, we also encourage the
597 users of all CIMS techniques to adopt the methods for calibration and background subtraction discussed herein
598 when sampling analytes that suffer from wall interactions, and encourage the reporting of all relevant sampling
599 and calibration method details in the publication of such research.

600 **Data Availability**

601 All data is available upon request to the authors.

602 **Author Contributions**

603 BBP and JAT designed, assembled, and tested the IMR, and determined the framework of wall interactions. BBP,
604 XL and JLJ conducted the characterization experiments at the CU Environmental Chamber facility and
605 participated in the analysis of the data. BBP wrote the manuscript. All authors contributed to revisions of the
606 manuscript.

607 **Competing Interests**

608 The authors declare that they have no conflict of interest.

609 **Acknowledgements**

610 This research was funded by grant AGS-1652688 from the U.S. National Science Foundation (NSF) and
611 by grant NA17OAR4310012 from the National Oceanic and Atmospheric Administration (NOAA). The authors
612 sincerely thank machinist Dennis Canuelle from the University of Washington for his contributions to the design
613 and manufacture of the IMR. XL and JLJ were supported by the Sloan Foundation Grant 2016-7173, and the US
614 DOE (BER/ASR) grant DE-SC0016559. We are grateful for many helpful discussions with Demetrios Pagonis,
615 Jordan E. Krechmer, and other members of the Aerodyne ToF-CIMS User's Group which helped to shape the
616 material presented herein. We thank Douglas A. Day for experimental support during the chamber
617 measurements.

618

619 **References**

- 620 Atkinson, R., Carter, W. P. L., Winer, A. M. and Pitts, J. N.: An Experimental Protocol For The Determination Of
 621 OH Radical Rate Constants With Organics Using Methyl Nitrite Photolysis As An OH Radical Source, *J. Air Pollut.*
 622 *Control Assoc.*, 31, 1090–1092, doi:10.1080/00022470.1981.10465331, 1981.
- 623 Bertram, T. H., Kimmel, J. R., Crisp, T. A., Ryder, O. S., Yatavelli, R. L. N., Thornton, J. A., Cubison, M. J., Gonin, M.
 624 and Worsnop, D. R.: A Field-Deployable, Chemical Ionization Time-Of-Flight Mass Spectrometer, *Atmos. Meas.*
 625 *Tech.*, 4, 1471–1479, doi:10.5194/amt-4-1471-2011, 2011.
- 626 Le Breton, M., McGillen, M. R., Muller, J. B. A., Bacak, A., Shallcross, D. E., Xiao, P., Huey, L. G., Tanner, D., Coe,
 627 H. and Percival, C. J.: Airborne Observations Of Formic Acid Using A Chemical Ionization Mass Spectrometer,
 628 *Atmos. Meas. Tech.*, 5, 3029–3039, doi:10.5194/amt-5-3029-2012, 2012.
- 629 Di Carlo, P., Brune, W. H., Martinez, M., Harder, H., Leshner, R., Ren, X. R., Thornberry, T., Carroll, M. A., Young,
 630 V., Shepson, P. B., Riemer, D., Apel, E. and Campbell, C.: Missing OH Reactivity In A Forest: Evidence For
 631 Unknown Reactive Biogenic VOCs, *Science*, 304, 722–725, doi:10.1126/science.1094392, 2004.
- 632 Çengel, Y. A. and Cimbala, J. M.: *Fluid Mechanics : Fundamentals And Applications*, 3rd ed., The McGraw-Hill
 633 Companies, Inc., New York, NY, 2014.
- 634 Crounse, J. D., McKinney, K. A., Kwan, A. J. and Wennberg, P. O.: Measurement Of Gas-Phase Hydroperoxides By
 635 Chemical Ionization Mass Spectrometry, *Anal. Chem.*, 78, 6726–6732, doi:10.1021/ac0604235, 2006.
- 636 Crutzen, P. J.: The Role Of NO And NO₂ In The Chemistry Of The Troposphere And Stratosphere, *Annu. Rev.*
 637 *Earth Planet. Sci.*, 7, 443–472, doi:10.1146/annurev.ea.07.050179.002303, 1979.
- 638 Deming, B. L., Pagonis, D., Liu, X., Day, D. A., Talukdar, R., Krechmer, J. E., de Gouw, J. A., Jimenez, J. L. and
 639 Ziemann, P. J.: Measurements Of Delays Of Gas-Phase Compounds In A Wide Variety Of Tubing Materials Due
 640 To Gas–Wall Interactions, *Atmos. Meas. Tech.*, 12, 3453–3461, doi:10.5194/amt-12-3453-2019, 2019.
- 641 Donahue, N. M., Kroll, J. H., Pandis, S. N. and Robinson, A. L.: A Two-Dimensional Volatility Basis Set – Part 2:
 642 Diagnostics Of Organic-Aerosol Evolution, *Atmos. Chem. Phys.*, 12, 615–634, doi:10.5194/acp-12-615-2012,
 643 2012.
- 644 Ehn, M., Thornton, J. A., Kleist, E., Sipilä, M., Junninen, H., Pullinen, I., Springer, M., Rubach, F., Tillmann, R., Lee,
 645 B., Lopez-Hilfiker, F., Andres, S., Acir, I.-H., Rissanen, M., Jokinen, T., Schobesberger, S., Kangasluoma, J.,
 646 Kontkanen, J., Nieminen, T., Kurtén, T., Nielsen, L. B., Jørgensen, S., Kjaergaard, H. G., Canagaratna, M., Maso,
 647 M. D., Berndt, T., Petäjä, T., Wahner, A., Kerminen, V.-M., Kulmala, M., Worsnop, D. R., Wildt, J. and Mentel, T.
 648 F.: A Large Source Of Low-Volatility Secondary Organic Aerosol, *Nature*, 506, 476–479,
 649 doi:10.1038/nature13032, 2014.
- 650 Eisele, F. L. and Tanner, D. J.: Measurement Of The Gas Phase Concentration Of H₂SO₄ And Methane Sulfonic
 651 Acid And Estimates Of H₂SO₄ Production And Loss In The Atmosphere, *J. Geophys. Res. Atmos.*, 98, 9001–9010,
 652 doi:10.1029/93JD00031, 1993.
- 653 Gaston, C. J., Lopez-Hilfiker, F. D., Whybrew, L. E., Hadley, O., McNair, F., Gao, H., Jaffe, D. A. and Thornton, J. A.:
 654 Online Molecular Characterization Of Fine Particulate Matter In Port Angeles, WA: Evidence For A Major Impact

655 From Residential Wood Smoke, *Atmos. Environ.*, 138, 99–107, doi:10.1016/j.atmosenv.2016.05.013, 2016.

656 Goldstein, A. H. and Galbally, I. E.: Known And Unexplored Organic Constituents In The Earth's Atmosphere,
 657 *Environ. Sci. Technol.*, 41, 1514–1521, doi:10.1021/es072476p, 2007.

658 Huang, Y., Zhao, R., Charan, S. M., Kenseth, C. M., Zhang, X. and Seinfeld, J. H.: Unified Theory Of Vapor–Wall
 659 Mass Transport In Teflon-Walled Environmental Chambers, *Environ. Sci. Technol.*, 52, 2134–2142,
 660 doi:10.1021/acs.est.7b05575, 2018.

661 Huey, L. G., Hanson, D. R. and Howard, C. J.: Reactions Of SF_6^- And I^- With Atmospheric Trace Gases, *J. Phys.*
 662 *Chem.*, 99, 5001–5008, doi:10.1021/j100014a021, 1995.

663 Hunter, J. F., Day, D. A., Palm, B. B., Yatavelli, R. L. N., Chan, A. W. H., Kaser, L., Cappellin, L., Hayes, P. L., Cross,
 664 E. S., Carrasquillo, A. J., Campuzano-Jost, P., Stark, H., Zhao, Y., Hohaus, T., Smith, J. N., Hansel, A., Karl, T.,
 665 Goldstein, A. H., Guenther, A., Worsnop, D. R., Thornton, J. A., Heald, C. L., Jimenez, J. L. and Kroll, J. H.:
 666 Comprehensive Characterization Of Atmospheric Organic Carbon At A Forested Site, *Nat. Geosci.*, 10, 748–753,
 667 doi:10.1038/ngeo3018, 2017.

668 Isaacman-VanWertz, G., Massoli, P., O'Brien, R., Lim, C., Franklin, J. P., Moss, J. A., Hunter, J. F., Nowak, J. B.,
 669 Canagaratna, M. R., Misztal, P. K., Arata, C., Roscioli, J. R., Herndon, S. T., Onasch, T. B., Lambe, A. T., Jayne, J. T.,
 670 Su, L., Knopf, D. A., Goldstein, A. H., Worsnop, D. R. and Kroll, J. H.: Chemical Evolution Of Atmospheric Organic
 671 Carbon Over Multiple Generations Of Oxidation, *Nat. Chem.*, 10, 462–468, doi:10.1038/s41557-018-0002-2,
 672 2018.

673 Isaacman, G., Kreisberg, N. M., Yee, L. D., Worton, D. R., Chan, A. W. H., Moss, J. A., Hering, S. V. and Goldstein,
 674 A. H.: Online Derivatization For Hourly Measurements Of Gas- And Particle-Phase Semi-Volatile Oxygenated
 675 Organic Compounds By Thermal Desorption Aerosol Gas Chromatography (SV-TAG), *Atmos. Meas. Tech.*, 7,
 676 4417–4429, doi:10.5194/amt-7-4417-2014, 2014.

677 Iyer, S., Lopez-Hilfiker, F., Lee, B. H., Thornton, J. A. and Kurtén, T.: Modeling The Detection Of Organic And
 678 Inorganic Compounds Using Iodide-Based Chemical Ionization, *J. Phys. Chem. A*, 120, 576–587,
 679 doi:10.1021/acs.jpca.5b09837, 2016.

680 Jokinen, T., Sipilä, M., Junninen, H., Ehn, M., Lönn, G., Hakala, J., Petäjä, T., Mauldin, R. L., Kulmala, M.,
 681 Worsnop, D. R., Mauldin III, R. L., Kulmala, M., Worsnop, D. R., Sipilä, M., Junninen, H., Ehn, M., Lönn, G., Hakala,
 682 J., Petaja, T., Mauldin, R. L., Kulmala, M., Worsnop, D. R., Sipilä, M., Junninen, H., Ehn, M., Lönn, G., Hakala, J.,
 683 Petäjä, T., Mauldin, R. L., Kulmala, M. and Worsnop, D. R.: Atmospheric Sulphuric Acid And Neutral Cluster
 684 Measurements Using CI-API-TOF, *Atmos. Chem. Phys.*, 12, 4117–4125, doi:10.5194/acp-12-4117-2012, 2012.

685 Kercher, J. P., Riedel, T. P. and Thornton, J. A.: Chlorine Activation By N_2O_5 : Simultaneous, In Situ Detection Of
 686 ClNO_2 And N_2O_5 By Chemical Ionization Mass Spectrometry, *Atmos. Meas. Tech.*, 2, 193–204, doi:10.5194/amt-
 687 2-193-2009, 2009.

688 Krechmer, J. E., Coggon, M. M., Massoli, P., Nguyen, T. B., Crounse, J. D., Hu, W., Day, D. A., Tyndall, G. S.,
 689 Henze, D. K., Rivera-Rios, J. C., Nowak, J. B., Kimmel, J. R., Mauldin, R. L., Stark, H., Jayne, J. T., Sipilä, M.,
 690 Junninen, H., Clair, J. M. St., Zhang, X., Feiner, P. A., Zhang, L., Miller, D. O., Brune, W. H., Keutsch, F. N.,
 691 Wennberg, P. O., Seinfeld, J. H., Worsnop, D. R., Jimenez, J. L. and Canagaratna, M. R.: Formation Of Low

692 Volatility Organic Compounds And Secondary Organic Aerosol From Isoprene Hydroxyhydroperoxide Low-NO
693 Oxidation, *Environ. Sci. Technol.*, 49, 10330–10339, doi:10.1021/acs.est.5b02031, 2015.

694 Krechmer, J. E., Groessl, M., Zhang, X., Junninen, H., Massoli, P., Lambe, A. T., Kimmel, J. R., Cubison, M. J., Graf,
695 S., Lin, Y.-H., Budisulistiorini, S. H., Zhang, H., Surratt, J. D., Knochenmuss, R., Jayne, J. T., Worsnop, D. R.,
696 Jimenez, J.-L. and Canagaratna, M. R.: Ion Mobility Spectrometry–Mass Spectrometry (IMS–MS) For On- And
697 Offline Analysis Of Atmospheric Gas And Aerosol Species, *Atmos. Meas. Tech.*, 9, 3245–3262, doi:10.5194/amt-
698 9-3245-2016, 2016.

699 Krechmer, J. E., Pagonis, D., Ziemann, P. J. and Jimenez, J. L.: Quantification Of Gas-Wall Partitioning In Teflon
700 Environmental Chambers Using Rapid Bursts Of Low-Volatility Oxidized Species Generated In Situ, *Environ. Sci.*
701 *Technol.*, 50, 5757–5765, doi:10.1021/acs.est.6b00606, 2016.

702 Krechmer, J. E., Day, D. A., Ziemann, P. J. and Jimenez, J. L.: Direct Measurements Of Gas/Particle Partitioning
703 And Mass Accommodation Coefficients In Environmental Chambers, *Environ. Sci. Technol.*, 51,
704 doi:10.1021/acs.est.7b02144, 2017.

705 Krechmer, J. E., Lopez-Hilfiker, F., Koss, A., Hutterli, M., Stoermer, C., Deming, B., Kimmel, J., Warneke, C.,
706 Holzinger, R., Jayne, J., Worsnop, D., Fuhrer, K., Gonin, M. and de Gouw, J.: Evaluation Of A New Reagent-Ion
707 Source And Focusing Ion–Molecule Reactor For Use In Proton-Transfer-Reaction Mass Spectrometry, *Anal.*
708 *Chem.*, 90, 12011–12018, doi:10.1021/acs.analchem.8b02641, 2018.

709 Lee, B. H., Lopez-Hilfiker, F. D., Mohr, C., Kurtén, T., Worsnop, D. R. and Thornton, J. A.: An Iodide-Adduct High-
710 Resolution Time-Of-Flight Chemical-Ionization Mass Spectrometer: Application To Atmospheric Inorganic And
711 Organic Compounds, *Environ. Sci. Technol.*, 48, 6309–6317, doi:10.1021/es500362a, 2014.

712 Lee, B. H., Mohr, C., Lopez-Hilfiker, F. D., Lutz, A., Hallquist, M., Lee, L., Romer, P., Cohen, R. C., Iyer, S., Kurtén,
713 T., Hu, W., Day, D. A., Campuzano-Jost, P., Jimenez, J. L., Xu, L., Ng, N. L., Guo, H., Weber, R. J., Wild, R. J.,
714 Brown, S. S., Koss, A., de Gouw, J., Olson, K., Goldstein, A. H., Seco, R., Kim, S., McAvey, K., Shepson, P. B., Starn,
715 T., Baumann, K., Edgerton, E. S., Liu, J., Shilling, J. E., Miller, D. O., Brune, W., Schobesberger, S., D’Ambro, E. L.
716 and Thornton, J. A.: Highly Functionalized Organic Nitrates In The Southeast United States: Contribution To
717 Secondary Organic Aerosol And Reactive Nitrogen Budgets, *Proc. Natl. Acad. Sci.*, 113, 1516–1521,
718 doi:10.1073/pnas.1508108113, 2016.

719 Lee, B. H., Lopez-Hilfiker, F. D., Veres, P. R., McDuffie, E. E., Fibiger, D. L., Sparks, T. L., Ebben, C. J., Green, J. R.,
720 Schroder, J. C., Campuzano-Jost, P., Iyer, S., D’Ambro, E. L., Schobesberger, S., Brown, S. S., Wooldridge, P. J.,
721 Cohen, R. C., Fiddler, M. N., Bililign, S., Jimenez, J. L., Kurtén, T., Weinheimer, A. J., Jaegle, L. and Thornton, J. A.:
722 Flight Deployment Of A High-Resolution Time-Of-Flight Chemical Ionization Mass Spectrometer: Observations
723 Of Reactive Halogen And Nitrogen Oxide Species, *J. Geophys. Res. Atmos.*, doi:10.1029/2017JD028082, 2018.

724 Liu, X., Deming, B., Pagonis, D., Day, D. A., Palm, B. B., Talukdar, R., Roberts, J. M., Veres, P. R., Krechmer, J. E.,
725 Thornton, J. A., de Gouw, J. A., Ziemann, P. J. and Jimenez, J. L.: Effects Of Gas–Wall Interactions On
726 Measurements Of Semivolatile Compounds And Small Polar Molecules, *Atmos. Meas. Tech.*, 12, 3137–3149,
727 doi:10.5194/amt-12-3137-2019, 2019.

728 Lopez-Hilfiker, F. D., Mohr, C., Ehn, M., Rubach, F., Kleist, E., Wildt, J., Mentel, T. F., Lutz, A., Hallquist, M.,
729 Worsnop, D. and Thornton, J. A.: A Novel Method For On-Line Analysis Of Gas And Particle Composition:

730 Description And Evaluation Of A Filter Inlet For Gases And AEROSols (FIGAERO), *Atmos. Meas. Tech.*, 6, 9347–
731 9395, doi:10.5194/amt-7-983-2014, 2013.

732 Lopez-Hilfiker, F. D., Iyer, S., Mohr, C., Lee, B. H., D'Ambro, E. L., Kurtén, T. and Thornton, J. A.: Constraining The
733 Sensitivity Of Iodide Adduct Chemical Ionization Mass Spectrometry To Multifunctional Organic Molecules Using
734 The Collision Limit And Thermodynamic Stability Of Iodide Ion Adducts, *Atmos. Meas. Tech.*, 9, 1505–1512,
735 doi:10.5194/amt-9-1505-2016, 2016.

736 Massoli, P., Stark, H., Canagaratna, M. R., Krechmer, J. E., Xu, L., Ng, N. L., Mauldin, R. L., Yan, C., Kimmel, J.,
737 Misztal, P. K., Jimenez, J. L., Jayne, J. T. and Worsnop, D. R.: Ambient Measurements Of Highly Oxidized Gas-
738 Phase Molecules During The Southern Oxidant And Aerosol Study (SOAS) 2013, *ACS Earth Sp. Chem.*, 2, 653–
739 672, doi:10.1021/acsearthspacechem.8b00028, 2018.

740 Matsunaga, A. and Ziemann, P. J.: Gas-Wall Partitioning Of Organic Compounds In A Teflon Film Chamber And
741 Potential Effects On Reaction Product And Aerosol Yield Measurements, *Aerosol Sci. Technol.*, 44, 881–892,
742 doi:10.1080/02786826.2010.501044, 2010.

743 Mohr, C., Lopez-Hilfiker, F. D., Zotter, P., Prévôt, A. S. H., Xu, L., Ng, N. L., Herndon, S. C., Williams, L. R., Franklin,
744 J. P., Zahniser, M. S., Worsnop, D. R., Knighton, W. B., Aiken, A. C., Gorkowski, K. J., Dubey, M. K., Allan, J. D. and
745 Thornton, J. A.: Contribution Of Nitrated Phenols To Wood Burning Brown Carbon Light Absorption In Detling,
746 United Kingdom During Winter Time, *Environ. Sci. Technol.*, 47, 6316–6324, doi:10.1021/es400683v, 2013.

747 Neuman, J. A., Huey, L. G., Ryerson, T. B. and Fahey, D. W.: Study Of Inlet Materials For Sampling Atmospheric
748 Nitric Acid, *Environ. Sci. Technol.*, 33, 1133–1136, doi:10.1021/es980767f, 1999.

749 Neuman, J. A., Huey, L. G., Dissly, R. W., Fehsenfeld, F. C., Flocke, F., Holecek, J. C., Holloway, J. S., Hübler, G.,
750 Jakoubek, R., Nicks, D. K., Parrish, D. D., Ryerson, T. B., Sueper, D. T. and Weinheimer, A. J.: Fast-Response
751 Airborne In Situ Measurements Of HNO_3 During The Texas 2000 Air Quality Study, *J. Geophys. Res.*, 107, 4436,
752 doi:10.1029/2001JD001437, 2002.

753 Pagonis, D., Krechmer, J. E., de Gouw, J., Jimenez, J. L. and Ziemann, P. J.: Effects Of Gas–Wall Partitioning In
754 Teflon Tubing And Instrumentation On Time-Resolved Measurements Of Gas-Phase Organic Compounds,
755 *Atmos. Meas. Tech.*, 10, 4687–4696, doi:10.5194/amt-10-4687-2017, 2017.

756 Pankow, J. F. and Asher, W. E.: SIMPOL.1: A Simple Group Contribution Method For Predicting Vapor Pressures
757 And Enthalpies Of Vaporization Of Multifunctional Organic Compounds, *Atmos. Chem. Phys.*, 8, 2773–2796,
758 doi:10.5194/acp-8-2773-2008, 2008.

759 Peng, J., Hu, M., Guo, S., Du, Z., Zheng, J., Shang, D., Levy Zamora, M., Zeng, L., Shao, M., Wu, Y.-S., Zheng, J.,
760 Wang, Y., Glen, C. R., Collins, D. R., Molina, M. J. and Zhang, R.: Markedly Enhanced Absorption And Direct
761 Radiative Forcing Of Black Carbon Under Polluted Urban Environments, *Proc. Natl. Acad. Sci.*, 113, 4266–4271,
762 doi:10.1073/pnas.1602310113, 2016.

763 Roscioli, J. R., Zahniser, M. S., Nelson, D. D., Herndon, S. C. and Kolb, C. E.: New Approaches To Measuring Sticky
764 Molecules: Improvement Of Instrumental Response Times Using Active Passivation, *J. Phys. Chem. A*, 120,
765 1347–1357, doi:10.1021/acs.jpca.5b04395, 2016.

766 Seinfeld, J. H. and Pandis, S. N.: *Atmospheric Chemistry And Physics: From Air Pollution To Climate Change*, 2nd

ed., John Wiley & Sons, Inc., Hoboken, New Jersey, USA, 2006.

Sekimoto, K., Li, S.-M., Yuan, B., Koss, A., Coggon, M., Warneke, C. and de Gouw, J.: Calculation Of The Sensitivity Of Proton-Transfer-Reaction Mass Spectrometry (PTR-MS) For Organic Trace Gases Using Molecular Properties, *Int. J. Mass Spectrom.*, 421, 71–94, doi:10.1016/j.ijms.2017.04.006, 2017.

Shrivastava, M., Cappa, C. D., Fan, J., Goldstein, A. H., Guenther, A. B., Jimenez, J. L., Kuang, C., Laskin, A., Martin, S. T., Ng, N. L., Petaja, T., Pierce, J. R., Rasch, P. J., Roldin, P., Seinfeld, J. H., Shilling, J., Smith, J. N., Thornton, J. A., Volkamer, R., Wang, J., Worsnop, D. R., Zaveri, R. A., Zelenyuk, A. and Zhang, Q.: Recent Advances In Understanding Secondary Organic Aerosol: Implications For Global Climate Forcing, *Rev. Geophys.*, 55, 509–559, doi:10.1002/2016RG000540, 2017.

Sparrow, E. M., Abraham, J. P. and Minkowycz, W. J.: Flow Separation In A Diverging Conical Duct: Effect Of Reynolds Number And Divergence Angle, *Int. J. Heat Mass Transf.*, 52, 3079–3083, doi:10.1016/j.ijheatmasstransfer.2009.02.010, 2009.

Veres, P., Roberts, J. M., Warneke, C., Welsh-Bon, D., Zahniser, M., Herndon, S., Fall, R. and de Gouw, J.: Development Of Negative-Ion Proton-Transfer Chemical-Ionization Mass Spectrometry (NI-PT-CIMS) For The Measurement Of Gas-Phase Organic Acids In The Atmosphere, *Int. J. Mass Spectrom.*, 274, 48–55, doi:10.1016/j.ijms.2008.04.032, 2008.

Veres, P. R., Roberts, J. M., Wild, R. J., Edwards, P. M., Brown, S. S., Bates, T. S., Quinn, P. K., Johnson, J. E., Zamora, R. J. and de Gouw, J.: Peroxynitric Acid (HO_2NO_2) Measurements During The UBWOS 2013 And 2014 Studies Using Iodide Ion Chemical Ionization Mass Spectrometry, *Atmos. Chem. Phys.*, 15, 8101–8114, doi:10.5194/acp-15-8101-2015, 2015.

Yatavelli, R. L. N., Stark, H., Thompson, S. L., Kimmel, J. R., Cubison, M. J., Day, D. A., Campuzano-Jost, P., Palm, B. B., Hodzic, A., Thornton, J. A., Jayne, J. T., Worsnop, D. R. and Jimenez, J. L.: Semicontinuous Measurements Of Gas–Particle Partitioning Of Organic Acids In A Ponderosa Pine Forest Using A MOVI-HRToF-CIMS, *Atmos. Chem. Phys.*, 14, 1527–1546, doi:10.5194/acp-14-1527-2014, 2014.

Yuan, B., Koss, A. R., Warneke, C., Coggon, M., Sekimoto, K. and de Gouw, J. A.: Proton-Transfer-Reaction Mass Spectrometry: Applications In Atmospheric Sciences, *Chem. Rev.*, 117, 13187–13229, doi:10.1021/acs.chemrev.7b00325, 2017.


Zaytsev, A., Breitenlechner, M., Koss, A. R., Lim, C. Y., Rowe, J. C., Kroll, J. H. and Keutsch, F. N.: Using Collision-Induced Dissociation To Constrain Sensitivity Of Ammonia Chemical Ionization Mass Spectrometry (NH_4^+ CIMS) To Oxygenated Volatile Organic Compounds, *Atmos. Meas. Tech.*, 12, 1861–1870, doi:10.5194/amt-12-1861-2019, 2019.

Zhao, J., Eisele, F. L., Titcombe, M., Kuang, C. and McMurry, P. H.: Chemical Ionization Mass Spectrometric Measurements Of Atmospheric Neutral Clusters Using The Cluster-CIMS, *J. Geophys. Res.*, 115, D08205, doi:10.1029/2009JD012606, 2010.

Zhao, Y., Chan, J. K., Lopez-Hilfiker, F. D., McKeown, M. A., D’Ambro, E. L., Slowik, J. G., Riffell, J. A. and Thornton, J. A.: An Electrospray Chemical Ionization Source For Real-Time Measurement Of Atmospheric Organic And Inorganic Vapors, *Atmos. Meas. Tech.*, 10, 3609–3625, doi:10.5194/amt-10-3609-2017, 2017.

804

805



807

808 **Figure 1.** Schematic of the new, co-axial, low-pressure IMR design for CIMS. This is a two-dimensional cross
809 section of the cylindrical IMR along the axis of flow, and it is not to exact scale. Black lines represent stainless
810 steel surfaces, green and blue lines represent PTFE Teflon, and red/yellow lines represent FEP or PFA Teflon.
811 Constant mass flow into the IMR is controlled using a variable orifice. Water vapor can be added through a port
812 in the orifice plate, in order to keep the environmental conditions in the IMR more constant. The sample flow
813 and ion flow are passed through laminizer elements to limit the effects of turbulent diffusion to the IMR walls.
814 Ion-molecule adducts are formed via diffusive mixing in the drift region. The ability to enhance the mixing of
815 ions into the sample flow by applying an electric field between the drift region wall and the exit of the sample
816 flow laminizer was also included (not shown), but led to only modest enhancement and was not used in the
817 measurements presented herein. A mirrored pumping scheme also prevents turbulence and limits the effects of
818 wall interactions. Adducts are sampled through a capillary into the time-of-flight mass spectrometer.

819





Figure 2. a) Schematic illustrating how wall interactions affect the measurement of low volatility or polar gases for several experimental conditions, and b) example of the fast zero method of background subtraction for the measurement of constant concentration of ~ 2 ppbv nitric acid from a permeation tube. The times corresponding to each panel in a) are labeled on the time series in panel b). The bottom of panel b) illustrates the benefits of performing frequent background signal subtractions as opposed to only subtracting the persistent background signal.




Figure 3. Delay time measurement of a DHN ($C_9H_{19}NO_5$) in an I⁻ CIMS with the new IMR. Prior to the start of the delay measurement, wall steady state had been achieved. The total signal is equal to the background-subtracted signal plus the background signal. Regular background measurements were performed for 6 s of each 1 min, illustrating that approximately half of the $C_9H_{19}NO_5$ that entered the IMR was interacting with the walls prior to desorbing and being sampled. The delay time for this DHN, defined as the time required for the signal to return to 10% of the original value, was determined to be 356 s.




Figure 4. Delay times for a variety of organic molecules as a function of saturation vapor concentration (C^* , $\mu\text{g m}^{-3}$), compared with previous IMR designs including a q-PTRMS, Vocus PTR-TOF-MS, and several I⁻ CIMS instruments with different IMRs. The delay time in a nitrate (NO_3^-) CIMS is also shown for comparison. The organic molecules are described in Table S1.