

Integrating wearable data into circadian models

Kevin M. Hannay¹ and Jennette P. Moreno²

Abstract

The emergence of wearable health sensors in the last decade has the potential to revolutionize the study of sleep and circadian rhythms. In particular, recent progress has been made in the use of mathematical models in the prediction of a patient's internal circadian state using data measured by wearable devices. This is a vital step in our ability to identify optimal circadian timing for health interventions. We review the available data for fitting circadian phase models with a focus on wearable data sets. Finally, we review the current modeling paradigms and explore avenues for developing personalized parameter sets in limit cycle oscillator models to further improve prediction accuracy.

Addresses

¹ Department of Mathematics, University of Michigan, Ann Arbor, MI, 48109, USA

² USDA/ARS Childrens Nutrition Research Center, Department of Pediatrics, Baylor College of Medicine, Houston, TX, 77030, USA

Corresponding author: Hannay, Kevin M (khannay@umich.edu)

Current Opinion in Systems Biology 2020, 22:32–38

This review comes from a themed issue on **Mathematical Modelling (2020)**

Edited by **Daniel Forger** and **Olivia Walch**

For complete overview of the section, please refer the article collection - [Mathematical Modelling \(2020\)](#)

Available online 21 August 2020

<https://doi.org/10.1016/j.coisb.2020.08.001>

2452-3100/© 2020 Elsevier Ltd. All rights reserved.

Keywords

Circadian rhythms, Mathematical models, Biological oscillators.

Introduction

Chronic sleep deprivation and circadian disorders have reached critical levels around the world by recent estimates. For instance, a large chronotype survey found that approximately 69% of the respondents reported social jet lag of 1 h or more [1]. Circadian disruption and misalignment with the external environment has been linked to a host of negative mental and physical health outcomes [2,3].

In mammals, the core circadian clock has been localized to a region of the hypothalamus known as the supra-chiasmatic nucleus (SCN). To remain entrained to the external environment, the SCN must receive temporal

information from external forces known as zeitgebers (time-givers). In mammals, the most important zeitgeber is the daily light cycle. The master clock, in the SCN, then coordinates a population of peripheral clocks producing physiological rhythms throughout the body.

The sleep and circadian research communities have made tremendous progress in elucidating the core mechanisms driving and connecting sleep and circadian rhythms. Translation of these scientific advances into improvements in circadian and sleep health of the populace has proved difficult [4]. For example, it has been found that the efficacy of a number of treatments vary in a circadian manner [5,6]. Application of this knowledge will require the ability to accurately predict the circadian phase of patients [7,8].

Circadian phase prediction is an inherently difficult problem. Circadian phase in humans can only be measured indirectly through the use of a peripheral marker rhythm driven by the core clock [9]. However, each of these peripheral rhythms has the potential to be obscured (masked) by behavioral rhythms such as the sleep–wake cycle [10]. This challenge has been met by the development of a series of laboratory protocols designed to isolate the circadian signatures from marker melatonin/core body temperature rhythms [10]. Although these protocols have been essential to determining many core circadian clock properties, the labor, expense, and significant patient burden make them impractical for use on large populations or children [4,11].

Personalizing circadian phase predictions will require novel data collection and mathematical modeling approaches. The advent and mass popularization of wearable technologies has provided an avalanche of data. However, this infusion of data will require careful integration into the accumulated body of knowledge using novel mathematical modeling techniques. Here, we will review the various data collection methods available and recent advances in building mathematical models.

Data overview

Phase assessment gold standards

The constant routine protocol is the current gold standard for human circadian phase estimation [10,12]. This protocol calls for participants to be held in a constant light environment free from time cues while a circadian marker is monitored. This constant environment

prevents the core circadian rhythms from being masked by exogenous periodic cycles and the sleep–wake cycle. Early circadian studies used the core body temperature as the circadian marker, although in recent years this has been replaced primarily by melatonin and other less invasive markers [13,14]. Melatonin rhythms may be collected from blood, saliva and urine samples, although using melatonin comes with the additional complication that even low levels of light suppress melatonin rhythms [14,15,12]. Thus, melatonin samples must be collected under dim light conditions of less than 30 lux—although many researchers strive for less than 5 lux [16]. Usually, samples are collected every 30–60 min and are used to estimate the onset of melatonin secretion which typically occurs just before bedtime. This time point is called the dim light melatonin onset (DLMO). Recently, positive progress has been made on an effort to move this protocol outside the laboratory by using self-collected melatonin samples from saliva, although compliance with this protocol could prove difficult in children and other groups [17].

These circadian phase assessment protocols may be applied to measure the effect of a light stimulus on the circadian phase. Plotting the timing of the stimulus (e.g. bright light) against the shift in circadian phase provides a phase response curve (PRC). PRCs are often used for treatment planning in clinical settings to determine the optimal timing to shift the circadian phase in the appropriate direction [18]. Several phase response curves to light have been found using a variety of light stimuli [13,19,20]. Moreover, several valuable data sets for the effect of light schedules of different duration's and intensities on the circadian phase have also been collected [21–23]. To date, PRC data have been the primary data used to fit and validate circadian prediction models [24,25].

Wearable data

A huge variety of wearable health and fitness trackers have entered the market place in the last decade. Generally, commercial versions of these devices will provide a time series of accelerometer and heart rate estimates based on photoplethysmography [26]. Wearable devices offer an objective alternative to self-reported data to measure circadian patterns of various clock outputs such as heart rate, sleep/wake patterns, activity, and wrist body temperature. Moreover, some of these devices have the ability to collect data on exposure to zeitgebers (e.g. ambient light exposure), although these are generally not found in commercial devices [4,27,2,28,26,29].

From a circadian phase estimation prospective, each of these sensor measurements should add additional information and improve circadian phase predictions. However, incorporation of these time series

measurements has proved challenging. These variables are typically highly correlated with one another, and any circadian signatures within the time series can potentially be masked by the sleep–wake cycle or other biological rhythms [10]. Moreover, with the exception of the light intensity variable, each of these physiological processes is both an input and output of the master circadian clock. Each of these factors must be accounted for to build an accurate circadian phase prediction system.

The large size of wearable data sets has the potential to override some of these difficulties. The incorporation of commercial devices will likely be required to make maximal use of this big data potential for circadian phase predictions. The diversity of commercial devices on the market has proved a challenge to their validation. Moreover, the algorithms used in processing the raw sensor data are typically closed source and updated more quickly; then, they could be validated for research purposes [26]. However, recent progress has been made in extracting the raw accelerometer and heart rate data from a subset of these devices [26]. Development of open-source methods for the extraction and processing of raw sensor data from commercial devices is a crucial step toward incorporating these data into circadian phase estimates [26,4].

The relationship between the physiological variables measured by wearable devices and the gold standard for circadian phase estimation, DLMO timing, is still somewhat unclear. However, recent studies have found significant correlations between many of these time series and DLMO measurements [30,31]. In particular, measurements of the body temperature at the wrist and mid sleep time have been found to be correlated with the DLMO and have been applied to estimate circadian phase [32,30,31,33,34]. It remains to be seen how these correlations can be used to refine circadian phase estimates. In the best case scenario, these relationships in combination with a lengthy time series could eliminate the need for DLMO estimation all together. However, the majority of the correlation results above were obtained under regular entrainment conditions for healthy adults. Before generalizing these results to larger populations, additional validation will be required across larger demographic ranges and environmental conditions.

Predicting phase from wearables

The lengthy time series data collected by wearable devices have allowed for machine learning techniques to be used in sleep and circadian prediction [35,26,36,30,37]. For example, artificial neural network (ANN) approaches have provided some of the most accurate predictions of circadian phase [35,36]. The ANN approach is appealing because of the success of

these techniques in predicting a variety of complex time series data [38]. Furthermore, these techniques do not require a mechanistic model and are therefore well-suited to deal with the modeling of complex phenomena. However, the predictions of ANN models have been found to perform poorly when applied to the more challenging problem of circadian phase prediction for shift workers [35].

Limit cycle oscillator models

Mathematically, the circadian clock is an example of an attracting limit cycle oscillator (LCO), meaning its solutions contain a stable closed trajectory comprising a self-sustained oscillator (Figure 1). The most widespread LCO models for human circadian rhythms are based on progressive modifications of a Van der Pol nonlinear oscillator model [39,25,40,41,18]. Recently, a class of models based on a systematic reduction of phase oscillators was proposed and fit to human phase response curve data [24,42]. These models are based on the network level physiology of the core clock, and therefore, the parameters are endowed with physiological interpretations [24]. By comparison, the lack of a physiological foundation for the Van der Pol-based


models makes parameter and variable interpretations more problematic. Each of these LCO models can take the light levels recorded by a wearable device and generate a circadian phase prediction as seen in (Figure 2).

A LCO model was recently applied to the prediction of circadian phase using wearable data for medical shift workers and compared against direct measurements of circadian phase [8]. These predictions were found to be accurate to within ± 1 hour for 70–80% of subjects in the study. This is comparable with the accuracy of predictions made for diurnal subjects in previous studies [43,44]. Being able to more accurately predict the circadian phase of shift workers from wearable data will help to facilitate the delivery of light therapy in the treatment of shift-work disorder.

Parameters for limit cycle oscillator models

LCO models have parameters (θ) which must be estimated using experimental phase response data (D_k). Optimization over a least squares cost function leads to a maximum likelihood estimation of the best-fit ($\hat{\theta}^*$) parameter set.

Figure 1



This figure shows an example simulation for a LCO model. The individual is started from a low amplitude rhythm (near 0) and then is entrained to a regularly repeated light schedule with a period of 24 h. The yellow color shows light exposure times, and the dark shade shows periods of darkness on all plots. On the left, the phase plane plot shows a parametric plot of the phase and amplitude variables growing toward a limit cycle solution (closed trajectory). Two circadian phase markers used in human circadian rhythms are shown (CBT_{min} is the minimum of the core body temperature rhythm, and the DLMO is the dim light melatonin onset). The right plots show the amplitude (top) and phase plotted against the time variable [24].

$$\vec{\theta}^* = \arg \min \frac{1}{2} \sum_{k=1}^N \left(\frac{D_k - M_k(\vec{\theta})}{\sigma_k} \right)^2 \quad (1)$$

These best-fit parameter sets have been applied to generate circadian phase predictions from the model [8,44,43]. This use of a single set of best-fit parameters ignores the inherent experimental, parameter and model uncertainty [46]. Significantly, parameter uncertainty results from the variation in the circadian parameters among the individuals used to construct the experimental curve. These uncertainties may be partially accounted for by adopting a Bayesian approach and generating an ensemble of parameters which are consistent with the experimental data rather than a single best-fit set [24,47]. From a clinical perspective the generation of statistical parameter ensembles could assist in the classification of circadian and sleep disorders. Whether these states are identifiable as distinct regions in parameter space remains to be seen, however the development of low-dimensional models whose


parameters can be tied to clock physiology should help in this regard [24].

Parameter estimation for circadian models

The intrinsic period (τ) parameter is influential in predictions, is variable across the human population, and can be measured directly using a forced desynchrony protocol [48–51]. Forced desynchrony decouples the sleep and circadian cycles, allowing the unmasked circadian cycle to be measured for long enough (10–20 days) to accurately assess the period [52,53].

Surveys can also be used to study circadian heterogeneity in larger samples than laboratory approaches, and the results of chronotype surveys have been found to be related to key circadian parameters [54,55,11,56]. Beyond the circadian period parameter, other parameters in circadian models have also been found to vary [57]. A striking example of this is provided by a recent study, which found that evening light intensities varying from dim reading light (10 lux) to bright indoor light (400 lux) caused equivalent suppression of melatonin rhythms in healthy young adults [57].

Figure 2



Wrist light level measurements recorded by an Actiwatch Spectrum wearable device [45]. The light recordings span one week but are repeated four times for illustration. The yellow blocks indicate light levels over 10 lux, and black areas indicate light levels below 10 lux. Red crosses show model predictions of the core body temperature (CBT) minimum circadian marker, and blue circles show model predicted DLMO timing [24].

The large data sets produced by wearable devices could be used to estimate key model parameters. A methodological challenge is how best to integrate the wearable device data alongside the laboratory and survey data. Some promising results have been found using a particle filter approach to integrate across these heterogeneous data sources [58]. Particle filters provide optimal state estimation by integrating a dynamical model with noisy observations [58]. As wearable data sets become more widely available, additional mathematical and computational tools will need to be developed to optimally perform this integration.

Conclusion

Here, we have reviewed the various data sources available for circadian phase prediction of which large-scale wearable data is an important component. In addition, we reviewed recent advances in mathematical and computational models for integrating wearable data into phase predictions.

Circadian phase prediction is a difficult problem, and a solution will require advancements in a number of fields. The first challenge will be to develop inexpensive methods for circadian phase measurement/validation which are suitable to be deployed on a massive scale outside the laboratory. Secondly, the challenge of incorporating and processing the heterogeneous and noisy measurements collected by wearable devices must be faced. Finally, new models and paradigms will need to be developed which can integrate laboratory, survey, and wearable data into predictions. Each of these data sources provide a complementary picture, and optimal circadian phase prediction will require that each of these are integrated into the predictive model.

Conflict of interest statement

Nothing declared.

Acknowledgements

KMH is grateful for support for this work through NSF DMS 1853506. JPM was supported by the Eunice Kennedy Shriver National Institute of Child Health and Human Development of the National Institutes of Health under award number R00HD091396 and the Department of Agriculture (USDA/ARS) Children's Nutrition Research Center, Department of Pediatrics, Baylor College of Medicine, Houston, TX, funded in part with federal funds from the USDA/ARS under Cooperative Agreement No. 58-3092-5-001.

References

- Roenneberg T, Allebrandt KV, Mellow M, Vetter C: **Social jetlag and obesity**. *Curr Biol* 2012, **22**:939–943.
 - Moreno JP, Crowley SJ, Alfano CA, Hannay KM, Thompson D, Baranowski T: **Potential circadian and circannual rhythm contributions to the obesity epidemic in elementary school age children**. *Int J Behav Nutr Phys Activ* 2019, **16**:1–10.
 - Loehfelm A, Boucsein A, Pretz D, Tups A: **Timing matters: circadian effects on energy homeostasis and alzheimer's disease**. *Trends Endocrinol Metabol* 2019, **30**:132–143, <https://doi.org/10.1016/j.tem.2018.12.001>.
 - Depner CM, Cheng PC, Devine JK, Khosla S, de Zambotti M, Robillard R, Vakulin A, Drummond SPA: **Wearable technologies for developing sleep and circadian biomarkers: a summary of workshop discussions**. *Sleep* 2019. URL <https://academic.oup.com/sleep/advance-article-abstract/doi/10.1093/sleep/zsz254/5602700>, <https://academic.oup.com/sleep/advance-article/doi/10.1093/sleep/zsz254/5602700>.
 - Hill RJ, Innominato PF, Lévi F, Ballesta A: **Optimizing circadian drug infusion schedules towards personalized cancer chemotherapy**. *PLoS Comput Biol* 2020, **16**, e1007218.
 - Rogers VE, Zhu S, Mandrell BN, Ancoli-Israel S, Liu L, Hinds PS: **Relationship between circadian activity rhythms and fatigue in hospitalized children with CNS cancers receiving high-dose chemotherapy**. *Support Care Canc* 2020, **28**:1459–1467.
 - Gaspar LS, Álvaro AR, Carmo-Silva S, Mendes AF, Relógio A, Cavadas C: **The importance of determining circadian parameters in pharmacological studies**. *Br J Pharmacol* 2019, **176**:2827–2847.
 - Stone JE, Aubert XL, Maass H, Phillips AJ, Magee M, Howard ME, Lockley SW, Rajaratnam SM, Sletten TL: **Application of a limit-cycle oscillator model for prediction of circadian phase in rotating night shift workers**. *Sci Rep* 2019, **9**:1–12.
- Circadian phase prediction using wearable data and a limit cycle oscillator. The model is applied to predict circadian phase for shift workers moving from a diurnal schedule to night shifts.
- Wirz-Justice A: **How to measure circadian rhythms in humans**. *Medicographia* 2007, **29**:84–90.
 - Duffy JF, Dijk DJ: **Getting through to circadian oscillators: why use constant routines?** *J Biol Rhythm* 2002, **17**:4–13.
 - Roenneberg T, Pilz LK, Zerbini G, Winnebeck EC: **Chronotype and social jetlag: a (self-) critical review**. *Biology* 2019, **8**:1–19.
 - Keijzer H, Smits MG, Duffy JF, Curfs LM: **Why the dim light melatonin onset (DLMO) should be measured before treatment of patients with circadian rhythm sleep disorders**. 2014.
 - Czeisler CA, Kronauer RE, Allan JS, Duffy JF, Jewett ME, Brown EN, Ronda JM: **Bright light induction of strong (type 0) resetting of the human circadian pacemaker**. *Science* 1989, **244**:1328–1333 (New York, N.Y.).
 - Broussard JL, Reynolds AC, Depner CM, Ferguson SA, Dawson D, Wright KP: **Circadian rhythms versus daily patterns in human physiology and behavior**. In *Biological time-keeping: clocks, rhythms and behaviour*. Edited by Kumar V, New Delhi: Springer India; 2017:279–295, https://doi.org/10.1007/978-81-322-3688-7_13.
 - Klerman EB, Gershengorn HB, Duffy JF, Kronauer RE: **Comparisons of the variability of three markers of the human circadian pacemaker**. *J Biol Rhythm* 2002, **17**:181–193.
 - Crowley SJ: **Assessment of circadian rhythms**. In *The Oxford handbook of infant, child, and adolescent sleep and behavior*, Oxford library of psychology. New York, NY, US: Oxford University Press; 2013:204–222.
 - Burgess HJ, Wyatt JK, Park M, Fogg LF: **Home circadian phase Assessments with measures of compliance yield accurate dim light melatonin onsets**. *Sleep* 2015.
 - Serkh K, Forger DB: **Optimal schedules of light exposure for rapidly correcting circadian misalignment**. *PLoS Comput Biol* 2014, **10**.
 - Khalsa SBS, Jewett ME, Cajochen C, Czeisler Ca: **A phase response curve to single bright light pulses in human subjects**. *J Physiol* 2003, **549**:945–952. <http://eutils.ncbi.nlm.nih.gov/entrez/eutils/elink.fcgi?dbfrom=pubmed%7B%7D&Dretmode=ref%7B%7D&Dcmd=prlinks%5Cnhttp://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=2342968&tool=pmcentrez&rendertype=abstract%5Cnhttp://www.jphysiol.org/>.
 - St Hilaire Ma, Gooley JJ, Khalsa SBS, Kronauer RE, Czeisler Ca, Lockley SW: **Human phase response curve to a 1 h pulse of bright white light**. *J Physiol* 2012, **590**:3035–3045. URL <http://doi.wiley.com/10.1113/jphysiol.2012.227892>.

21. Chang AM, Santhi N, St Hilaire M, Gronfier C, Bradstreet DS, Duffy JF, Lockley SW, Kronauer RE, Czeisler CA: **Human responses to bright light of different durations**. *J Physiol* 2012, **590**:3103–3112. URL <http://www.pubmedcentral.nih.gov/articlerender.fcgi?&artid=3406393&tool=pmcentrez&rendertype=abstract> <http://dx.doi.org/10.1113/jphysiol.2011.226555>.
 22. Gronfier C, Wright KP, Kronauer RE, Jewett ME, Czeisler CA: **Efficacy of a single sequence of intermittent bright light pulses for delaying circadian phase in humans**. *Am J Physiol Endocrinol Metab* 2004, **287**:E174–E181.
 23. Rimmer DW, Boivin DB, Shanahan TL, Kronauer RE, Duffy JF, Czeisler CA: **Dynamic resetting of the human circadian pacemaker by intermittent bright light**. *Am J Physiol Regul Integr Comp Physiol* 2000, **279**:1574–1579. <http://ajpregu.physiology.org/content/279/5/R1574.abstract>.
 24. Hannay KM, Booth V, Forger DB: **Macroscopic models for human circadian rhythms**. *J Biol Rhythm* 2019. 074873041987829URL, <http://journals.sagepub.com/doi/10.1177/0748730419878298>; 2019.
- Introduction of a new limit cycle oscillator model for human circadian rhythms with physiological variables and parameters. The model is tested against wearable data and compared with predictions made by other circadian models.
25. Forger DB, Jewett ME, Kronauer RE: **A simpler model of the human circadian pacemaker**. *J Biol Rhythm* 1999, **14**:533–538. <http://jbr.sagepub.com/cgi/doi/10.1177/074873099129000867>.
 26. Walch O, Huang Y, Forger D, Goldstein C: **Sleep stage prediction with raw acceleration and photoplethysmography heart rate data derived from a consumer wearable device**. *Sleep* 2019, **42**:1–19.
- Recent work on using commercial wearable device data in sleep research. Makes use of wearable data from apple watches and an artificial neural network to improve sleep predictions
27. Joyce DS, Zele AJ, Feigl B, Adhikari P: **The accuracy of artificial and natural light measurements by actigraphs**. *J Sleep Res* 2019:1–7.
 28. Flynn JI, Coe DP, Larsen CA, Rider BC, Conger SA, Bassett DR: **Detecting indoor and outdoor environments using the actigraph GT3X+ light sensor in children**. *Med Sci Sports Exerc* 2014, **46**:201–206.
 29. Arguelles-Prieto R, Bonmati-Carrion MA, Rol MA, Madrid JA: **Determining light intensity, timing and type of visible and circadian light from an ambulatory circadian monitoring device**. *Front Physiol* 2019, **10**:1–10.
 30. Bonmati-Carrion MA, Middleton B, Revell V, Skene DJ, Rol MA, Madrid JA: **Circadian phase assessment by ambulatory monitoring in humans: correlation with dim light melatonin onset**. *Chronobiol Int* 2014, **31**:37–51.
 31. Kantermann T, Burgess HJ: **Average mid-sleep time as a proxy for circadian phase**. *Psych J* 2017, **6**:290–291.
 32. Capella MdM, Martinez-Nicolas A, Adan A: **Circadian rhythmic characteristics in men with substance use disorder under treatment. Influence of age of onset of substance use and duration of abstinence**. *Front Psychiatry* 2018, **9**:1–13.
 33. Kitamura S, Hida A, Aritake S, Higuchi S, Enomoto M, Kato M, Vetter C, Roenneberg T, Mishima K: **Validity of the Japanese version of the Munich ChronoType questionnaire**. *Chronobiol Int* 2014, **31**:845–850.
 34. Terman JS, Terman M, Lo ES, Cooper TB: **Circadian time of morning light administration and therapeutic response in winter depression**. *Arch Gen Psychiatr* 2001, **58**:69–75.
 35. Stone JE, Phillips AJ, Ftouni S, Magee M, Howard M, Lockley SW, Sletten TL, Anderson C, Rajaratnam SM, Postnova S: **Generalizability of a neural network model for circadian phase prediction in real-world conditions**. *Sci Rep* 2019, **9**:1–17, <https://doi.org/10.1038/s41598-019-47311-4>.
- Use of an artificial neural network for circadian phase prediction using wearable data for shift workers. The ANN model generalizes well for diurnal subjects but not for night shift workers.
36. Kolodyazhnyi V, Späti J, Frey S, Götz T, Wirz-Justice A, Kräuchi K, Cajochen C, Wilhelm FH: **An improved method for estimating human circadian phase derived from multichannel ambulatory monitoring and artificial neural networks**. *Chronobiol Int* 2012, **29**:1078–1097.
 37. Kolodyazhnyi V, Späti J, Frey S, Götz T, Wirz-Justice A, Kräuchi K, Cajochen C, Wilhelm FH: **Estimation of human circadian phase via a multi-channel ambulatory monitoring system and a multiple regression model**. *J Biol Rhythm* 2011, **26**:55–67.
 38. Goodfellow I, Bengio Y, Courville A: *Deep learning*. MIT press; 2016.
 39. Kronauer RE, Czeisler CA, Pilato SF, Moore-Ede MC, Weitzman ED: **Mathematical model of the human circadian system with two interacting oscillators**. *Am J Physiol Regul Integr Comp Physiol* 1982, **242**.
 40. Hilaire MAS, Klerman EB, Khalsa SBS, Wright KP, Czeisler CA, Kronauer RE: **Addition of a non-photic component to a light-based mathematical model of the human circadian pacemaker**. *J Theor Biol* 2007, **247**:583–599.
 41. Asgari-Targhi A, Klerman EB: **Mathematical modeling of circadian rhythms**. *Wiley Interdiscipl Rev: Systems Biol Med* 2019, **11**:e1439. URL <http://doi.wiley.com/10.1002/wsbm.1439>.
 42. Hannay KM, Forger DB, Booth V: **Macroscopic models for networks of coupled biological oscillators**. *Sci Adv* 2018, **4**, e1701047.
 43. Woelders T, Beersma DG, Gordijn MC, Hut RA, Wams EJ: **Daily light exposure patterns reveal phase and period of the human circadian clock**. *J Biol Rhythm* 2017, **32**:274–286.
 44. Phillips AJ, Clerx WM, O'Brien CS, Sano A, Barger LK, Picard RW, Lockley SW, Klerman EB, Czeisler CA: **Irregular sleep/wake patterns are associated with poorer academic performance and delayed circadian and sleep/wake timing**. *Sci Rep* 2017, **7**.
 45. Patel SR, Weng J, Rueschman M, Dudley KA, Loreda JS, Mossavar-Rahmani Y, Ramirez M, Ramos AR, Reid K, Seiger AN, et al.: **Reproducibility of a standardized actigraphy scoring algorithm for sleep in a US hispanic/latino population**. *Sleep* 2015, **38**:1497–1503. <http://www.ncbi.nlm.nih.gov/pubmed/25845697> <http://www.pubmedcentral.nih.gov/articlerender.fcgi?artid=PMC4531418> <https://academic.oup.com/sleep/article-lookup/doi/10.5665/sleep.4998>.
 46. Kennedy MC, O'Hagan A: **Bayesian calibration of computer models**. *J Roy Stat Soc B* 2001, **63**:425–464.
 47. Stack N, Zeitzer JM, Czeisler C, Diniz Behn C: **Estimating representative group intrinsic circadian period from illuminance-response curve data**. *J Biol Rhythm* 2020, **35**:195–206.
- An innovative approach for recovering human circadian period averages at the group level from light response data without having to use a forced desynchrony protocol.
48. Swaminathan K, Klerman EB, Phillips AJK: **Are individual differences in sleep and circadian timing amplified by use of artificial light sources?** *J Biol Rhythm* 2017, **32**:165–176. <http://journals.sagepub.com/doi/10.1177/0748730417699310>.
 49. Czeisler CA, Duffy JF, Shanahan TL, Brown EN, Jude F, Rimmer DW, Ronda JM, Silva EJ, Allan JS, Jonathan S, et al.: **Stability, precision, and near-24-hour period of the human circadian pacemaker**. *Science* 1999, **284**:2177–2181.
 50. Carskadon MA, Labyak SE, Acebo C, Seifer R: **Intrinsic circadian period of adolescent humans measured in conditions of forced desynchrony**. *Neurosci Lett* 1999, **260**:129–132. <https://linkinghub.elsevier.com/retrieve/pii/S0304394098009719>.

51. Eastman CI, Suh C, Tomaka VA, Crowley SJ: **Circadian rhythm phase shifts and endogenous free-running circadian period differ between African-Americans and European-Americans.** *Sci Rep* 2015, **5**:1–8.
52. Stack N, Barker D, Carskadon M, Diniz Behn C: **A model-based approach to optimizing ultradian forced desynchrony protocols for human circadian research.** *J Biol Rhythm* 2017, **32**: 485–498.
53. Klerman EB, Dijk DJ, Kronauer RE, Czeisler CA: **Simulations of light effects on the human circadian pacemaker: implications for assessment of intrinsic period.** *Am J Physiol Regul Integr Comp Physiol* 1996, **270**.
54. Horne JA, Östberg O: **A self-assessment questionnaire to determine morningness-eveningness in human circadian rhythms.** *Int J Chronobiol* 1976.
55. Roenneberg T, Wirz-Justice A, Mrosovsky M, Roenneberg T, Mrosovsky M: **Life between clocks: daily temporal patterns of human chronotypes.** *J Biol Rhythm* 2003, **18**:80–90. <http://jbr.sagepub.com/cgi/content/abstract/18/1/80%5Cnhttp://jbr.sagepub.com/cgi/reprint/18/1/80.pdf>.
56. Duffy JF, Rimmer DW, Czeisler CA: **Association of intrinsic circadian period with morningness–eveningness, usual wake time, and circadian phase.** *Behav Neurosci* 2001, **115**: 895–899. <http://proxy.lib.sfu.ca/login?url=http://search.ebscohost.com/login.aspx?direct=true&db=pdh&AN=2001-07804-015&site=ehost-live%5Cnhttp://content.ebscohost.com/ContentServer.aspx?T=P&P=AN&K=2001-07804-015&S=L&D=pdh&OEbscoContent=dGJyMNxb4kSep7Q4yOvqOLCmr0yep7>.
57. Phillips AJK, Vidafar P, Burns AC, McGlashan EM, Anderson C, Rajaratnam SMW, Lockley SW, Cain SW: **High sensitivity and interindividual variability in the response of the human circadian system to evening light.** *Proc Natl Acad Sci USA* 2019. 201901824URL, <http://www.pnas.org/lookup/doi/10.1073/pnas.1901824116>; 2019.

This work establishes a high degree of variability in the sensitivity to evening light. Also demonstrates that important parameter variations exist outside of the circadian period parameter.

58. Bonarius Jochem H, Linnartz JPM: **Particle filter-based parameter estimation in a model of the human circadian rhythm introduction.** In *2018 symposium on information theory and signal processing in the benelux (SITB 2018), june; 2018*: 35–45.

Application of a particle filter approach to recover circadian parameters from wearable data.