

DSCC2020-3163

**STEERABLE NEEDLE TRAJECTORY FOLLOWING IN THE LUNG:
TORSIONAL DEADBAND COMPENSATION AND FULL POSE ESTIMATION WITH
5DOF FEEDBACK FOR NEEDLES PASSING THROUGH FLEXIBLE ENDOSCOPES**

Tayfun Efe Ertop*
Maxwell Emerson
Margaret Rox
Josephine Granna
Robert Webster

Department of Mechanical Engineering
Vanderbilt University
Nashville, TN 37212

Fabien Maldonado
Erin Gillaspie
Michael Lester

Vanderbilt University Medical Center
Nashville, TN 37212

Alan Kuntz
Robotics Center and
School of Computing
University of Utah
Salt Lake City, UT 84112

Caleb Rucker

The Department of Mechanical, Aerospace,
and Biomedical Engineering
University of Tennessee
Knoxville, TN 37996

Mengyu Fu
Janine Hoelscher
Inbar Fried
Ron Alterovitz

Department of Computer Science
University of North Carolina at Chapel Hill
Chapel Hill, NC 27599

ABSTRACT

Bronchoscopic diagnosis and intervention in the lung is a new frontier for steerable needles, where they have the potential to enable minimally invasive, accurate access to small nodules that cannot be reliably accessed today. However, the curved, flexible bronchoscope requires a much longer needle than prior work has considered, with complex interactions between the needle and bronchoscope channel, introducing new challenges in steerable needle control. In particular, friction between the working channel and needle causes torsional windup along the bronchoscope, the effects of which cannot be directly measured at the tip of thin needles embedded with 5 degree-of-freedom magnetic tracking coils. To compensate for these effects, we propose a new torsional deadband-aware Extended Kalman Filter to

estimate the full needle tip pose including the axial angle, which defines its steering direction. We use the Kalman Filter estimates with an established sliding mode controller to steer along desired trajectories in lung tissue. We demonstrate that this simple torsional deadband model is sufficient to account for the complex interactions between the needle and endoscope channel for control purposes. We measure mean final targeting error of 1.36 mm in phantom tissue and 1.84 mm in ex-vivo porcine lung, with mean trajectory following error of 1.28 mm and 1.10 mm, respectively.

INTRODUCTION

The development of bevel tip steerable needles [1] has been motivated by their potential to steer around obstacles and correct

*Corresponding author. Email: tayfun.efe.ertop@vanderbilt.edu

for factors like tissue deformation, needle bending, and other effects that cause traditional needles to miss their desired targets (see [2] and [3] for reviews). Steerable needles have been envisioned in applications including the liver [4], lungs [5], kidneys [4], prostate [6], and brain [7].

Interventions in the lung [5] are particularly promising, given the large number of potentially cancerous small nodules for which biopsy is desired [8], but for which complication rates are high for percutaneous approaches [9]. Similarly, existing bronchoscope-delivered needles either cannot reach these nodules (due to lack of an airway in close proximity), or cannot reach them with sufficient accuracy, leading to low yields [10]. Bevel tip steerable needles have the potential to enhance accuracy and steer to targets far from airways large enough to admit bronchoscopes. However, controlling the needle once delivered into the lung through a bronchoscope is challenging, for several reasons.

First, the bronchoscope requires an insertion pathway (bronchoscope channel) that is much longer than has previously been considered for percutaneous needle insertion. Second, it is desirable to use small-diameter needles for reducing required bronchoscope port diameter (and hence potentially facilitating the use of smaller diameter bronchoscopes, which can reach deeper into the lung). Another reason a small diameter is desirable is that it reduces needle bending stiffness, which reduces the loads required on bronchoscope tendons to deflect the tip of the bronchoscope with the needle loaded inside of it. Yet another motivation for smaller diameter needles is reduction of invasiveness for the patient; smaller needles require less tissue disruption during insertion [11]. Using 5 degree-of-freedom (DOF) tracking coils (0.41 mm diameter), rather than the larger 6-DOF versions (0.8 mm diameter) [12], for feedback enables smaller diameter needles.

Both long needle length and small diameter contribute to torsional windup along the needle in the presence of needle-port friction. Torsional friction acting along the long flexible needle shaft causes the needle tip rotation (about the needle axis) to be different than the needle base rotation. In general, it is challenging to model the interaction of a flexible metal cylinder passing through a larger tube (see e.g. [13]). Furthermore, the one DOF lost in moving from a larger 6-DOF magnetic tracking coil to a smaller 5-DOF coil is precisely the one we most care about – the axial angle at the needle’s tip, which defines the bevel’s direction, and hence the direction of forward progression of the needle as it is inserted.

The torsional friction challenge has been studied previously for steerable needles in the context of torsion induced by needle-tissue interactions [14–17]. These results assume continuous contact with tissue, as well as a length of unsupported needle outside the tissue, i.e. the context one would expect in a percutaneous insertion. Thus, these results do not apply directly to the non-uniform contact conditions along the interior of a bronchoscope port. In this paper, rather than model this complex

interaction explicitly, we instead assume a simplified model with Coulomb friction which induces a torsional deadband (i.e. a range of values for which the tip does not axially rotate when one changes the direction of base axial rotation), with the goal of determining whether such a model is sufficient for closed-loop control.

Various forms of observer models, including Extended Kalman Filters (EKF) have been used in a variety of ways to estimate state in both 2D and 3D needle steering. Needle-tissue interaction forces [18, 19] and needle deflection [20] have both been estimated using this method. In another study, an EKF was used to approximate belief dynamics in belief space planning for steerable needles [21]. EKFs have also been used to estimate a steerable needle’s planar, single-dimensional orientation from the needle’s two-dimensional position, sensed using a vision system [22, 23]. This concept has also been applied to full 6-DOF pose estimation from a sensed sequence of 3D needle tip positions [24, 25]. Observers have also been used previously to estimate the axial orientation of a steerable needle while sensing the position using a vision system [23, 26].

Our approach in this paper is inspired by these prior papers and adapts their EKF paradigm to the context of the torsional deadband mentioned previously. In terms of control, a wide variety of controllers have been developed for bevel-tip steerable needles (see [27] for a review). In this paper, we use the established sliding-mode controller proposed by Rucker et al. [28]. The primary contribution of our paper is to determine whether a simple Coulomb friction model can be used to effectively control a trans-bronchoscopic steerable needle with Kalman Filter-based roll estimation. We find that this simple modeling and control approach produces excellent accuracy when steering in phantom tissues and in inflated ex-vivo porcine lung tissue.

METHODS

In this section, we explain the integration of the torsional deadband into the steerable needle model, the Extended Kalman Filter estimation, and the needle steering controller. All vectors are written in bold lower case letters, and matrices are denoted with bold capital letters. The steerable needle model is illustrated in Figure 1, and the needle state estimation and control scheme is shown in Figure 2.

Review of Steerable Needle Kinematics

We use the nonholonomic steerable needle kinematic model described in [1]. In this model, the steerable needle is assumed to follow a constant curvature path when inserted, as shown in Figure 1. The needle curves in the direction of its bevel, and axial rotation changes the plane in which the needle curves. The




Figure 1. Steerable needle kinematics with the actuation variables at the base (v_1 and v_2) and the tip of the needle (u_1 and u_2).

needle motion is described by

$$\begin{aligned} \dot{\mathbf{p}} &= \mathbf{R} \begin{bmatrix} 0 \\ 0 \\ u_1 \end{bmatrix} \\ \dot{\mathbf{R}} &= \mathbf{R} [\boldsymbol{\omega}]_{\times} = \mathbf{R} \begin{bmatrix} 0 & -u_2 & 0 \\ u_2 & 0 & -\kappa u_1 \\ 0 & \kappa u_1 & 0 \end{bmatrix}, \end{aligned} \quad (1)$$

where the position vector of the needle tip is denoted by $\mathbf{p}$, and the orientation of the needle tip is denoted by the rotation matrix $\mathbf{R}$. Actuation variables for the needle are the insertion and rotation speeds, denoted by u_1 and u_2 , respectively. The angular velocity of the needle tip expressed in the body frame is given by $\boldsymbol{\omega} = [\kappa u_1 \ 0 \ u_2]^T$. The operator $[\]_{\times}$ represents the transformation to a skew-symmetric matrix. κ denotes the curvature of the needle.

The needle tip frame rotation can also be expressed using quaternions as

$$\begin{aligned} \begin{bmatrix} 0 \\ \dot{\mathbf{p}} \end{bmatrix} &= \mathbf{q} * \begin{bmatrix} 0 \\ 0 \\ 0 \\ u_1 \end{bmatrix} * \mathbf{q}^{-1} \\ \dot{\mathbf{q}} &= \frac{1}{2} \mathbf{q} * \begin{bmatrix} 0 \\ 0 \\ \kappa u_1 \\ u_2 \end{bmatrix} = \frac{1}{2} \mathbf{q} * \begin{bmatrix} 0 \\ \kappa u_1 \\ 0 \\ u_2 \end{bmatrix}, \end{aligned} \quad (2)$$

where the orientation of the needle tip is denoted by the unit quaternion $\mathbf{q}$ (using the convention that the scalar is the first com-

ponent), and the $*$ represents the quaternion multiplication operator. We will adopt this representation for the rest of our paper, and use it in our Extended Kalman Filter implementation.

Torsional Deadband-Aware Needle Model

In the needle model (2), the inputs u_1 and u_2 are the linear insertion speed and axial rotation speed of the needle tip, respectively. These are along and about the tip-frame z -axis, respectively, as shown in Figure 1. In the past, the needle has been assumed to be infinitely stiff axially and torsionally, meaning that velocities applied at the needle base are directly transmitted to the needle tip. Here, we continue to assume that the needle is axially stiff, such that the insertion speed at the base v_1 is transferred exactly to the needle tip ($u_1 = v_1$). However, we relax the assumption that the needle is infinitely torsionally stiff, and allow the needle's tip angle to lag behind its base angle, to account for torsional frictional effects. In other words, we assume tip rotation and base rotation speeds, u_2 and v_2 respectively, can differ. As mentioned in the introduction, such torsional dynamics have been previously modeled for steerable needles embedded in tissue [14, 16] using complex interaction models.

In this paper, we seek to determine whether the complex interactions between the needle and bronchoscope port, as well as needle-tissue interactions beyond the bronchoscope tip, can be modeled well enough for closed-loop control using a simple Coulomb friction model. Such a model implies a torsional deadband of fixed angular magnitude. Figure 3 illustrates the relationship between the base and tip of the needle, where the needle base rotation angle (as a function of time t) is denoted by $\theta_b(t)$, and the needle tip rotation angle is denoted by $\theta_t(t)$. These angles are given by integrating the rotation speeds at their corresponding locations:

$$\begin{aligned} \theta_b(t) &= \int_0^t v_2(\tau) d\tau \\ \theta_t(t) &= \int_0^t u_2(\tau) d\tau. \end{aligned} \quad (3)$$

In this deadband model, the input u_2 in the kinematic model (2) is a function of the tip angle $\theta_t(t)$, the actuator angle $\theta_b(t)$ and its derivative $v_2 = \dot{\theta}_b$ as follows:

$$u_2 = v_2 \ C(\theta_t, \theta_b, v_2), \quad (4)$$

where the function $C(\theta_t, \theta_b, v_2)$ checks the conditions for deadband and returns 1 or 0 as described below:

$$C(\theta_t, \theta_b, v_2) = \begin{cases} 1 & \text{if } \theta_t \leq \theta_b - c \ \wedge \ v_2 > 0 \\ 1 & \text{if } \theta_t \geq \theta_b + c \ \wedge \ v_2 < 0 \\ 0 & \text{otherwise.} \end{cases} \quad (5)$$




Figure 2. Overall control scheme for the proposed torsional deadband aware needle pose estimator and steering controller. The pose estimator components are shown in the Extended Kalman Filter box (in red). The controller takes the current needle pose estimate $\hat{\mathbf{x}}$ and determines the actuator commands at the needle base (v_1, v_2) to reach the target point $\mathbf{p}_d$. The feedback from the physical needle tip is obtained with the 5-DOF magnetic tracker, and sent to the EKF pose estimator.




Figure 3. The deadband model with fixed magnitude $2c$ that incorporates the Coulomb friction model is presented. An example hysteresis curve relating the base rotation θ_b to the tip rotation θ_t around the deadband range is shown by the path following counterclockwise direction. The shaded region represents the entire deadband range, and extends to infinity.

We define $2c$ as the total width of the rotational deadband which depends on the frictional forces on the needle. We assume c to be a constant based on our Coulomb friction model, and it can be determined prior to steering.

Extended Kalman Filter with 5-DOF Magnetic Tracker

Our Extended Kalman Filter uses magnetic tracker information from an embedded 5-DOF coil at the needle tip, in combination with the torsional deadband-aware needle model described in the previous section. We use the following notation: $\hat{\cdot}$ de-

notes an estimated state, $\cdot'$ denotes a predicted state, and $\tilde{\cdot}$ denotes a sensed quantity. The full 6-DOF pose of the needle tip is estimated by the EKF as $\hat{\mathbf{x}} = [\hat{\mathbf{p}}^T \hat{\mathbf{q}}^T]^T$, where $\hat{\mathbf{p}}$ denotes the estimated tip position and $\hat{\mathbf{q}}$ denotes the estimated unit quaternion representing the tip orientation. The actuation variables are taken as v_1 (insertion) and v_2 (rotation speed at the base, which affects u_2 through the deadband model). The torsion-aware needle model is used to predict the needle state based on the commanded actuation. The EKF prediction step is written as

$$\begin{aligned} \mathbf{x}'_n &= f(\hat{\mathbf{x}}_{n-1}, v_1, v_2, \Delta t) \\ \mathbf{P}'_n &= \mathbf{J}_{n-1} \mathbf{P}_{n-1} \mathbf{J}_{n-1}^T + \mathbf{Q}, \end{aligned} \quad (6)$$

where $\mathbf{x}'_n$ and $\mathbf{P}'_n$ are the new predicted state and state covariance matrix, respectively. The variable n is used for discrete time. Δt is the time step between each prediction. $f(\cdot)$ is the torsional deadband aware needle model function which is computed by discretely integrating (2) using a single forward-Euler step over Δt , with $u_1 = v_1$ and u_2 computed by the deadband model in (4) and (5). $\mathbf{J}$ is the 7×7 Jacobian matrix for the needle model, given by $\mathbf{J}_{n-1} = \frac{\partial f}{\partial \mathbf{x}}|_{\mathbf{x}=\hat{\mathbf{x}}_{n-1}}$. The matrix $\mathbf{Q}$ is the covariance of the process noise.

Sensor information is then used to update the estimate. The 5-DOF information retrieved from the electromagnetic tracker consists of the position, $\tilde{\mathbf{p}}_n$, and the needle bearing, which is the unit vector along the z-axis, $\tilde{\mathbf{v}}_{z,n}$. The discrete derivative of this needle bearing vector provides a vector indicating the direction in which the needle is curving, which is along the negative y-axis of the needle as shown in Figure 1. With normalization of the derivative, y-axis unit vector is given by

$$\tilde{\mathbf{v}}_{y,n} = -\frac{\tilde{\mathbf{v}}_{z,n} - \tilde{\mathbf{v}}_{z,n-1}}{|\tilde{\mathbf{v}}_{z,n} - \tilde{\mathbf{v}}_{z,n-1}|}. \quad (7)$$

The sensing vector is then constructed as $\tilde{\mathbf{y}}_n = [\tilde{\mathbf{p}}_n^T \tilde{\mathbf{v}}_{z,n}^T \tilde{\mathbf{v}}_{y,n}^T]^T$.

This is compared to the predicted value $\mathbf{y}'_n = h(\mathbf{x}'_n)$ to calculate the residuals as

$$\mathbf{r}_n = \tilde{\mathbf{y}}_n - \mathbf{y}'_n, \quad (8)$$

where $h(\cdot)$ is the function relating the state variables to the measurement. The following Kalman equations are used to update the estimate based on the residuals:

$$\begin{aligned} \mathbf{S}_n &= \mathbf{H}_n \mathbf{P}'_n \mathbf{H}_n^T + \mathbf{Z} \\ \mathbf{K}_n &= \mathbf{P}'_n \mathbf{H}_n^T \mathbf{S}_n^{-1} \\ \hat{\mathbf{x}}_n &= \mathbf{x}'_n + \mathbf{K}_n \mathbf{r}_n \\ \mathbf{P}_n &= (\mathbf{I} - \mathbf{K}_n \mathbf{H}_n) \mathbf{P}'_n, \end{aligned} \quad (9)$$

where $\mathbf{S}_n$ is the innovation covariance, $\mathbf{K}_n$ is a matrix containing the Kalman gains, and $\mathbf{I}$ is the identity matrix. $\mathbf{H}$ is the 9×7 Jacobian matrix for measurement function, defined by $\mathbf{H}_n = \frac{\partial h}{\partial \mathbf{x}} \big|_{\mathbf{x}=\mathbf{x}'_n}$. The matrix $\mathbf{Z}$ is the sensor noise covariance matrix. The brand new state estimate $\hat{\mathbf{x}}_n$ is used to inform the controller of the needle tip position.

Needle Steering Control

We used the basic Sliding Mode Controller from [28] for needle steering. The current needle tip state estimate $\hat{\mathbf{x}}_n$ from the EKF and the desired point, $\mathbf{p}_d$, are used to calculate the error vector in the needle tip frame, $\mathbf{e}$:

$$\begin{bmatrix} 0 \\ \mathbf{e} \end{bmatrix} = \begin{bmatrix} 0 \\ e_x \\ e_y \\ e_z \end{bmatrix} = \hat{\mathbf{q}}_n^{-1} * \begin{bmatrix} 0 \\ \mathbf{p}_d - \hat{\mathbf{p}}_n \end{bmatrix} * \hat{\mathbf{q}}_n. \quad (10)$$

Using the components of $\mathbf{e}$, angular error can be determined as

$$\sigma = \text{atan2}(e_x, -e_y), \quad (11)$$

where σ is the angular error between the desired curving direction and current estimated curving direction. Finally, the base insertion and rotation speeds are determined as

$$\begin{aligned} v_1 &= \lambda_1 \\ v_2 &= \lambda_2 \text{sign}(\sigma), \end{aligned} \quad (12)$$

where λ_1 and λ_2 are the pre-selected sliding mode controller parameters for insertion speed and rotation speed, respectively. The function $\text{sign}()$ is the basic signum function that returns -1 when




Figure 4. The robotic actuation unit, coupled with an Ambu aScope™ 4 bronchoscope.

the input is negative, +1 when the input is positive, and 0 when the input is 0.

Note that this control strategy can be used to steer toward a single target point as well as along a desired trajectory. To follow trajectories, the desired point $\mathbf{p}_d$ for the controller is moved along the trajectory at the insertion speed. A short lead distance is used to ensure the desired point is ahead of the needle along the trajectory.

EXPERIMENTS & RESULTS

We validated our proposed method by performing needle steering experiments in gelatin and ex-vivo porcine lung tissue. For all experiments, we used the transbronchial robotic needle insertion system shown in Figure 4 which is described in [5], with an updated actuation unit described in [29], connected to a flexible bronchoscope (Ambu aScope™ 4, Ambu, Inc., USA) with an 81 cm long working channel.

The needle we used in the experiments was utilizing a flexure tip concept, originally proposed in [30], with a cutout design for the flexure element [31, 32]. The head of the needle was made from a superelastic Nitinol tube with an inner diameter (ID) of 0.61 mm and an outer diameter (OD) of 0.86 mm. We machined the flexure hinge and the bevel into a short length of this tube. The finished needle head was then glued to the needle shaft, a Nitinol tube with 0.41 mm ID and 0.53 mm OD. We incorporated a 5-DOF magnetic tracking sensor (Northern Digital Inc., Canada) through the inner lumen of the needle for real-time feedback during steering. The sensor coil was placed distal to the needle hinge and fixed with glue. We measured the distance between the glued sensor location and the actual needle tip, and the offset distance was accounted for in software. The needle was deployed into the tissue for steering through a plastic intro-




Figure 5. The experimental setup for gel experiments.

duction tube that was passed through the bronchoscope. Note that there is some ambiguity in experiments associated with determining the initial torsional state within the deadband range, since the tip angle remains constant as the base angle begins to rotate, in that case. To overcome this in a practical setting, we simply rotated the needle in one direction before each insertion, to wind it up torsionally and cause it to start in a known torsional state.

The proposed state estimator and the controller were implemented using ROS (Robot Operating System). The control loop is operated at 40 Hz, which is the maximum update rate of the magnetic tracking system. The controller gains were selected to be $\lambda_1 = 5.0 \text{ mm/s}$ and $\lambda_2 = 90^\circ/\text{s}$. A lead distance of 3 mm was used for moving the desired point along trajectories in the controller. The sensor noise matrix for the Kalman Filter was chosen to be $\mathbf{Z} = \text{diag}([0.12 \mathbf{o}_{1 \times 3} \ 0.01 \mathbf{o}_{1 \times 3} \ 0.5 \mathbf{o}_{1 \times 3}])$ based on the sensor data sheet and the observed noise from the sensor with our system, where $\mathbf{o}$ is a vector full of ones. The process noise was selected as $\mathbf{Q} = \text{diag}([0.001 \mathbf{o}_{1 \times 3} \ 0.0005 \mathbf{o}_{1 \times 4}])$.

Steering in Phantom Tissue

As a preliminary validation of our proposed method, we performed steering experiments in 10% by weight Knox gelatin (Kraft Foods Global Inc., IL). The experimental setup for this experiment is shown in Figure 5.

Prior to the experiments, the width of the deadband was calibrated as $2c = 30^\circ$. This was done by observing the needle tip rotation with a glued 6-DOF tracking coil at the tip, and comparing it to the base rotation while the needle was in the introduction tube deployed through the bronchoscope. A standard 5 French flexible medical catheter (Cordis, USA) was used as the introduction tube. We also calibrated the radius of curvature of the needle in gelatin as $1/\kappa = 120 \text{ mm}$. The same gelatin recipe men-




Figure 6. Needle tip paths and the desired trajectories in the gel experiments. Kalman filter tip positions are plotted.

Table 1. Accuracy results for gel steering experiments. Both EKF and raw magnetic tracker sensor based needle tip positions were recorded for each insertion, and they are used to calculate the steering errors. The distance between the target and the needle tip position at the end of the insertion is given as the final targeting error. The trajectory error is calculated as the average of the needle tip distances to the desired trajectory throughout the steer.

Insertion Number	Final Targeting Error (mm)		Mean Trajectory Following Error (mm)	
	EKF Estimate	Raw Sensor	EKF Estimate	Raw Sensor
1	1.17	0.53	0.98	1.05
2	0.44	0.34	1.13	1.12
3	1.17	2.05	1.33	1.39
4	1.34	1.38	1.65	1.76
5	1.95	1.91	1.00	1.26
6	1.88	1.54	1.71	1.75
7	1.95	2.08	1.35	1.36
8	0.94	1.44	1.10	1.10
Mean	1.36	1.41	1.28	1.35

tioned above was used for this procedure as well. All calibrated values were incorporated into our EKF and needle model.

We generated eight 60 mm circular trajectories leading to arbitrarily selected target points within the needle's workspace to test targeting and trajectory following accuracy. The needle was




Figure 7. The experimental setup for ex-vivo lung experiments.

steered to these target points, and the Kalman Filter state estimate data was recorded during the insertions, as shown in Figure 6. Raw 5-DOF magnetic tracker sensor data was also recorded during these insertions to compare with the EKF accuracy. As can be seen in Figure 6, the needle successfully steered to the desired target points with 1.36 mm final mean targeting error, shown in Table 1. The needle was able to follow the desired trajectories with an average deviation of 1.28 mm based on the EKF. Table 1 also shows that the tip accuracy based on EKF is in agreement with the raw sensor accuracy, indicating that the EKF state estimates remain close to the physical needle tip positions.

Ex-vivo Lung Steering

To further test our methods in a clinical scenario, we performed steering experiments in ex-vivo porcine lungs with the experimental setup shown in Figure 7. The ex-vivo lungs were statically inflated with an air compressor. A pre-operative Computed Tomography (CT) scan (xCAT, Xoran Technologies, USA) was taken at the beginning of the experiments to identify lung anatomy and to determine target points. Airways and vasculature were segmented using the method of [33]. Fiducial markers were used to register the electromagnetic tracker to the CT scanner using point-based registration.

After performing registration, we deployed the bronchoscope into the airways and used a 21 gauge trans-bronchial needle (eXcelon, Boston Scientific Inc.) to pierce through the wall of the airway. We then slid a plastic sheath over this needle, and then removed the needle from the sheath. Next, we deployed our steerable needle through the plastic sheath. The steerable needle consisted of a coaxial solid flexure tip needle surrounded by a nitinol cannula, which supported the needle between the bronchoscope tip and the bronchial wall.

Once the needle was in place, we selected a target point in the needle's workspace. We then used a motion planner to generate a trajectory for the steerable needle from its current pose to




Figure 8. An example needle insertion from an ex-vivo experiment. The recorded needle tip position estimates and the planned trajectory are shown with respect to segmented anatomy. The airways are highlighted in blue, with major blood vessels in green.

this target while avoiding the obstacles identified by the segmentation, including vasculature, airways, and the pleura [33, 34]. Needle steering was initiated along this trajectory, and stopped when the needle reached the target point. In order to rapidly pass through the deadband during insertions, the needle's insertion velocity was reduced by a factor of 2 and the needle's axial rotation velocity was increased by a factor of 2.

The targeting results in ex-vivo lung are shown in Table 2. Figure 8 shows one of the insertions. As can be seen from these results, the needle was able to follow the desired trajectories with good accuracy.

CONCLUSION

This paper focuses on the challenge of controlling the path of a long, thin steerable needle delivered through the working channel in a bronchoscope. We sought to determine whether a simple Coulomb friction model (implying a torsional deadband) can be used to model the complex interactions between the needle and the bronchoscope port sufficiently for closed-loop control, acting on only 5-DOF electromagnetic tracking feedback. We found that, in combination with an Extended Kalman Filter and an established sliding mode controller, this approach enables accurate steering along desired trajectories to specified targets in phantom tissue and in inflated ex-vivo porcine lung. In future work, we plan to perform additional ex-vivo experiments

Table 2. Accuracy results for ex-vivo lung experiments. The needle tip positions obtained from our EKF and raw magnetic tracker sensor readings were both recorded for each insertion, and they were used to calculate the errors. The Final Targeting Errors are the distances between the needle tip position and the target, at the end of the insertion. The average of the needle tip distances to the desired trajectory throughout the insertion are reported as the Mean Trajectory Following Error.

Insertion Number	Final Targeting Error (mm)		Mean Trajectory Following Error (mm)	
	EKF Estimate	Raw Sensor	EKF Estimate	Raw Sensor
1	0.34	0.69	1.00	1.61
2	3.33	3.58	1.20	1.30
Mean	1.84	2.14	1.10	1.46

and eventually use this approach in-vivo in porcine lungs. As we move toward these experiments we will explore whether the coaxial cannula used in our ex-vivo experiments is necessary or whether the needle's overall sized can be reduced by omitting it. We will also explore the best nominal speeds for insertion and axial rotation, and whether speed variation is needed within the deadband. Furthermore, as we move toward in-vivo experiments, the system will require accurate image registration approaches as well as the potential need to compensation for respiratory motion. However, these challenges have been overcome in prior image guidance and robotic needle insertion studies so they do not seem insurmountable. Thus, this approach paves the way for accurate control of smaller diameter needles that do not hinder the bending capabilities of the tendon actuated bronchoscope tip, and reduce invasiveness, compared to the larger diameters used in the past to accommodate 6-DOF tracking coils.

ACKNOWLEDGMENT

This material is based upon work supported in part by the National Institutes of Health under grants T32EB021937 and R01EB024864. Any opinions, findings, and conclusions or recommendations expressed in this material are those of the authors and do not necessarily reflect the views of the NIH.

REFERENCES

- [1] Webster III, R. J., Kim, J. S., Cowan, N. J., Chirikjian, G. S., and Okamura, A. M., 2006. "Nonholonomic modeling of needle steering". *International Journal of Robotics Research*, **25**(5-6), pp. 509–525.
- [2] Abolhassani, N., Patel, R., and Moallem, M., 2007. "Needle insertion into soft tissue: a survey". *Medical Engineering and Physics*, **29**(4), pp. 413–431.
- [3] Reed, K. B., Majewicz, A., Kallem, V., Alterovitz, R., Goldberg, K., Cowan, N. J., and Okamura, A. M., 2011. "Robot-assisted needle steering". *IEEE Robotics and Automation Magazine*, **18**(4), pp. 35–46.
- [4] Majewicz, A., Marra, S. P., Van Vledder, M. G., Lin, M., Choti, M. A., Song, D. Y., and Okamura, A. M., 2012. "Behavior of tip-steerable needles in ex vivo and in vivo tissue". *IEEE Transactions on Biomedical Engineering*, **59**(10), pp. 2705–2715.
- [5] Swaney, P. J., Mahoney, A. W., Hartley, B. I., Ramirez, A. A., Lamers, E., Feins, R. H., Alterovitz, R., and Webster III, R. J., 2017. "Toward transoral peripheral lung access: combining continuum robots and steerable needles". *Journal of Medical Robotics Research*, **2**(1).
- [6] Alterovitz, R., Goldberg, K., and Okamura, A., 2005. "Planning for Steerable Bevel-tip Needle Insertion Through 2D Soft Tissue with Obstacles". *IEEE International Conference on Robotics and Automation*, pp. 40–45.
- [7] Minhas, D., Engh, J. A., and Riviere, C. N., 2009. "Testing of neurosurgical needle steering via duty-cycled spinning in brain tissue in vitro". *Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pp. 258–261.
- [8] Krishna, G., and Gould, M., 2008. "Minimally invasive techniques for the diagnosis of peripheral pulmonary nodules". *Current opinion in pulmonary medicine*, **14**, 07, pp. 282–6.
- [9] Kothary, N., Lock, L., Sze, D. Y., and Hofmann, L. V., 2009. "Computed tomography-guided percutaneous needle biopsy of pulmonary nodules: Impact of nodule size on diagnostic accuracy". *Clinical Lung Cancer*, **10**(5), pp. 360 – 363.
- [10] Baaklini, W. A., Reinoso, M. A., Gorin, A. B., Sharafkaneh, A., and Manian, P., 2000. "Diagnostic yield of fiberoptic bronchoscopy in evaluating solitary pulmonary nodules". *Chest*, **117**(4), pp. 1049 – 1054.
- [11] van de Berg, N. J., Dankelman, J., and van den Dobbelsteen, J. J., 2015. "Design of an actively controlled steerable needle with tendon actuation and FBG-based shape sensing". *Medical Engineering and Physics*, **37**(6), pp. 617–622.
- [12] Northern Digital Inc., (2020). "Aurora Electromagnetic Tracking System," [Online]. Available: <https://www.ndigital.com/medical/products/tools-and-sensors/>.
- [13] Kesner, S. B., and Howe, R. D., 2011. "Position Control of Motion Compensation Cardiac Catheters". *IEEE Transactions on Robotics*, **PP**(99), July, pp. 1–11.
- [14] Reed, K. B., Okamura, A. M., and Cowan, N. J., 2009.

- “Modeling and control of needles with torsional friction”. *IEEE Transactions on Biomedical Engineering*, **56**(12), pp. 2905–2916.
- [15] Reed, K. B., Okamura, A. M., and Cowan, N. J., 2009. “Controlling a robotically steered needle in the presence of torsional friction”. *IEEE International Conference on Robotics and Automation*, pp. 3476–3481.
- [16] Swensen, J. P., Lin, M., Okamura, A. M., and Cowan, N. J., 2014. “Torsional dynamics of steerable needles: Modeling and fluoroscopic guidance”. *IEEE Transactions on Biomedical Engineering*, **61**(11), nov, pp. 2707–2717.
- [17] Asadian, A., Patel, R. V., and Kermani, M. R., 2014. “Dynamics of translational friction in needle – Tissue interaction during needle insertion”. *Annals of Biomedical Engineering*, **42**(1), pp. 73–85.
- [18] Asadian, A., Kermani, M. R., and Patel, R. V., 2011. “A novel force modeling scheme for needle insertion using multiple Kalman filters”. *IEEE Transactions on Instrumentation and Measurement*, **61**(2), pp. 429–438.
- [19] Yan, K. G., Podder, T., Xiao, D., Yu, Y., Liu, T.-I., Ling, K. V., and Ng, W. S., 2006. “Online parameter estimation for surgical needle steering model”. *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pp. 321–329.
- [20] Sadjadi, H., Hashtrudi-Zaad, K., and Fichtinger, G., 2013. “Fusion of electromagnetic trackers to improve needle deflection estimation: Simulation study”. *IEEE Transactions on Biomedical Engineering*, **60**(10), pp. 2706–2715.
- [21] Sun, W., and Alterovitz, R., 2014. “Motion planning under uncertainty for medical needle steering using optimization in belief space”. *IEEE International Conference on Intelligent Robots and Systems*, pp. 1775–1781.
- [22] Wood, N. A., Shahrour, K., Ost, M. C., and Riviere, C. N., 2010. “Needle steering system using duty-cycled rotation for percutaneous kidney access”. *Annual International Conference of the IEEE Engineering in Medicine and Biology*, pp. 5432–5435.
- [23] Reed, K. B., Kallem, V., Alterovitz, R., Goldberg, K., Okamura, A. M., and Cowan, N. J., 2008. “Integrated planning and image-guided control for planar needle steering”. *IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechanics*, pp. 819–824.
- [24] Kallem, V., and Cowan, N. J., 2009. “Image guidance of flexible tip-steerable needles”. *IEEE Transactions on Robotics*, **25**(1), pp. 191–196.
- [25] Kallem, V., 2008. “Vision-based control on lie groups with application to needle steering”. PhD Thesis, Johns Hopkins University, Baltimore, MD.
- [26] Li, X., Lehocky, C. A., and Riviere, C. N., 2013. “Efficient 3d control for needle steering using duty-cycled rotation”. *International Conference on Informatics in Control, Automation and Robotics*, pp. 192–199.
- [27] Rossa, C., and Tavakoli, M., 2017. “Issues in closed-loop needle steering”. *Control Engineering Practice*, **62**, pp. 55–69.
- [28] Rucker, D. C., Das, J., Gilbert, H. B., Swaney, P. J., Miga, M. I., Sarkar, N., and Webster III, R. J., 2013. “Sliding mode control of steerable needles”. *IEEE Transactions on Robotics*, **29**(5), pp. 1289–1299.
- [29] Amack, S., Rox, M., Mitchell, J., Ertop, T. E., Emerson, M., Kuntz, A., Maldonado, F., Akulian, J., Gafford, J., Alterovitz, R., and Webster III, R. J., 2019. “Design and control of a compact modular robot for transbronchial lung biopsy”. *SPIE Medical Imaging: Image-Guided Procedures, Robotic Interventions, and Modeling*, **10951**, pp. 114 – 121.
- [30] Swaney, P. J., Burgner, J., Gilbert, H. B., and Webster III, R. J., 2013. “A flexure-based steerable needle: High curvature with reduced tissue damage”. *IEEE Transactions on Biomedical Engineering*, **60**(4), pp. 906–909.
- [31] Amack, S., 2019. “Design of a Compact, Workflow-Oriented Robot for Transbronchial Lung Biopsy”. Master’s thesis, Vanderbilt University, Nashville, TN.
- [32] Swaney, P. J., York, P., Gilbert, H. B., Webster III, R. J., Mahoney, A. W., and Wellborn, P., Dec. 2016. Surgical device tip with arc length varying curvature. U.S. Patent 20160346513A1.
- [33] Fu, M., Kuntz, A., Webster III, R. J., and Alterovitz, R., 2018. “Safe motion planning for steerable needles using cost maps automatically extracted from pulmonary images”. *IEEE International Conference on Intelligent Robots and Systems*, pp. 4942–4949.
- [34] Kuntz, A., Torres, L. G., Feins, R. H., Webster III, R. J., and Alterovitz, R., 2015. “Motion planning for a three-stage multilumen transoral lung access system”. *IEEE International Conference on Intelligent Robots and Systems*, Sept., pp. 3255–3261.