

Online Beam Current Estimation in Particle Beam Microscopy Through Time-Resolved Measurement

S. W. Seidel*[§], L. Watkins*, M. Peng*, A. Agarwal*, C. Yu[§], and V. K. Goyal*

**Department of Electrical and Computer Engineering,
Boston University, Boston, MA 02215, sseidel@bu.edu*

*§Charles Stark Draper Laboratory,
155 Technology Square, Cambridge, MA 02139*

Conventional particle beam microscopy requires knowledge of the beam current for accurate micrograph formation and sample milling. In practice, beam current fluctuations while the sample is horizontally raster scanned give rise to horizontal stripe artifacts in the micrograph, as shown for a synthetic SEM scenario in Figure 1(b). Existing mitigation techniques apply image processing tools to remove striped content post facto¹ (ex. Figure 1(c)) and may also remove true features of the underlying image. Time-resolved (TR) measurement, which divides the dwell time into n shorter sub-acquisitions, can be implemented without changes to the instrument and has enabled reconstruction techniques that mitigate the effects of source shot noise.² These techniques (Figure 1(d)) are naturally more robust to imperfect knowledge of the beam current.³ In this work we establish that beam current can be explicitly estimated *online* from time-resolved data, without the use of a calibrated sample. Our method prevents artifacts in the micrograph and provides the operator with novel information about instrument fitness.⁴

In Figure 1, we show our reconstruction results without (e) and with (f) total variation regularization of the micrograph. Both estimates have lower root-mean square error (RMSE) than the state-of-the-art mitigation technique. Figure 2 demonstrates that our beam current estimate is a close match to the true beam current. This estimate could be used to determine milling time or for feedback control of the current itself.

This work was supported in part by a Draper Fellowship, a Boston University Clare Boothe Luce Scholar Award, and by the US National Science Foundation under Grant No. 1815896.

¹ Barlow, A. J., Portoles, J. F., Sano, N., and Cumpson, P.J., “Removing beam current artifacts in helium ion microscopy: A comparison of image processing techniques,” *Microsc. Microanal.* **22**, 939 (2016).

² Peng, M., Murray-Bruce, J., Berggren, K. K., and Goyal, V. K., “Source shot noise mitigation in focused ion beam microscopy by time-resolved measurement,” *Ultramicroscopy*. **211**, 112948 (2020).

³ Watkins, L., Seidel, S. W., Peng, M., Agarwal, A., Yu, C., and Goyal, V. K., “Robustness of time-resolved measurement to unknown and variable beam current in particle beam microscopy,” *Proc. IEEE Int. Conf. Image Process.*, 3487 (2020).

⁴ Seidel, S. W., Watkins, L., Peng, M., Agarwal, A., Yu, C., and Goyal, V. K., “Online beam current estimation in particle beam microscopy,” (2021), arXiv:2111.10611 [physics.med-ph].




Figure 1: Synthetic Microscopy Results: Estimation algorithms assume a mean dose of $\bar{\lambda} = 200$ incident particles per pixel, although the actual dose is a Gaussian autoregressive random process with standard deviation $\sigma_{\lambda} = 40$. The nominal dose per sub-acquisition is 0.1 particles. Inset plots show the error between estimates and ground truth for a small image patch. Root-mean-square error (RMSE) is used as an error metric. Ground truth image taken from the Thermo Fisher Scientific microscopy image database scaled to secondary electron yield in $[0.2, 1]$.




Figure 2: Beam current estimate $\hat{\lambda}$, in units of incident particles corresponding to the estimate in Figure 1(c). Estimate RMSE = 6.6761.