

REVIEW ARTICLE

Biological mass spectrometry enables spatiotemporal 'omics: From tissues to cells to organelles

Leena R. Pade | Kaitlyn E. Stepler | Erika P. Portero | Kellen DeLaney | Peter Nemes

Department of Chemistry & Biochemistry, University of Maryland, College Park, Maryland, USA

Correspondence

Peter Nemes, Department of Chemistry & Biochemistry, University of Maryland, 8051 Regents Dr, College Park, MD 20742, USA.
Email: nemes@umd.edu

Abstract

Biological processes unfold across broad spatial and temporal dimensions, and measurement of the underlying molecular world is essential to their understanding. Interdisciplinary efforts advanced mass spectrometry (MS) into a tour de force for assessing virtually all levels of the molecular architecture, some in exquisite detection sensitivity and scalability in space-time. In this review, we offer vignettes of milestones in technology innovations that ushered sample collection and processing, chemical separation, ionization, and 'omics analyses to progressively finer resolutions in the realms of tissue biopsies and limited cell populations, single cells, and subcellular organelles. Also highlighted are methodologies that empowered the acquisition and analysis of multidimensional MS data sets to reveal proteomes, peptidomes, and metabolomes in ever-deepening coverage in these limited and dynamic specimens. In pursuit of richer knowledge of biological processes, we discuss efforts pioneering the integration of orthogonal approaches from molecular and functional studies, both within and beyond MS. With established and emerging community-wide efforts ensuring scientific rigor and reproducibility, spatiotemporal MS emerged as an exciting and powerful resource to study biological systems in space-time.

KEYWORDS

imaging, mass spectrometry, metabolomics, peptidomics, proteomics

1 | INTRODUCTION

Interdisciplinary efforts sculpted mass spectrometry (MS) into an indispensable tool in the molecular study of life processes (Figure 1). Virtually all levels of the molecular architecture—from genes to transcripts to proteins to peptides to metabolites—can now be measured using this technology, some in exquisite detail. MS detection and quantification of proteomes, peptidomes, and metabolomes ('omes) expanded the analytical toolbox of biology across the taxa, helping to generate and test new hypotheses. Over the last decades, technological advances broke down classical limitations in detection

limits (sensitivity), speed, and scalability to enable the analysis of limited amounts of materials. This review offers examples of innovations that extended MS-based 'omics to study the spatiotemporal organization of diverse biological systems at the realms of the organism, organ, tissue, cell, and organelle.

1.1 | MS meets research needs

Life processes respond to intrinsic and extrinsic events with complex and dynamic molecular changes across broad spatial and temporal dimensions, and measuring




FIGURE 1 Understanding biological processes hinges on the analysis of molecular heterogeneity that they orchestrate over broad spatial and temporal dimensions. Recent technological innovations and the development of specialized methodologies in sample collection and preparation, chemical separation, and HRMS analysis and detection provide sufficient sensitivity and spatiotemporal resolution for studying molecular processes at the levels of transcriptomes, proteomes, peptidomes, and metabolomes ('omes) in ever-deepening information content. Information from these studies fuels basic and translational research forward. Representative examples are highlighted from whole organismal to subcellular studies. Adapted with permission from (Gnann et al., 2021; Lombard-Banek et al., 2021). [Color figure can be viewed at [wileyonlinelibrary.com](https://onlinelibrary.wiley.com)]

these perturbations is key to their understanding. Systems biology promotes holistic understanding, in part, by characterizing all types of molecules and their potential interactions from whole organismal to sub-cellular scales. Vast signal amplification via polymerase chain reaction (PCR) facilitated next-generation sequencing of minute amounts of transcripts and genes, thus leading single-cell sequencing to become quasi-routine. Subcellular RNA-seq, for example, recently uncovered transcriptomic differences between the cell body (soma) and dendrites in individual GABAergic and glutamatergic neurons during biological functions (Perez et al., 2021), raising the question of subcellular heterogeneity for functional molecules downstream of transcription. Without technologies capable of amplifying whole proteomes to metabolomes, information from sequencing is often used as a proxy for the proteome.

However, the relationship between messenger RNA (mRNA) expression, protein production, and metabolism is complex in dynamic systems (Y. S. Liu et al., 2016). Posttranscriptional and -translational regulation as well as protein turnover rates contribute to widely variable correlation between transcription and translation. This relationship is further complicated by processes occurring

heterogeneously in space and time, as was recently found during cell culturing (Vogel & Marcotte, 2012), signal transduction between cells and neurons, differentiation of stem cells in developing organisms (Lombard-Banek, Reddy, et al., 2016; Peshkin et al., 2015), cancer invasion and metastasis (Saadatpour et al., 2015), and responses to drug treatments (Tian et al., 2004). Similarly, bioactive peptides and metabolites, which also carry out important physiological functions, are the result of interconnected molecular pathways, complicating their predictions from information upstream. Elucidating these molecular relationships over space-time is important for understanding states of health and disease, but this goal requires direct detection of proteins, peptides, and metabolites.

MS, especially high-resolution MS (HRMS), emerged as the technology of choice to bridge this knowledge gap. It integrates exceptional molecular specificity with a capability for quantification, typically without the need for functional probes. Readers interested in the fundamentals of MS and HRMS, ranging from ion generation to mass analysis, are referred to recent reviews (e.g., Feng et al., 2008). From tissues and cell cultures, hundreds to tens of thousands of biomolecules are routinely assessed in discovery studies, with near-complete coverage of proteomes (Geiger et al.,

2012; Hebert et al., 2014). For example, HRMS on two- and three-dimensional (2D and 3D, resp.) tissue cultures of human embryonic stem cells (hESCs) and induced pluripotent stem cells helped delineate molecular pathways critical for cell differentiation, cellular repair, and degeneration (Arnold et al., 2022; Cervenka et al., 2021; Sperber et al., 2015). Quantitative HRMS on proliferating neuronal stem cells revealed proteomic alterations between cellular and secreted proteomes in response to the microenvironment that would be complicated to predict from the transcriptome (Cervenka et al., 2021). HRMS quantified ~2800 proteins and their temporal expression profiles at progressive stages of differentiation, offering a glimpse into molecular mechanisms toward developing next-generation cell-based therapies such as cell grafting (Cervenka et al., 2021). Similarly, metabolic profiling of naïve and primed hESCs revealed alterations in epigenetic dynamics underpinning cell fate determination (Sperber et al., 2015).

The advent of highly sensitive HRMS propelled 'omics to dynamic and microscale biological systems. This group of technologies integrated existing approaches and invented new ones to accomplish trace-level sensitivity, sufficient to probe molecules at physiological concentrations. In doing so, they also allowed for scaling the spatial and temporal resolution of an analysis to the specific biological systems and questions at focus. Figure 1 conceptualizes broad spatiotemporal dimensions that contemporary HRMS is able to address. With sample collection capable of spatiotemporal scalability, high-sensitivity HRMS opened even more investigative possibilities. Rapid sampling enabled HRMS to assist studies on fast biochemical processes, including those underpinning cellular processes. For example, methods capable of fast sampling, such as capillary microaspiration with duty cycles on the order of seconds, can collect cellular materials even within a ~15–45 min cell cycle in cleavage-stage embryos, as shown from the South African clawed frog (*Xenopus laevis*) in Figure 1. Sampling over minutes to hours brought molecular studies to important physiological processes, including the maintenance of circadian cycles, feeding, and response to stress (see examples later). Additionally, scalability in physical dimensions raised the possibility of capturing spatial information on the distribution of molecules in specimens on the orders of millimeters to micrometers. MS imaging (MSI) played an important role by enabling the mapping of proteins, peptides, and metabolites across biological tissues, supporting functional studies and drug discovery. Profiling and imaging HRMS were recently adapted to subcellular structures. Chemical labeling enhanced spatial resolution to the domain of interacting molecules, thus vastly enhancing

knowledge of the biochemistry underlying biology, such as signaling between neurons in the brain.

These advancements in HRMS support basic and translational studies with information on the spatiotemporal organization of the molecular composition in biological systems. Next-generation separations, enhanced mass analyzers, and intelligent data acquisition strategies facilitated molecular identifications and quantification from minute amounts of materials. Thousands of proteins and metabolites can now be measured, even from single cells. For targeted proteins, antibodies tagged with heavy elements extended HRMS to MSI at single-cell resolution. Novel mathematical algorithms helped processing complex primary data in depth. User-friendly software packages and community-wide efforts to freely share them aided adoptability across laboratories needing sensitive 'omics, raising an opportunity to support applications in biological and translational studies. Molecular information that emerges from these measurements holds answers to existing questions and help generate new hypotheses. Efforts in pursuit of sensitivity, scalability, and 'omic coverage progressively expand our technological toolbox, thus accelerating the scientific cycle of exploration and discovery. This review highlights technological and methodological milestones in the development of spatiotemporal HRMS 'omics.

2 | SENSITIVE HRMS FOR SPATIOTEMPORAL 'OMICS

Spatiotemporal HRMS builds on the general workflow of MS-based 'omics. The use of MS in systems biology was recently reviewed elsewhere (Feng et al., 2008). For spatiotemporal analysis, the interconnected steps of the analytical pipeline are presented in Figure 2. During sampling, specimens, such as tissues and cells, are collected and lysed to release their contents by a variety of methods and technologies, with each tailoring to specific benefits. The lysate is typically enriched for the biomolecules of interest, such as proteins or metabolites, then subjected to sample clean-up/purification to render compatibility for instrumental analysis.

The biomolecules are measured using a combination of technologies. The chemical complexity of the sample is spread out over time through separation in solution, e.g., by using liquid chromatography (LC) and capillary electrophoresis (CE), or the gas phase using, e.g., ion mobility spectrometry (IMS) or field asymmetric IMS (FAIMS). The biomolecules are next ionized, detected, identified, and quantified by HRMS. An array of fragmentation (tandem/multistage MS) and analyzer-detector systems provide complementary speed and




FIGURE 2 General experimental workflow and representative technologies enabling spatiotemporal MS. Scalable sample collection and processing allow for probing biological specimens across broad scales in space and time. Separation in the liquid and/or gas phase reduces spectral complexity. High-resolution mass analyzers, tandem/multistage fragmentation, and efficient detector systems support molecular identifications with high specificity, sensitivity, and a capacity for quantification. Development of advanced tools from bioinformatics and chemometrics facilitate the analysis and interpretation of MS data with spatiotemporal insights. Technological abbreviations are in the text. [Color figure can be viewed at [wileyonlinelibrary.com](https://onlinelibrary.wiley.com)]

spectral resolution, ranging from quadrupole to time-of-flight (TOF), orbitrap, and Fourier-transform ion cyclotron resonance (FTICR) instruments.

The primary data from these measurements are processed and analyzed, usually with specialized MS bioinformatics software packages supporting identification of proteins, peptides, metabolites, and their post-translational modifications as well as relative or absolute quantification. Statistical and multivariate models from chemometrics allow for systematic evaluation of the observed effects between different experimental conditions under study. The results are interpreted in relation to canonical knowledge and the biological context. In successful investigations, the specifics of each step of the workflow are matched to the goals of the study. Following sections focus on technologies that advanced spatiotemporal analysis in different biological specimens.

2.1 | Sample collection

The type of question and biological system at hand *de facto* determines the method of sampling. In tissues displaying heterogeneous biochemistry in space or time, such as the brain and kidney as well as tumors, local sampling with accuracy ensures contamination-free

analysis from neighboring cell populations, thus aiding the interpretation of results. Likewise, sample collection with single-cell resolution unmasks differences between single cells (cell heterogeneity) within the same tissue, which are otherwise obscured due to signal averaging during cell pooling. As exemplified in Figure 3, various strategies were developed to collect and purify tissue biopsies, small populations of cells, single cells, or subcellular fractions of interest with different speeds and operational conditions.

2.1.1 | From biopsies to single cells

Manual dissection is simple and can have high accuracy. Surgical tools, ranging from sharp forceps to tungsten needles to eyebrow knives, are readily adaptable to excising distinct regions of tissues. With sufficient finesse, this approach also enables isolation of limited cell populations and individual cells from broad types of biological models. Dissection techniques, for example, enabled isolation of spatially distinct colon tumor biopsies from humans (Arnadottir et al., 2020), brain tissues from the mouse (X. H. Zhou et al., 2019), large single cells (Onjiko et al., 2015) and small cell populations (Baxi et al., 2018) from developing *X. laevis* embryos, neurons from *A. californica*




FIGURE 3 Representative strategies enabling scalable spatiotemporal sampling. Diverse technologies enabled HRMS with sufficient spatial and temporal resolving power to assess 'omics in broad types of biological systems and investigative contexts, ranging from ex vivo to in situ to in vivo analyses and from small populations of cells to single cells to subcellular organelles. Technological abbreviations are in the text. Adapted with permission from Azizian et al. (2021); Brown et al. (2018); Choi et al. (2022); Espino et al. (2019); Lombard-Banek, Moody, et al. (2019); Moroz et al. (2013); OuYang et al. (2015); Simone et al. (1998); von Eggeling et al. (2020); Yuan et al. (2021). [Color figure can be viewed at wileyonlinelibrary.com]

(Hanson et al., 2009), and distinct regions of nervous tissue from *Callinectes sapidus* (Hu, Helfenbein, et al., 2021). Needle-based biopsies, such as core needle (Satpathy et al., 2020) or fine needle (Piga et al., 2021) aspiration, were also used to collect small tissues or cells from humans. Labeling with dyes or fluorescent reporters enhances the accuracy of dissections under a microscope. Figure 3 demonstrates one such example, where identified cholinergic neurons from *A. californica* were dissected, permitting transcriptomics studies on aging (Moroz & Kohn, 2013).

Spatially resolved sampling reveals information on the physical location of the biomolecules. With narrow focusing and adjustable power (fluence), lasers naturally became a method of choice in this research field. As illustrated in Figure 3, laser capture microdissection (LCM) concentrates infrared or ultraviolet light to lift off

small tissue areas to groups of cells to single cells (e.g., 3–5 μm in size), with a capability for resolutions down to micrometers (Simone et al., 1998). LCM-based HRMS facilitated biological studies on genes, proteins, and metabolites in broad contexts including cancer biopsies (Shen et al., 2021) and single-cell proteomics (Zhu, Dou, et al., 2018). The technology was recently reviewed elsewhere (von Eggeling & Hoffmann, 2020). Because the dimensions of laser ablation are scalable, LCM can be employed to access cellular and subcellular structures. Representative examples include analysis of growth cones of a retinal ganglion cell (Zivraj et al., 2010), soma and dendrites of single neurons (Perez et al., 2021), and invadosomes coordinating motility of normal and cancer cells (Ezzoukhry et al., 2018). Dissection by LCM enables accurate isolation of tissue areas that is scalable to even

smaller regions. For example, LCM with LC-MS profiled lipids between the cell bodies and synapses of *Drosophila* neurons (Hebbar et al., 2014). Likewise, HRMS of the optic tectum and telencephalon regions in the zebrafish brain identified over 400 proteoforms with heterogeneous distribution, including many with important activities for visual and motor functions and neuropeptide signaling (Lubeckyj & Sun, 2022). These approaches provide an effective snapshot of the 'omic state of the specimen at a given time.

For smaller cell populations or dispersed cells, alternative strategies enhance the fidelity of sampling. Flow cytometry-based techniques, such as fluorescence-activated cell sorting (FACS), were employed in several studies to purify heterogeneous cell populations (Chappell et al., 2016; L. Chen et al., 2020; Sergent-Tanguy et al., 2003). As illustrated in Figure 3, FACS enables accurate and high-throughput purification of cell populations to single cells based on fluorescent markers. Although specialized cytometers are able to sort particles ranging from 0.2 to 150 μm in diameter, commercial FACS instruments usually have a low-pass limit of $\sim 20 \mu\text{m}$ cell diameter (Davey & Kell, 1996). Required expertise, expense, and technical limitations for larger cells culminated into alternative methods. In magnetic-activated cell sorting, for example, antibody-conjugated magnetic beads of several nanometers in diameter can be used to recognize surface markers before pulling down specific types of cells and subcellular organelles (Miltenyi et al., 1990).

Microfluidics renders added benefits in sample collection with minimal manual intervention. Cell sorting in microfluidic chips employs orthogonal mechanisms for isolation, ranging from acoustic waves to radiation forces to hydrodynamic flow to microfiltration and cellular immobilization. These and other emerging technologies were recently reviewed elsewhere (Y. Y. Zhang et al., 2021). It follows that these approaches can be extended to collecting single cells. However, due to mixing during processing, information on the spatial origin of each single cell is lost, which may be important in the context of biology.

2.1.2 | Subcellular organization

For targeted subcellular analysis, distinct organelles may be isolated prior to HRMS detection. Sequential lysis of cellular compartments and density-based centrifugation are well-established for cells and cell types from multiple biological models. Organelles such as nuclei, mitochondria, peroxisomes, lysosomes, and the Golgi apparatus can be separated based on their size, shapes, and surface charges (Tharkeshwar et al., 2018). Similar

centrifugation-based strategies allowed for the isolation of exosomes from myeloid-derived suppressor cells involved in cancer tumor signaling (Burke et al., 2014) and from *Bacillus subtilis* at different developmental stages (Y. Kim et al., 2016). Organelles with similar densities, however, may pose challenges for high-purity isolation.

Biochemical approaches and bioinformatics provide alternatives for subcellular analyses. Subcellular labeling HRMS was the topic of a previous review (Y. H. Lee et al., 2010). Localization of organelle proteins by isotope tagging (LOPIT) and hyperLOPIT precluded the need for high-purity organelle purification by taking a multipronged approach, in which organelles were partially separated into multiple fractions using multistep density centrifugation for multiplexed quantification via LC-HRMS, followed by deep machine learning to extract information on the subcellular localization of detected proteins (Geladaki et al., 2019). The HRMS analysis of organelle fractionation is improved in MS-compatible buffers by reducing chemical interferences during detection. For example, radio-immunoprecipitation assay (RIPA) buffer with ethylene glycol and phase transfer surfactants aided isolation of cytoplasmic, nuclear, and organelle fractions from small populations of cells without the need for traditional proteomic cleanup steps (Masuda et al., 2020).

Integration of orthogonal methods of isolation enhances capacity, and so molecular specificity and accuracy. In fluorescence-activated organelle sorting (FAOS), subcellular targets are fluorescently labeled, for example, using antibodies for two-step purification via ultracentrifugation and FACS. This strategy enabled purification of synaptosomes without neuronal and glial contaminants, identifying novel proteins released in glutaminergic synapses (Biesemann et al., 2014). Magnetic isolation using differentially coated superparamagnetic iron oxide nanoparticles (SPIONs) helped purify endosomes during the study of Niemann-Pick disease, a lysosomal storage disorder, by pulling down endosomes enriched with dimercaptosuccinic acid-coated SPIONs and plasma membranes adhering to aminolipid-coated SPIONs (Tharkeshwar et al., 2017).

MSI techniques such as micro- and nano-probe HRMS with bioinformatics emerged for precision organelle analysis, precluding the need for physically isolating microscopic organelles. By tightly focusing beams of projectiles, from ions to photons (recall Figure 2), spatial 'omics recently opened the door to subcellular 'omics, as was the focus of a recent review (Seydel, 2021). High-definition multiplex ion beam imaging (HD-MIBI) secondary ion MS (SIMS) offered a 30-nm resolution, sufficient to localize cisplatin enrichment in nuclear

speckles with exclusion from chromatin regions, providing new information on subcellular microenvironments for drug development (Rovira-Clave et al., 2021). Spatial single nuclear metabolomics (SEAM) integrated TOF-SIMS with novel computational data analysis modules (Figure 3) to identify and localize single nuclei with variable metabolic patterns in healthy and fibrotic liver tissue (Yuan et al., 2021). In the 3D OrbiSIMS, the coupling of TOF-SIMS with an ultrahigh-resolution orbitrap empowered high-speed 3D molecular imaging, providing information on the subcellular localization of lipids and neurotransmitters in the hippocampus and single macrophages (Passarelli et al., 2017). Similarly, using image-guided matrix-assisted laser desorption ionization (MALDI) coupled to an ultrahigh-resolution mass analyzer, neuropeptides and lipids were identified from 0.5 to 2 μm vesicles in *A. californica* to reveal distinct metabolic subpopulations within these organelles (Castro et al., 2021). These and other advancements enabling subcellular and single-organelle spatial 'omics are expected to continue informing us of molecular profiles among cells and within their components.

In situ sampling can take a molecular snapshot of the tissues or cells under innate conditions with minimal external perturbations during collection. As illustrated in Figure 3, fast photochemical oxidation of proteins (FPOP) utilizes laser-induced photolysis of hydrogen peroxide to irreversibly label proteins in their native state within a cell, thus enabling studies on protein structure and folding and protein-protein interactions (Espino & Jones, 2019). Pioneering a new approach in epitope mapping, oxidative labeling was successfully used to map antibody-binding regions of the protein serine protease thrombin (Jones et al., 2011). Such detailed information on molecular identity, structure, and interactions is essential to develop monoclonal antibody-based therapeutics for targeted proteins.

2.1.3 | In situ/vivo

Chemical labeling enables the isolation of molecular partners in interaction. A plethora of labeling techniques allow for tracing biomolecules in subcellular compartments, and some in vivo. Bioorthogonal reactions, for example, use a wide array of click chemistry to anchor specific biomolecules without disturbing the intracellular environment. Such molecular handles were used to label proteins, lipids, glycans, nucleic acids, and enzymes, and for targeted delivery of small molecules, as was reviewed in (Bird et al., 2021). Bioorthogonal noncanonical amino acid tagging (BONCAT), Figure 3, enabled the measurement of newly synthesized proteins by incorporating biotin-tagged

noncanonical amino acids in the nascent proteins with follow-up affinity purification or imaging by fluorescence microscopy (Landgraf et al., 2015). The approach recently allowed for mapping the proteome of tumor cells in ductal adenocarcinoma in the mouse (Azizian et al., 2021). BONCAT was recently used to find local translation of 42 proteins in the axoplasm during regeneration of the sciatic nerve following injury in rats (Di Paolo et al., 2021). Alternatively, proximity labeling using ascorbate peroxidase, APEX or APEX2 (Lam et al., 2015), or biotin ligase, BioID (Roux et al., 2012), localized enzymatic activity to map their proteomic environment in organelles, thus opening the door to proteomic studies on synaptic clefts of excitatory and inhibitory neurons (Loh et al., 2016). Isotope tracing adds a dynamic component to these studies via spatial fluxomics. Glutamine metabolism was monitored in mitochondria and the cytosol of cancer cells to elucidate metabolome remodeling (W. D. Lee et al., 2019). Alternatively, subcellular fractionation with activity-based protein profiling (ABPP) monitored proteomic changes in secretory pathways in the endoplasmic reticulum in response to cellular stress (Bechtel et al., 2020).

Sampling in vivo extends HRMS 'omics to physiological states. Figure 3 shows examples of in vivo probes capable of minimally invasive sampling. Microdialysis has a long history in the research of neuropeptides, neurotransmitters, and metabolites with ca. millimeter resolution in live biological models. This strategy was used to access molecular information on physiologically relevant timescales (e.g., from minutes to hours) in functional, behavioral, and clinical studies. Representative examples helped decipher the spatiotemporal dynamics of neuropeptides and their roles in feeding using crabs (Schmerberg & Li, 2013; Schmerberg et al., 2015), decision-making and drug response in rats (Zestos & Kennedy, 2017), and the tumor-promoting impact of radiation in glioblastoma using human xenografts in mice (Gupta et al., 2020), as well as clinically for cerebral metabolic analysis of human patients with aneurysmal subarachnoid hemorrhage among others (Helbok et al., 2017).

Real-time biopsies with rapid chemical analysis opened exciting new capabilities for health research (Figure 3). Liquid biopsies helped capture circulating cancer cells for diagnosis (Abouleila et al., 2019) or the assessment of effectiveness during therapy (Bensen et al., 2021). In the clinical setting, desorption electrospray ionization (DESI) and MasSpec Pen (Figure 3) gained recognition for its ability to downscale microextraction into <10 μl of water droplet on tissue regions of interest, even in morphologically complex samples (Brown et al., 2018; J. L. Zhang et al., 2017). With fast sample transfer and HRMS detection, the approach aided cancer

diagnoses of ovarian, lung, thyroid, and breast biopsies *ex vivo*, and recently also *in vivo*, thus empowering surgical applications with a new tool.

Microsampling confined to capillaries and nanochannels downscales sampling to the realm of cells and organelles. Whole cells, portions of a cell, or subcellular contents may be aspirated using precision-fabricated capillaries, usually with assistance from fine translation control and optical microscopy. The introduction of ambient ionization and ultra-sensitive HRMS ushered the extension of capillary microsampling to endogenous and exogenous metabolites in broad types of single cells, as was recently reviewed elsewhere (Y. Y. Yang et al., 2017; Zhang & Vertes, 2018; Zhu, Shao, et al., 2021). Early on, for single cells, live-video MS with direct electrospray ionization (ESI) detected metabolites from single adherent embryonic fibroblasts (mouse) and plant cells (Fujii et al., 2015; Tsuyama et al., 2008), T-probe (Figure 3) profiled metabolic perturbations in HeLa cells on anticancer treatment (R. M. Liu et al., 2018), and laser ablation electrospray ionization (LAESI) surveyed secondary metabolites in *Allium cepa* cells under guidance by optical microscopy (Taylor, Mattson, et al., 2021). The live cell analysis device (LCAD) exchanged capillaries for nanofabricated channels, where localized electroporation facilitated tyrosine phosphatase extraction from cells during proliferation and hydrogen peroxide treatment in human triple-negative breast cancer cells with protein detection enabled by self-assembled monolayers for matrix-assisted laser desorption/ionization (SAMDI) MS (Mukherjee et al., 2020).

To detect deeper into single-cell 'omes, microsampling can be combined with high-efficiency separations before *in situ/vivo* HRMS. By aspirating <5% of the total cell volume using a pulled borosilicate capillary, proteomic and metabolomic cell heterogeneity was found in embryos of *X. laevis* and zebrafish (Lombard-Banek, Moody, et al., 2019; Onjiko et al., 2017; Portero & Nemes, 2019; Saha-Shah et al., 2019). Capillary microsampling facilitated integration of HRMS 'omics with functional biology. Patch-seq has a history of leveraging this strategy for sequencing mRNAs from neurons after their electrophysiological characterization (Cadwell et al., 2016), as was recently reviewed elsewhere (Lipovsek et al., 2021). After eavesdropping on the electrical activity of single neurons using a patch pipette, neurotransmitters in the mouse (Aerts et al., 2014), the American cockroach (Neupert et al., 2018), and recently also proteins in the mouse (Choi et al., 2022) were measured in identified single neurons in the central nervous systems of these model organisms. Multi-step capillary microaspiration also enabled *in vivo* analysis of dynamic reorganization in the metabolome and proteome of single identified

blastomeres dividing in the live *X. laevis* embryo, which successfully developed to a functioning tadpole with indistinguishable behavior from wild-type siblings in a background color preference assay (Lombard-Banek et al., 2021). Accessing such information on molecular interactions in space-time offers an important piece of the puzzle towards the holistic understanding of molecular mechanisms underlying biological functions and disease states.

2.2 | Sample preparation

Sample preparation can make or break studies requiring trace detection, and those addressing spatiotemporal processes are no exception. Specimens encompassing few cells, single cells, or subcellular organelles provide only micro- to picogram quantities of starting materials (Ctorteccka & Mechtler, 2021; Zhang & Vertes, 2018). Processing such limited amounts poses nontrivial challenges in handling of miniscule volumes without solvent evaporation and with minimal-to-no losses for proteins, peptides, and metabolites due to surface adsorption (Z. C. Yang & Sun, 2021). Advances in small-scale sample preparation was the topic of several reviews (Hu, Li, et al., 2023; Kelly, 2020; Matula et al., 2020; Qin et al., 2022; Taylor, Lukowski, et al., 2021; Z. C. Yang & Sun, 2021). The following section highlights developments that ushered biological HRMS to spatiotemporal 'omics from small tissue biopsies down to the single-cell and subcellular level.

2.2.1 | Proteomics

HRMS provides critical information on proteomes during states of health and disease. Methods and technologies enabling bottom-up, middle-down, and top-down proteomics are able to identify and quantify broad types of proteins, assess their structure, gauge their interactions, and determine modifications during biological processes in space and time. Deep proteome coverage is an enabling feature of bottom-up proteomics, which is the most widely used of the three workflows. In this approach, proteins are enzymatically digested to peptides, which can be efficiently ionized and sequenced to pinpoint their source proteins using advanced bioinformatics. Thousands of different proteins can be identified using this approach, even from limited samples. Bottom-up proteomics was applied to study complex spatiotemporal processes at the organismal to subcellular levels in development (Baxi et al., 2021; L. W. Yang et al., 2020), neuroscience (Hobson et al.,

2022; Velasquez et al., 2019), and various disease states (Vo & Palsson, 2007; Waas & Kislinger, 2020). The information from these measurements provides the foundation for hypothesis-driven and functional studies.

HRMS-based proteomics executes multiple analytical steps with each capable of affecting sensitivity. Lysis, extraction, reduction, alkylation (of disulfide bonds), enzymatic digestion, and desalting are routine stages of sample processing requiring care for limited samples. Simplification of the workflow was recently found to improve performance. For single-cell measurements, for instance, MS-compatible detergents and mechanical lysis methods (i.e., freezing/thawing or laser pulses) and exclusion of reduction and alkylation eliminated the need for downstream clean-up, and reduced nonspecific adsorptive losses and dilution (Lombard-Banek, Moody, et al., 2019; Schoof et al., 2021). Select strategies enabling high-fidelity proteomics are presented in Figure 4 and discussed further.

“One-pot” strategies confined the analytical steps to a single location. For amount-limited samples, immobilization of molecular processing and reduction or elimination of sample transfers proved particularly beneficial. For example, gel-aided sample preparation (GASP) (Fischer & Kessler, 2015), suspension trapping

filter (S-Trap) (HaileMariam et al., 2018), and filter-aided sample preparation (FASP) (Wisniewski et al., 2009) used gels and filter membranes, respectively, to reduce sample transfer steps and improve proteome coverage from hundreds of micrograms of protein. By miniaturizing and confining sample processing onto a multi-well plate, micro-FASP was able to identify ~3400 proteins from ~200 ng of yolk-free proteome from single blastomeres that were isolated from early-stage *X. laevis* embryos (Zhang, Dubiak, et al., 2020). Single-pot solid-phase-enhanced sample preparation (SP3) demonstrated similar capabilities for comparable protein amounts from 1000 HeLa cells and *Drosophila* embryos by immobilizing proteins and peptides on carboxylate-coated paramagnetic beads throughout sample preparation (Hughes et al., 2014).

Sample processing in situ provided an opportunity to streamline the analytical workflow. As illustrated in Figure 4, single-cell proteomics by MS (SCoPE-MS) confined sample processing to one pot, allowing for multiplexed protein quantification among single cells (Budnik et al., 2018). Capillary-based systems may serve as alternative substrates for single-cell sampling and processing. As shown in Figure 4, the integrated proteome analysis device (iPAD) directly supplied a




FIGURE 4 Scalable and sensitive sample processing for metabolites and proteins for spatiotemporal MS. One-pot sample processing (SCoPE), direct cell capture and processing (iPAD), and miniaturization of sample processing steps (nanoPOTS) deepen the detectable proteome by reducing analyte losses. Direct cell capture (SCP-LVC), multi-solvent processing, and in situ metabolite extraction (LMJ) after MALDI imaging improve coverage of the single-cell metabolome. Technological abbreviations are in the text. Adapted with permission from Budnik et al. (2018); Cahill et al. (2019); Comi et al. (2017); Kelly (2020); Onjiko et al. (2016); Shao, Wang, et al. (2018). [Color figure can be viewed at [wileyonlinelibrary.com](https://onlinelibrary.wiley.com)]

suspension of living cells (iPAD-100) for online cell lysis, protein digestion, and peptide desalting and trapping to a 10-port valve-based system on a commercial nanoLC, identifying ~600–650 proteins from a single injection of 100 cells (~10 ng protein) using nanoLC-HRMS (Q. Chen et al., 2015). In pursuit of cell heterogeneity, iPAD-1 (Figure 4) used an ultranarrow-bore analytical column to identify ~270 proteins from a single HeLa cell (220 pg proteome) in 1 h of analysis with nanoLC-HRMS (Shao et al., 2018). These and other simplified devices facilitate technology adaptation.

Microfabricated devices naturally give way to “one-pot” workflows. ProteoCHIP used polytetrafluoroethylene (PTFE)-based microfluidics to automate single-cell isolation toward small-volume handling, enabling the processing of up to 192 samples simultaneously for LC-MS (Ctortekka, Hartlmayr, et al., 2022). As shown in Figure 4, nanodroplet processing in one pot for trace samples (nanoPOTS) miniaturized the processing of 100 cells to a single mammalian cell onto nano-wells on microfabricated glass chips using nanoliters of reagents for deep proteomics (Zhu, Clair, et al., 2018; Zhu, Piehowski, et al., 2018). Using LCM to dissect ~100 μm squares from mouse tissues, the approach helped uncover spatial differences between uterine cell types during pregnancy in preparation for blastocyst implantation (Piehowski et al., 2020). NanoPOTS also enabled proteomics on single sensory hair cells measuring ~0.6–1 pl in volume, facilitating the study of cell differentiation in chicken embryos (Y. Zhu et al., 2019). Nested nanoPOTS (N2) arrayed multiple wells to enhance measurement throughput for multiplexed quantification using nanoLC-HRMS (J. Woo et al., 2021). Recent efforts sought to further refine one-pot sample preparation strategies to make the setups cost-effective and allow scalable sample preparation, as was recently reviewed in Slavov (2022). Automated preparation in one pot for trace samples (autoPOTS), the latest upgrade of nanoPOTS, now allows automated processing of 1–500 cells using commercially available instrumentation (Liang et al., 2021).

Similar strategies also emerged for top-down proteomics. Detection of intact proteins provides insights into proteoforms and posttranslational modifications (Catherman et al., 2014). With fewer analytical steps than its bottom-up sibling, miniaturized formats for detergent removal and desalting (Nickerson et al., 2021; Yu et al., 2021) and “one-pot” sample preparation (Dagley et al., 2019; M. W. Zhou et al., 2020) extended top-down proteomics to limited sample amounts as low as hundreds to tens of nanograms. As mentioned earlier, analyte recovery can be enhanced by eliminating the routines of sample handling

and desalting. Freeze-thaw lysis and nonionic surfactant-based extraction from ~200 brain cells enabled spatial top-down proteomics for ~400 proteoforms between the optic tectum and telencephalon regions of the zebrafish brain (Lubeckyj & Sun, 2022). Further advancements in technology raise exciting potentials to leverage proteoforms as biomarkers of diseases (Smith et al., 2021).

2.2.2 | Peptidomics

Endogenous peptides, such as neuropeptides, can provide additional insight into biological processes, particularly related to the nervous and neuroendocrine systems. These peptides are synthesized from larger proteins via enzymatic cleavage, and their biological activity is often distinct from the prohormone or precursor-related peptides (Hook et al., 2018). Many neuropeptides contain similar sequences that may vary by only one or a few amino acids, but their biological functions could be identical or largely different (DeLaney et al., 2018). Furthermore, neuropeptides are thought to be the most diverse class of signaling molecules, are present in low abundances in vivo, and are prone to rapid degradation. As such, analyzing endogenous peptides in space and time presents unique challenges and requires a rapid, sensitive workflow with specificity capable of differentiating between peptide isoforms. MS emerged as the gold standard for profiling neuropeptides, as it is capable of rapidly detecting neuropeptides at sub-attomole levels and can identify peptides without prior knowledge (DeLaney, Phetsanthad, et al., 2022).

Peptidomics consists of peptide extraction, purification via molecular weight cutoff filtration, and desalting before MS analysis. A critical aspect of this process is deactivation of proteolytic enzymes to prevent peptide degradation and achieve highly sensitive peptidomics analyses, which can be accomplished by flash-freezing the tissue immediately after dissection, heat stabilization, or acidification. Acidification is particularly effective for deactivating proteases when the acid is in an organic solvent (e.g., acidified methanol or acidified acetonitrile), as while proteases are deactivated by the low pH, large proteins are precipitated from the solution. For example, acidified methanol was recently used to successfully extract neuropeptides from single mollusk neurons, enabling sequencing of peptides in the cytoplasm and nucleus of the neurons and identification of differences in peptide abundances based on subcellular location (L. W. Zhang, Khattar, Kemenes, et al., 2018). In MALDI-MS workflows, 2,5-dihydroxybenzoic acid (DHB) was found ideal for peptide extraction, as it served as both the extraction buffer and the matrix (Romanova et al., 2008; J. H. Wang et al., 2009). Though other stabilizing agents

such as EDTA and protease inhibitor cocktails can also be added to the extraction buffer to improve peptide stability (R. B. Chen et al., 2009; Kwok & Tobe, 2006; Onorato et al., 2019), MS compatibility is important to consider, particularly with limited samples, so as to reduce additional cleanup steps and sample losses.

Peptidomics from limited samples may also benefit from molecular enrichment to boost sensitivity. Affinity-enhanced magnetic beads offered one such approach to improve detection for low-abundance peptides (Schmerberg & Li, 2013; Vocat et al., 2020). The incorporation of magnetic beads coated with NPY02 antibodies during sample preparation led to femtomolar sensitivity and quantitation of neuropeptide Y in human plasma (Vocat et al., 2020). Considerations of sample recovery for the desalting step are similar as in bottom-up proteomics workflows. Additionally, solid-phase extraction (Bardsen et al., 2019) and on-plate droplet desalting for MALDI-MS (S. Wang et al., 2017; Yoon et al., 2018) improved the sensitivity of neuropeptide analyses. Solid-phase extraction was also brought online to seamlessly clean up neuropeptide samples before LC-MS (Axel et al., 2017).

Alternatively, MSI strategies are popular for direct measurement and spatial localization of bioactive peptides, particularly in limited samples. Sampling with a focused laser beam requires minimal sample preparation, yet yields a sufficient amount of analytes for ionization and detection. Development of validated protocols made sample preparation for MSI straightforward and robust (Chatterji & Pich, 2013; DeLaney, Phetsanthad, et al., 2022; Reyzer & Caprioli, 2013). They predominantly involve sectioning of frozen or embedded tissues or dissection of cells; however, careful steps need to be undertaken to maintain sample integrity and reproducibility within measurements. To remove interfering contaminants and reduce signal suppression, washing or desalting steps are common before freezing. By stretching the tissue, the spatial resolution of MSI was improved to enable localization of neuropeptides in individual neurons of *A. californica* (Zimmerman et al., 2011). Similarly, electroblotting through trypsin-containing membrane was found to enhance tryptic digestion directly in tissues to improve protein identification without compromising their spatial location (Andrews et al., 2021). Employing a secondary step of ionization on the expanding MALDI plume improved ion yields, boosting the sensitivity of spatial imaging of the tryptic peptides (McMillen et al., 2021).

2.2.3 | Metabolomics

MS-based metabolomics can be streamlined to deliver high reproducibility facilitating both absolute and

relative quantification of metabolites. Common sample preparation protocols in metabolomics include quenching of enzymatic activity, analyte extraction, purification, and desalting, and/or chemical derivatization before separation. To deepen metabolite coverage, detection sensitivity can be improved by tailoring the chemical composition, pH, and temperature of the extraction solvent to the chemical properties of the compound classes of interest (e.g., amino acids, lipids, nucleotides). For example, a discovery study in *Escherichia coli* culture reported that polar solvents (e.g., water, ethanol) in combination with high temperature effectively extracted hydrophobic and hydrophilic metabolites as compared to less polar solvents (e.g., acetone) (Yanes et al., 2011). Several advances for cellular and subcellular sample preparation were developed, including optimized extraction solvent mixtures (Figure 4; Onjiko et al., 2016), on-column preconcentration or sample stacking (Kawai et al., 2019; Liao et al., 2020; J. X. Liu et al., 2014), and microfluidic devices (Wu et al., 2004). Recent technologies in droplet microfluidics were reviewed elsewhere (Matula et al., 2020).

Detection sensitivity benefits from sample cleanup. These strategies aim to reduce ion suppression while minimizing the risk of concentration dilution to help distinguish analyte signals from chemical interference. Recently, six targeted metabolites were detected in MCF-7 cells without lipid interference using simplified microextraction, which used the same pulled glass capillary to dispense a droplet of solvent onto the targeted cell for selective extraction and sampling by aspiration (X. C. Zhang et al., 2016). Alternatively, an on-probe derivatization step was reported to selectively enhance the signal of target metabolites, thus reducing ion suppression (Cao, Zhang, et al., 2020). Simplified extraction methods can also be employed to broaden the types of metabolites detectable in single cells and subcellular organelles, as was accomplished for polar and non-polar metabolites in single cells using analyte partitioning within a microcapillary (Hamilton et al., 2017). The single-step method successfully decreased ion suppression during lipidomics by MALDI-MS.

As with other 'omic areas, droplet-based microfluidic chips allow limited sample handling by confining single-cell workflows to low-volume droplets. These approaches were the focus of several comprehensive reviews (Feng et al., 2015; Gao et al., 2013; Joensson & Svahn, 2012; Y. F. Liu et al., 2019; Matula et al., 2020; Shang et al., 2017). A major advantage of microfluidic methods is the resulting high throughput that enables the rapid interrogation of hundreds of cells in a single experiment (Cahill et al., 2019; X. C. Zhang et al., 2016).

For example, by combining a single-cell printer (SCP) with liquid-vortex capture (LVC)-MS (Figure 4), a recent study demonstrated direct analysis of metabolites in single cells at a rate of ~30 cells/s (Cahill et al., 2019). LVC-MS has also been combined with LCM to enable matrix-free spatially resolved direct analysis of metabolites from 20 to 40 μm tissue sections (Cahill et al., 2016). An important aspect to consider when using microfluidics for cell isolation is the possible perturbation to the cell's native metabolome due to external factors, such as mechanical forces.

Techniques enabling in situ or minimally intrusive cellular or subcellular isolation can assist in these challenges. Recent methods using, for example, sharpened or dual-barrel capillaries, reproducibly targeted and aspirated picoliter-to-nanoliter sample volumes from neurons (Aerts et al., 2014; Choi et al., 2022), *Allium cepa* cells (Huang et al., 2021; Saha-Shah et al., 2016), and *X. laevis* embryonic cells (Lombard-Banek, Choi, et al., 2021; Onjiko et al., 2017; Saha-Shah et al., 2019). Likewise, a single-lysosome MS (SLMS) platform was recently reported which combines patch-clamp recording and induced nanoESI (InESI)-MS to enable electrophysiological profiling followed by direct metabolic analysis of individual enlarged lysosomes within ~80 s (Zhu, Li, et al., 2021). The same glass micropipette used as a recording electrode served as the electrospray emitter for subsequent ESI-MS analysis. Further advances aiming to extract and measure mitochondrial, cytosolic, and nuclear metabolites can open new avenues for investigating metabolic regulation within these subcellular compartments. Overall, the fast sampling afforded by these and other capillary probes help minimize metabolic changes and stress arising from endogenous metabolism, transitions in cell cycle, or metabolic degradation, thus providing a more precise view of the cell's metabolic state (Onjiko et al., 2017).

Integration of HRMS with light microscopy, fluorescent labeling, and MSI opened the possibility of high-throughput single-cell and -organelle metabolomics. SpaceM, an open-source method for in situ single-cell metabolomics, demonstrated the detection of ~100 metabolites from more than 1000 individual cells per hour to classify cell types with high accuracy (~91%) (Rappez et al., 2021). Microscopy-guided MALDI allowed for profiling over 30,000 individual rat cerebellar cells, revealing cellular heterogeneity (Neumann et al., 2019). Image-guided MALDI on a FTICR mass spectrometer enabled surveying lipid and peptide heterogeneity between 0.5- to 2- μm -diameter dense-core and lucent vesicles in *A. californica* cells (Castro et al., 2021). As shown in Figure 4, a liquid microjunction (LMJ) probe further enabled coupling optically guided MALDI with

CE-HRMS to extend the metabolic coverage from single pancreatic islet cells (Comi et al., 2017).

Minimization or elimination of chemical matrix for the purpose of sampling or ionization facilitated higher measurement throughput and reduced chemical interferences. Under vacuum, nano laser probe-based LDI achieved a resolution of 300 nm and was able to capture the localization of anticancer drugs within organelles of a HeLa cell (Meng et al., 2020). Nanostructure-initiator MS (NIMS) (H. K. Woo et al., 2008) and silicon nanopost array (NAPA) laser desorption/ionization (Korte et al., 2019) used specialized substrates to promote desorption and ionization for high-resolution spatial imaging of tissues, cells, and organelles. SIMS has a demonstrated history of imaging small biomolecules with ~650 yoctomole sensitivity (Greving et al., 2011), with resolutions on the order of a few tens of nanometers, as was reviewed recently (Massonnet & Heeren, 2019). Early on, for example, the technology found changes in lipid domains in mating *Tetrahymena thermophila* (Kurczy et al., 2010). NAPA imaged triglycerides from tissues, which are otherwise difficult to ionize (Fincher et al., 2020). To facilitate chemical identifications, a TOF instrument equipped with dual C_{60} -SIMS and MALDI increased the impact area of molecular-ion projectiles (Lanni et al., 2014), while matrix-enhanced SIMS used a matrix as an energy sink, allowing for profiling metabolites and neuropeptides in tissues (Potocnik et al., 2017) and single cells with a high 6 s/cell measurement throughput (Do et al., 2017). These and other developments in SIMS and MALDI were the focus of recent reviews (Lanni et al., 2012; X. P. Zhu et al., 2022).

Ambient MS lifted spatial and temporal imaging to the ordinary environment. Live single-cell video-MS (Mizuno et al., 2008), DESI, and LAESI MS provided convenient detection of primary and secondary metabolites in scalable resolution from tissues to cells, as was reviewed (Nemes et al., 2010; Wiseman et al., 2006; L. W. Zhang & Khattar, Kemenes, et al., 2018). LAESI MS pioneered spatial interrogation of tissues from plants and mouse brain sections in 2D (Nemes et al., 2010, Stopka et al., 2019) and 3D (Nemes et al., 2009). Sequential tissue sectioning with ambient MSI on each section allowed DESI (Eberlin et al., 2010) and nanoDESI (Lanekoff et al., 2015) to reconstruct the 3D distribution of metabolites in the mouse brain. The ability to spatially resolve metabolic signatures from tissues to cells promises to drive new biological discoveries and applications, for example, in the study of cancer cell populations and their microenvironment. Moreover, the development of methods to measure subcellular metabolic signatures offers an avenue for new biological insights.

2.2.4 | Multi-'Omics

The integration of multiple 'omics information on the same sample provides a deeper and more complete picture of molecular processes in a biological system. Furthermore, measuring the levels of multiple types of biomolecules across space and time is necessary to better understand dynamic biological processes, such as development, which would not be captured by a single-time snapshot (Lopes et al., 2020). Protein, peptide, and metabolite levels also change as a result of intrinsic and extrinsic factors (i.e., conditions of sample collection), such that simultaneous collection of multiple biomolecules from the same samples is favorable to enhance the fidelity, robustness, and reproducibility of analyses, particularly on limited samples.

Several methods were developed to enable sample processing for multi-omics studies on the same sample. Representative examples range from simple molecular weight cutoff filtration for proteomics and peptidomics (Zougman et al., 2008) to multiphase partitioning via simultaneous metabolite, protein, lipid extraction (SIMPLEX) (Coman et al., 2016). A more recent tip-based extraction, simultaneous trapping (SiTrap), was developed to allow proteomics and metabolomics on tens of micrograms of proteins from the same cell or tissue (Zougman et al., 2020). With further miniaturization, this method could be advantageous for multi-'omic analysis of limited samples, as fewer transfer steps help to minimize sample handling and reduce analyte loss compared to traditional extractions using multiple phases. Although recent efforts introduced methods to collect samples from the same tissue or cell for dual proteomics and metabolomics (Lombard-Banek et al., 2021; Pace et al., 2021), techniques enabling simultaneous multi-'omics have yet to become routine.

2.3 | Separation

Chemical separation aids biological HRMS 'omics. Separation provides another layer of compound-dependent information, the time of separation, to aid molecular identifications. It also helps minimize or eliminate chemical interferences during ionization and detection, thus improving sensitivity and the confidence of molecular identifications. Further, by spreading the chemical complexity of the sample over time, separation enhances the efficiency of duty cycle utilization during tandem and multi-stage MS, which in turn deepens the detectable and quantifiable portion of the 'omes. This section highlights approaches, shown in Figure 5, that facilitate spatiotemporal studies on limited specimens,

ranging from cell populations to single cells to organelles, using some form of separation.

2.3.1 | LC

As a result of extensive research in stationary phase chemistries and engineered systems capable of withstanding high pressures, automation, and commercialization, LC became the mainstream separation technology for HRMS. Virtually all types of proteins, peptides, metabolites, lipids, and drug molecules can be separated using high-performance LC, or HPLC (Aydogan et al., 2020; Broeckhoven & Desmet, 2021). Reversed-phase LC (RPLC) employing packed-bed stationary phases with C_{18} chemistries (Figure 5) is efficient in resolving peptides and hydrophobic metabolites including lipids, whereas C_4 or C_8 chemistries were found to work well for intact proteins (Catherman et al., 2014). The development of hydrophilic-interaction LC (HILIC) extended analyses to small and polar molecules, including neuropeptides, amino acids, and glycans (Aydogan et al., 2020; Mihailova et al., 2008).

Combining orthogonal mechanisms of separation enhances trace-sensitive analyses. Higher theoretical plate numbers resulting from the hyphenation of multiple dimensions of separation deepens the molecular depth of detection. The multidimensional protein identification technology (MudPIT) recognized these fundamental advantages for bottom-up proteomics early on by supplementing strong cation exchange (SCX) with reversed-phase chromatography (Schirmer et al., 2003). To reduce salt load from SCX, a modification of the approach used RPLC chemistries in both dimensions, first to fractionate at high pH (e.g., pH 8–10), then to separate each fraction at the conventional low pH (e.g., pH 2–3). High pH fractionation doubled the overall depth of the detectable phosphoproteome from mouse cell lines compared to SCX, identifying over 30,000 phosphopeptides (Batth et al., 2014). For limited cell populations, this approach recently yielded 70% improvement in the coverage of the measurable proteome, allowing for the molecular characterization of the developing neural ectoderm tissue during embryogenesis (Baxi et al., 2018). Fractionation using stage tips (H. Kim et al., 2019) or nano-liter volumes (Reubsaet et al., 2020) further aided in reducing sample loss for deeper molecular coverage.

Recent advances in HPLC extended HRMS to ever-decreasing amounts of materials (Figure 5). To reduce analytical requirements to few micrograms to nanograms of proteomes, new-generation columns enhanced the peak capacity of partition chromatography by packing




FIGURE 5 Advances in separation and ionization aided the chemical analysis of limited amounts of materials. Reducing the size of packing material and diameter of LC columns, newer generation columns minimize dispersion of molecules for high sensitivity during chromatography. Diverse stacking methods, including field-amplified sample stacking (FASS) and large-volume sample stacking (LVSS), boost sensitivity during CE. Specialized technologies and methods improve ion yields during electrospray ionization by controlling the spraying regime, charging of droplets, and reducing analyte dilution. Technological abbreviations are in the text. OR, orifice of mass spectrometer; Tc, Taylor cone. Adapted with permission from Bongaerts et al. (2020); Choi et al. (2017); DeLaney, Jia, et al. (2022); Desmet et al. (2020); Kenderdine et al. (2018); Lombard-Banek, Choi, et al. (2019); Lombard-Banek, Moody, et al. (2019); Marginean et al. (2007); Moini (2007); Nemes et al. (2007). [Color figure can be viewed at [wileyonlinelibrary.com](https://onlinelibrary.wiley.com)]

narrow-bore columns (inner diameters $<50\ \mu\text{m}$) with small silica particles ($<5\ \mu\text{m}$). These columns operate at nanoflow rates ($<50\ \text{nL/min}$), thus concurrently reducing sample dilution and increasing ionization efficiency towards ultrasensitive detection. Recently, 362–874 protein groups became identifiable from a single HeLa

cell ($<150\ \text{pg}$) using flow rates of $\sim 20\ \text{nL/min}$ through an in-house-packed nanoLC column with $20\ \mu\text{m}$ inner diameter (Cong et al., 2020). With 10–100-fold improvement in sensitivity than afforded narrow-bore packed columns ($\sim 15\text{--}30\ \mu\text{m}$ inner diameter), these HPLC systems were recently coupled with microfluidic systems

capable of automating sample preparation, such as nanoPOTS (S. M. Williams et al., 2020), to address the challenge of protein loss on surfaces due to adsorption. Similarly, metabolomic coverage also benefited from reducing the flow rate from the analytical (0.5 ml/min) to the microflow (8 μ l/min) regime, especially for sample-limited studies (Geller et al., 2022).

New-generation LC columns overcame challenges associated with packing narrow columns with robust performance. As illustrated in Figure 5, open tubular, monolithic, and micropillar array columns (μ PAC) reduced band broadening and increased peak capacities using flow rates decreased to nanoliters to subnanoliters per minute (Desmet et al., 2020). The resulting enhancements in separation improved sensitivity and reproducibility for trace-level proteomics and peptidomics (Aydogan et al., 2020; Broeckhoven & Desmet, 2021; Yi et al., 2017) as well as metabolomics (Maes et al., 2014; Turiak et al., 2018; Zemenova et al., 2017). Capillary monolithic columns with 12 nl/min flow rate enabled identification of over 2300 proteins from 1 ng of HeLa digest (Gregus et al., 2020). Flows on the order of picoliters per minute through narrow open tubular columns with $\sim 2 \mu$ m inner diameter recently elevated identifications to ~ 1000 proteins from 75 pg of *Shewanella oneidensis* tryptic digest, which was equivalent to half the amount obtained from a single HeLa cell (Xiang et al., 2020). Monolithic capillary columns were used to spatially profile the proteome and phosphoproteome of the plasma and the vacuolar membrane in *Oryza sativa* (Whiteman et al., 2008). Alternatively, fabricated monolithic capillaries were used for solid phase extraction and concentration enrichment of peptides released from spatially distinct regions of *A. californica* brain tissue followed by MALDI analysis (Iannacone et al., 2009).

Improvements in HPLC-ESI interfaces played an equally important role in advancing the sensitivity of HRMS analyses. In the conventional ESI interfaces, the LC eluent is fed through a metal emitter or pulled glass capillary, and sufficient potential difference is established between the spray tip and the mass spectrometer to form a stable electrospray. Electrospays, however, can operate in different ways (Nemes et al., 2007), and transitions between these modalities can impact performance and the higher-order structure of molecules (Nemes et al., 2008). Figure 5 shows the impact the onset of different spraying regimes has on spray dynamics and MS ion signal when the ESI potential was adjusted. The relative ion signals, stability, and chemical composition of the electrospray are all impacted by alterations in the spraying regime. As illustrated, the cone-jet regime generates small droplets continuously, thus maximizing ionization

efficiency (Nemes et al., 2007). The interface called UniSpray (Figure 5) placed an impactor rod of high voltage between the grounded capillary tip and inlet of the MS to facilitate electrospray ionization of the primary electrosprayed droplets, thus boosting S/N for neuropeptides (Bongaerts et al., 2020). Following the scaling laws of ESI, emitter tips with submicron-diameters (e.g., $\sim 0.6 \mu$ m) were recently demonstrated to improve protein and oligonucleotide sensitivity by efficiently reducing their formation of cluster and adduct ions (Kenderdine et al., 2018).

2.3.2 | CE

While LC is a powerful separation method for complex biological samples, separation with CE can offer distinct advantages. Unlike HPLC, CE separates charged compounds in solution based on differences in their electrophoretic mobilities. With typical sample consumption on the order of a few nanoliters, this technology is a natural fit for volume- and mass-limited samples. Several methods allow for efficient preconcentration of analytes via in-column sample stacking (Osborn et al., 2000). Field-amplified sample stacking (FASS) and exquisite separation power permit efficient differentiation of trace amounts of proteins, peptides, and metabolites, as was recently reviewed in multiple studies (DeLaney et al., 2019; Drouin & Ramautar, 2021; Lombard-Banek, Yu, et al., 2019; Zhao et al., 2012). With large-volume sample stacking (LVSS) using a dynamic pH junction, a 10-fold enhancement in sensitivity (Figure 5) enabled the detection of 5700 proteoforms in the *E. coli* proteome (McCool et al., 2018) and peptides from the renin-angiotensin system in the paraventricular nucleus of the mouse central nervous system (DeLaney, Jia, et al., 2022). An international study encompassing 13 different laboratories found CE robust and reproducible (Drouin et al., 2020), projecting this technology to complement chromatography-based HRMS for spatiotemporal 'omics.

Ultrasensitive measurements also benefited from the innovation of specialized CE ion sources for HRMS. The development of low sheath-flow and sheathless interfaces supported the coupling of CE with ESI-HRMS, reviewed elsewhere (Gomes & Yates, 2019; Lindenburg et al., 2015). Figure 5 highlights representative CE-ESI interfaces that enabled stable and sensitive CE-ESI operation using blunt-tip and tapered-tip metal as well as microfabricated borosilicate and fused silica emitters. These custom-built CE-ESI instruments extended 'omics measurements to attomole to zeptomole to yoctomole levels of metabolites and peptides from limited samples and broad types of biological models (Bonvin et al., 2012;

Choi et al., 2017; DeLaney, Jia, et al., 2022; Hirayama et al., 2018; Moini, 2007; Peuchen et al., 2017; Vermeire et al., 2020).

Microsampling and custom-built CE-HRMS platforms enabled spatiotemporal profiling of important biomolecules. Using micropunch sampling, the technology recently permitted characterization of nearly all the angiotensin peptides of the renin-angiotensin system in the subfornical organ and paraventricular nucleus of the mouse brain with a 30-zmol (~18,000 copies) limit of detection, raising a potential to facilitate future studies on thirst and water homeostasis and regulation of the neuroendocrine system to stress (DeLaney, Jia, et al., 2022; Lombard-Banek, Yu, et al., 2019). Further, microdissection and capillary microsampling extended ultrasensitive CE-HRMS to hundreds of different metabolites (Aerts et al., 2014; Onjiko et al., 2015) and up to ~2000 different proteins in single neurons in the central nervous system of the mouse (Choi et al., 2022) as well as blastomeres in *X. laevis* and zebrafish embryos (Lombard-Banek, Moody, et al., 2019; Lombard-Banek, et al., 2021; Lombard-Banek, Moody, et al., 2016). Results from these studies revealed previously unknown molecular differences between identified embryonic cells and reorganization of the single-cell proteome and metabolome during formation of cell lineages (Lombard-Banek, Choi, et al., 2019; Onjiko et al., 2017). Microdialysis combined with CE has enabled in vivo study of cerebral metabolism of peptide E in rodent brain (Behrens & Li, 2010).

Information from these studies supported the generation of hypotheses and led to biological discoveries. CE-HRMS metabolomics on dissected identified neurons revealed distinct metabolomes in *A. californica* neurons and their regional composition (Lapainis et al., 2009; Nemes et al., 2013). CE-HRMS played a central role in the discovery of noncanonical cell fate-altering molecules and mechanisms of embryonic patterning (Onjiko et al., 2015). Most recently, the technology enabled the characterization of ~700 proteins and ~150 metabolites from ~10 nl aspirates from single-identified blastomeres in cleavage-stage *X. laevis* embryos, enabling in vivo proteomics in live embryos (Lombard-Banek et al., 2021). Using top-down HRMS, the technology also enabled detection of ~400 different proteoforms in distinct regions of the zebrafish brain, containing ~250 cells (Lubeckyj & Sun, 2022). CE-HRMS aided temporal studies on the expression of N-glycans and N-glycosylated peptides during vertebrate embryonic development, revealing extensive reorganization of the N-glycome expression during neurulation (Qu et al., 2020).

Integration of CE with orthogonal HRMS technologies opened exciting new directions in basic research. CE-HRMS with CE modes (Xu et al., 2012) and LC

fractionation (Z. C. Yang et al., 2019) deepened detection of proteomes. Coupling high pH fractionation followed by CE analysis helped identify 6500 proteins from 500 ng of protein from a human cell line. When coupled with ion mobility detection, optical isomers became distinguishable in single-cell bodies isolated from *A. californica* pleurin neurons (Mast et al., 2021). With MALDI-MS, CE enabled rapid fingerprinting of individual pancreatic islet cells with single-cell resolution, revealing metabolic differences between the different cell types (Comi et al., 2017).

Further automation and miniaturization through microfluidic chips also facilitated ultrasensitive measurements. Microfluidic CE systems expanded HRMS to ultrasensitive analysis by constraining critical steps of the analytical workflow, from sample lysis to separation, in a single platform with minimal sample transfer and, in turn, reducing sample loss. For example, an integrated microfluidic device enabled the analysis of α and β subunits of hemoglobin in 12 human erythrocytes per minute (Mellors et al., 2010). High throughput provided by this platform was used to analyze intracellular levels of metabolites in a large number of neuronal cells under different physiological conditions (X. T. Li et al., 2016). Such automated systems enabled high throughput 'omics on large populations of single cells, as was reviewed elsewhere (Abraham et al., 2019).

2.3.3 | IMS

Gas-phase ion separation provides speed and added resolution to aid molecular detection. With ion mobility separations taking less than 100 ms, ion mobility can be nested into most liquid-phase separations before HRMS detection. Most IMS techniques yield information to enable the calculation of collision cross section (CCS), which serves as compound-dependent information to aid molecular identifications. This approach enabled the elucidation of compound classes with large molecular diversity, including D- and L-amino acid-containing peptides (Fouque et al., 2017) and neuropeptides of the renin-angiotensin system in the paraventricular nucleus of the mouse central nervous system (DeLaney et al., 2022). IMS facilitated a broad range of investigations, ranging from proteins to small metabolites to lipids to glycans, as was reviewed elsewhere (Burnum-Johnson et al., 2019).

IMS techniques combined with other separation techniques, such as HPLC, can help distinguish isomers, isobars, and conformers. As a dual separation device and charge-state filter, FAIMS increased unique peptide identifications by ~30% and doubled the number of

quantifiable peptides over non-IMS from HEK293 cells due to a 10-fold enhancement in ion signal abundance and improved utilization of MS/MS duty cycles (Pfammatter et al., 2018). FAIMS was employed to discard singly charged ions to improve detection of peptides from single HeLa cells processed using nano-POTS, providing over twofold improvement (Cong et al., 2020). Trapped IMS (TIMS) enabled regio-specific profiling of novel D-amino acid-containing peptides in the *A. californica* central nervous system (Mast et al., 2020). IMS also aids deepening molecular coverage from complex biological samples by providing an additional dimension of separation, the CCS, and reducing chemical noise and isobaric interferences. HPLC with TIMS-MS provided orthogonal separation to facilitate speciation for over ~1400 lipids and capture their spatial distribution in mouse brain tissues (X. Chen et al., 2022). HPLC with traveling wave IMS (TWIMS) was used to identify metabolic changes during colorectal cancer progression (M. D. Williams et al., 2015). Distinct metabolic profiles were captured to differentiate colorectal cancer from nonmalignant tissues, and across different stages of cancer progression.

IMS provides separation in cases when classical solution-based separations are restrictive or not feasible. Such is the case in most ambient ionization experiments, where LC- or CE-based separation would hinder analytical throughput and complicate sample processing. For example, capillary microsampling coupled with TWIMS-MS enabled neuropeptide measurements from cytoplasmic and nuclear components of single neurons from *Lymnaea stagnalis*, yielding previously unavailable data for functional biological experiments on their roles (L. W. Zhang, Khattar, Kemenes, et al., 2018). Similarly, dual-probe microsampling with drift-tube IMS-MS reported 73 lipids and 79 metabolites from live mammalian microglial cells (Domenick et al., 2020). Further, IMS boosts the reproducibility and confidence of identifications and quantification in MSI (Sans et al., 2018). LAESI with IMS imaged the distribution of isobaric primary and secondary metabolites, such as flavonoids, in plant species and lipids and metabolites in mouse brain sections (H. Li et al., 2015). MALDI with TIMS-MS enabled spatial mapping of isobaric and isomeric lipids in crustacean tissues (Fu et al., 2020). Recently, ambient ionization using liquid extraction surface analysis with TWIMS allowed imaging of intact proteins from mouse kidney tissue (Hale et al., 2020).

2.4 | Mass analysis and detection

Spatial HRMS 'omics profits from a still-expanding array of technological innovations. Development of specialized

ion optics enabling efficient ion collection and transfer (e.g., ion funnel) as well as systems integrating new-generation mass analyzers, digitizers, and ion detectors (e.g., TOF, orbitrap, FTICR, SIMS) ushered in a new era of sensitivity (detection limit), speed, and spectral resolution for limited specimens. A subset of developments are shown in Figure 6. High-repetition lasers and fast electronics enabled record 50 pixels/s data acquisition rates in MALDI-TOF, sufficiently high to enable fast 2- and even 3D imaging. In a recent example, 3D-MALDI-TOF MS (Figure 6) revealed differences between the lipid architecture of metastatic and non-metastatic medulloblastoma tumors, yielding new information on disease progression (Paine et al., 2019). Customized nano-flow ESI (nanoESI) ion sources on TIMS-TOF MS (Figure 6) quadrupled the detectable ion signal, deepening proteome coverage to >3900 proteins from 1 ng of protein from single HeLa cells (Brunner et al., 2022).

Ultrahigh resolution from orbitrap and FTICR instruments supported molecular identifications and coverage of the 'omes. Late-generation orbitrap tribrid mass spectrometers increased the detectable proteome from a single HeLa cell by ~36% and ~20% at the peptide and protein level, respectively, compared to the earlier designs (Cong et al., 2020). Additionally, transferring identification based on FAIMS filtering (TIFF) permitted identification of ~1700 proteins from single HeLa cells using varying compensation voltages (J. Woo et al., 2022). A 3D SIMS instrument equipped with an orbitrap analyzer (Figure 6) permitted high-resolution spatial imaging while obtaining accurate mass-to-charge (m/z) information, revealing lipid dysregulation during macrophage differentiation (Suvannapruk et al., 2022). Microscopy-guided MALDI with FTICR MS (Figure 6) enabled profiling ca. 500 lipids from 30,000 cells from the rat cerebellum, finding previously unknown molecular cell heterogeneity in the tissue in record time (Neumann et al., 2019). In situ microsampling by fiber-optic LAESI, or f-LAESI (Figure 6), with 21 T FTICR MS tracked metabolite responses in single plant cells upon bacterial infection (Samarah et al., 2020). These and other advancements continue to expand the capabilities of spatial measurements down to the building block of life, the cell.

Spatiotemporal 'omics naturally benefited from advances in tandem MS with increasing throughput, sensitivity, and specificity. An intelligent DDA method monitored the elution order of peptides based on prior discovery measurements to target tandem MS events, resulting in 80% more protein identifications than the standard (Bailey et al., 2014). BoxCar DDA leveraged multiple narrow m/z windows to improve ion accumulation during MS scans, realizing an order of




FIGURE 6 High- to ultrahigh resolution analyzers and instrumental configurations with enhanced ion transfer efficiency support spatiotemporal 'omics. TOF analyzers are capable of fast operation, enabling spatial interrogation of molecular distributions in tissues and cells using focused laser beams (MALDI) and in situ ablation/ionization (LAESI). Orbitrap and FTICR analyzer-detector systems provide ultrahigh resolution to distinguish isobaric species and their spatial and temporal distributions. Abbreviations are in the text. Adapted with permission from Brunner et al. (2022); Cong et al. (2020); Neumann et al. (2019); Paine et al. (2019); Samarah et al. (2020); Suvannapruk et al. (2022). [Color figure can be viewed at wileyonlinelibrary.com]

magnitude improvement in quantification (Meier et al., 2018). A DDA ladder employed a nested exclusion series of abundant peptide ions to reduce MS/MS redundancy, allowing identification of 35% more proteins than standard DDA and quantification of 415 proteins from ~5 ng of protein digest (approx. 10 neurons) from the mouse nervous system (Choi et al., 2021). In parallel, instruments with multiple fragmentation cells opened the door to using structurally diagnostic fragment ions to trigger sequential or multiplexed dissociation toward deeper sequence coverage. For example, electron transfer dissociation and higher-energy collisional dissociation (EThcD) discovered 4 different N-linked and 14 distinct O-linked glycosylated neuropeptides as well as 91 novel neuropeptides in identified crustacean neural tissues (Q. J. W. Cao, Yu, et al., 2020). The advent of data-independent acquisition (DIA) remedied limited tandem MS duty cycles during DDA, approximately doubling the number of neuropeptides in limited neural tissues (DeLaney & Li, 2019). This approach detected ~1500 proteins from single cells in the developing *X. laevis* embryo (Saha-Shah et al., 2019).

Spatial HRMS provides several strategies for quantification (recall Figure 2). Both label-free quantification (LFQ) and label-based quantification were adapted to

limited specimens in the context of spatiotemporal biology. LFQ uses spectral counts, peak intensities, or peak areas to estimate the abundance of proteins, peptides, and metabolites without barcoding (Dupree et al., 2020; Lai et al., 2013). A reduction in sample processing steps aids analytics on trace amounts of samples, as was reviewed recently (Slavov, 2021). Chemical derivatization helps analytical throughput and reproducibility, and when performed online, enabled low-picomolar detection limits for peptides in microdialysates continuously collected from rat brains (Wilson et al., 2018). To enhance throughput, label-based quantification may integrate stable isotopes during sample preparation in vivo (metabolic labeling) or in vitro (chemical labeling) to enable multiplexing, as was reviewed elsewhere (Arul & Robinson, 2019).

Label-based quantification became quasi-routine for high throughput quantification across broad spatial and temporal processes, lately even for single cells. Reductive methylation (DeLaney et al., 2021) and DiLeu labeling (Sauer & Li, 2021) aided neuropeptidomics. The carrier approach pioneered the use of tandem mass tags (Thermo Fisher Scientific) in abundance to boost tandem MS on isobarically tagged protein digests from single cells (Slavov, 2022), while usage of a sacrificial protein helped reduce nonspecific protein losses during sample

preparation (Slavov, 2021). SCoPE-MS (recall Figure 4) is one such method, which distinguished cancer cell types at the single-cell level and characterized cell-to-cell heterogeneity during embryonic stem cell differentiation (Budnik et al., 2018). The new-generation SCoPE2 further downscaled volumes, automated sample preparation, increased throughput, and optimized MS data acquisition and analysis to quantify ~1000 proteins per TMT experiment with only 1 h of LC separation (Specht et al., 2021). Integration of carrier-based approaches with newer instrumentation further compounded the demonstrated improvements in proteome coverage and quantification from single cells. For example, FAIMS and real-time search in both MS/MS and synchronous precursor selection (SPS)-MS³ methods on an orbitrap mass spectrometer attained higher proteome coverage and more accurate quantification in single acute myeloid leukemia cells by lengthening ion trap times, thus boosting sensitivity (Furtwängler et al., 2022). Combining narrow-window DIA acquisition with TMT tagging yielded high throughput multiplexing for trace samples (Ctortecka, Krssakova, et al., 2022). Additional improvements in efficient duty cycle utilization, sensitivity, and real-time search capabilities made possible by continued technological advances will further enhance performance of these carrier-based quantification methods for proteomics analysis of limited samples.

2.5 | Data analysis

Bioinformatics plays a critical and enabling role for spatiotemporal HRMS. This field of HRMS yields challenging data, not only in terms of complexity and size but also metadata to assess in relation to space and time. In Figure 7, representative examples are grouped based on their abilities to access these orthogonal pieces of information. General steps for data analysis include raw data processing (feature extraction or identification), data preprocessing, statistical analysis (quantitative information), and annotation and molecular network analysis (qualitative information). Data from imaging microscopy (optical to electron) and functional tests, such as electrophysiology and behavioral assays, complement the chemical data from MS 'omics with information on the biological phenotype. Analysis of such complex data poses several challenges such as handling large data sets, integrating orthogonal information obtained by combining different data acquisition techniques, accuracy of identification, and data processing speeds. Here we highlight select examples of software packages that helped address these challenges.

The first step of data analysis includes processing raw files obtained from MS analysis to identify proteins, peptides, or metabolic features. With ever-decreasing sample amounts to achieve spatial resolution at cellular and subcellular level, distinguishing true signal from noise is a fundamental challenge. New and next-generation bioinformatics aided the processing of low ion signals that could be mistaken for noise (false negative) or noise that could be mistaken for signal (false positive). The *Trace* software package employed machine learning on CE-HRMS data sets to create signal images from *m/z* and separation time data. These signal images effectively recognized true signal from noise with a record 95% confidence in single-cell metabolomics (Liu, Portero, et al., 2019). NPFimg recently extended the approach to metabolomics based on gas chromatography-MS (Jirayupat et al., 2021). OffsampleAI employed artificial intelligence for spatial metabolomics to recognize background interfering ions during MSI experiments. A machine learning tool was trained using ~2300 manually curated ion images to recognize background ions, thus improving metabolite identification from tissues (Ovchinnikova et al., 2020).

In parallel, a series of software developments increased speed, accuracy, and depth of coverage of the analyzed 'omes. To improve identifications from complex convoluted DIA data sets, newer machine learning tools, such as DeepMass (Tiwarly et al., 2019), accurately predicted peptide fragment spectra to build spectral libraries in silico, essentially eliminating restrictions on laborious experimental reference libraries from precious, rare, or limited specimens. Similarly, PRESnovo, a de novo sequencing tool developed for neuropeptidomics, directly predicted peptide sequences (DeLaney et al., 2020), leading to the identification of 13 novel, putative neuropeptides from *Carcinus maenas* and 77 from *Cancer borealis* brain tissues. Most recently, several software tools were developed to overcome limitations in the processing of complex DIA and ion mobility data necessitating enhanced computational resources, which were recently discussed elsewhere (Allison et al., 2020; Rosenberger et al., 2017; Szymanska et al., 2016; F. F. Zhang et al., 2020).

As in other areas of HRMS, scientific rigor calls for stringent quality controls not just during measurements but also during data processing. After extracting qualitative (molecular identifications) and quantitative (abundance or intensities) information, data preprocessing uses normalization and advanced data checks to disentangle biological variations from artifacts (Slavov, 2022). Programs such as SCPCompanion helped credential the quality of single-cell proteomics data (Cheung et al., 2021) and suggest optimization of instrument




FIGURE 7 Chemometrics plays a pivotal role in analyzing metadata from spatiotemporal HRMS. Image analysis by optical microscopy and functional measurements, such as electrophysiology, help identify tissues and cells, guiding subsequent molecular characterizations to transcripts, proteins, peptides, and metabolites. Statistical and multivariate data analysis and knowledge bases are integral to interpreting molecular processes in specific biological contexts, as illustrated here for dissected rat sciatic nerves (HPLC), temporally developing whole *Xenopus laevis* embryos (HPLC), and dense core vesicles in *Autographa californica* (MALDI). Working in vivo, LAESI-MSI imaged chlorophyll and secondary metabolites in the variegated leaves of *Aphelandra squarrosa* and CE-HRMS enabled metabolomics-proteomics in single differentiating cells in *X. laevis* and neurons in the mouse brain. Models of artificial intelligence (AI) and machine learning (ML) assist in extracting trace-level signals (Trace) or uncovering relationships in the metadata. Technological abbreviations are in the text. Adapted with permission from (Budnik et al., 2018; Castro et al., 2021; Choi et al., 2022; Di Paolo et al., 2021; Y. F. Liu et al., 2019; Nemes et al., 2009; Onjiko et al., 2017; Peshkin et al., 2015). [Color figure can be viewed at [wileyonlinelibrary.com](https://onlinelibrary.wiley.com)]

parameters based on the preliminary data. In addition, single-cell proteomics readout of expression (SCeptre) and NOREVA provided normalization strategies to assist with reducing nonbiological variations in multiplexed single-cell 'omics (B. Li et al., 2017; Schoof et al., 2021). SpectralAnalysis was adapted to multiple instruments and provides a framework from data preprocessing to multivariate analysis within a single software package (Race et al., 2016).

Figure 7 represents examples of spatial profiling to imaging by biological HRMS. Chemometrics is crucial for the interpretation of metadata resulting from spatiotemporal HRMS (recall Figure 2). This field of research offers various solutions for statistical and multivariate analyses. For example, volcano plots are widely used to examine molecular differences between spatially or temporally distinct sample types. Unsupervised and supervised dimensionality reduction and clustering algorithms such as principal component analysis (PCA), linear discriminant analysis (LDA), t-distributed stochastic neighbor embedding (t-SNE), hierarchical cluster analysis (HCA), heatmaps, k- and fuzzy-mean clustering are broadly adaptable to exploring systematic molecular differences between samples, such as types of tissues or cells. Sciatic nerves dissected from the rat were surveyed using unsupervised HCA with LC-HRMS (Di Paolo et al., 2021), neuropeptides in dense core vesicles in *A. californica* neurons using unsupervised PCA/t-SNE with 2 D MALDI MSI (Castro et al., 2021), and chlorophyll and secondary metabolites in areas of the variegated *A. squarrosa* leaves with 3 D LAESI MSI (Nemes et al., 2009). Pearson correlation analysis and LC-HRMS permitted comparison of temporal expression between transcription and translation at the level of whole developing *X. laevis* embryos (Peshkin et al., 2015), single differentiating embryonic stem cells (Budnik et al., 2018). Live stereomicroscopy and micro-probe CE-MS used fuzzy means cluster analysis to find reorganization of the single-cell proteome as cells formed a neural-tissue-fated clone in the live *X. laevis* embryo (Onjiko et al., 2017). STRING was used to predict functional association networks on proteins that were quantified in single dopaminergic neurons in the mouse brain after electrophysiology using patch CE-MS (Choi et al., 2022). Software packages like MetaboAnalyst (Pang et al., 2021), Perseus (Tyanova et al., 2016), Orange (Demšar et al., 2013), and IOAT (Wu et al., 2021) provide user-friendly platforms for performing such analyses. EXIMS (Wijetunge et al., 2015) is another such tool developed to recognize spatial distribution patterns in MSI data sets.

A broadening number of software tools enable systems biology for spatiotemporal HRMS. Apart from

classical tools from chemometrics, freeware programs, such as MetaboAnalyst, helped integrate transcriptomic to proteomic to metabolomic data sets, with added capabilities for enrichment studies and analyses of pathway networks. STATegra, another open-source tool, merged and identified important features from multi-'omics data in a stepwise manner (Planell et al., 2021). Similarly, Mergeomics tailored to extracting information on key biological pathways and networks, with demonstrated utility in aiding the understanding of disease progression (Ding et al., 2021). These and other statistical and data visualization pipelines help generate and test new hypotheses at the levels of proteomes, peptidomes, and metabolomes in biological processes occurring in space and time.

3 | CONCLUSIONS

Interdisciplinary efforts molded HRMS 'omics into a powerful tool to advance basic and translational research for molecular processes occurring in space and time. These developments are the result of still rapidly progressing interdisciplinary and collaborative efforts innovating essentially all steps of biological HRMS. Inventions of specialized strategies, equipment, and miniaturization to collect and process samples with speed and high fidelity is pivotal for studies on limited and dynamic biological systems. Advances in chemical separations, ionization, ion transfer and accumulation, and HRMS detection increase detection sensitivity and measurement throughput, thus ushering 'omics to increasingly finer spatial and temporal resolutions. Continued developments in data acquisition and user-friendly software packages to handle the resulting complex data promise to further deepen the detectable and quantifiable portion of proteomes, peptidomes, and metabolomes from trace amounts of specimens. Grand challenges associated with the acquisition, integration, and distribution of complex data from spatiotemporal multi-'omics pose exciting opportunities to move science forward toward a holistic understanding of life processes.

Sustaining the scientific momentum is as important as using the new knowledge that we learn to best benefit society at large. Collection of proven protocols, formulation of community standards and best practices, and deposition of primary and secondary data into public data repositories—such as pioneering works in metabolomics (Sansone et al., 2007), proteomics (Taylor et al., 2008), MS informatics (Binz et al., 2008), and now also in single-cell analyses (Gatto et al., 2022)—is pivotal for stringent, reproducible, and rigorous science. Investigators continue to break down classical boundaries in

bioanalyses to reach ever-deeper into the molecular world. In parallel, it is equally important to make these developments available and adaptable to broad users, including but not limited to investigators in the basic life sciences, translational research, pharmaceutical industry, and healthcare. Furnishing methods with versatility and compatibility across scientific disciplines, automation, and cost-efficiency are necessary and proven paths toward addressing pressing biological questions at present and to pose new ones to continue leading us to the future.

AUTHOR CONTRIBUTIONS

All authors contributed to the writing and revision of this work.

ACKNOWLEDGMENTS

Parts of this work were sponsored by the National Institute of Health (award no. 1R35GM124755 to P. N.), the National Science Foundation (award no. IOS-1832968 to P. N.), the University of Maryland–National Cancer Institute Partnership Program (seed award to P. N.), and the COSMOS Club Foundation (research award to L. R. P.).

REFERENCES

- Abouleila Y, Onidani K, Ali A, Shoji H, Kawai T, Lim CT, Kumar V, Okaya S, Kato K, Hiyama E, Yanagida T, Masujima T, Shimizu Y, Honda K. 2019. Live single-cell mass spectrometry reveals cancer-specific metabolic profiles of circulating tumor cells. *Cancer Sci.* 110:697-706.
- Abraham DH, Anttila MM, Gallion LA, Petersen BV, Proctor A, Allbritton NL. 2019. Design of an automated capillary electrophoresis platform for single-cell analysis. In: Allbritton NL, Kovarik ML, Editors. *Enzyme Activity in Single Cells*. London: Academic Press Ltd-Elsevier Science Ltd. p 191-221.
- Aerts JT, Louis KR, Crandall SR, Govindaiah G, Cox CL, Sweedler JV. 2014. Patch clamp electrophysiology and capillary electrophoresis-mass spectrometry metabolomics for single cell characterization. *Anal. Chem.* 86:3203-3208.
- Allison TM, Barran P, Benesch JLP, Cianferani S, Degiacomi MT, Gabelica V, Grandori R, Marklund EG, Menneteau T, Migas LG, Politis A, Sharon M, Sobott F, Thalassinou K. 2020. Software requirements for the analysis and interpretation of native ion mobility mass spectrometry data. *Anal. Chem.* 92:10881-10890.
- Andrews WT, Bickner AN, Tobias F, Ryan KA, Bruening ML, Hummon AB. 2021. Electroblobbing through enzymatic membranes to enhance molecular tissue imaging. *J. Am. Soc. Mass Spectrom.* 32:1689-1699.
- Arnadottir SS, Mattesen TB, Vang S, Madsen MR, Madsen AH, Birkbak NJ, Bramsen JB, Andersen CL. 2020. Transcriptomic and proteomic intra-tumor heterogeneity of colorectal cancer varies depending on tumor location within the colorectum. *PLoS One* 15:18.
- Arnold PK, Jackson BT, Paras KI, Brunner JS, Hart ML, Newsom OJ, Alibeckoff SP, Endress J, Drill E, Sullivan LB, Finley LWS. 2022. A non-canonical tricarboxylic acid cycle underlies cellular identity. *Nature* 603:477-481.
- Arul AB, Robinson RAS. 2019. Sample multiplexing strategies in quantitative proteomics. *Anal. Chem.* 91:178-189.
- Axel M, Ewelina K, Jenny-Maria B, Leif K. 2017. An online SPE LC-MS/MS method for the analysis of antibiotics in environmental water. *Environ. Sci. Pollut. Res.* 24:8692-8699.
- Aydogan C, Rigano F, Krcmova LK, Chung DS, Macka M, Mondello L. 2020. Miniaturized LC in molecular omics. *Anal. Chem.* 92:11485-11497.
- Azizian NG, Sullivan DK, Nie LT, Pardo S, Molleur D, Chen JJ, Weintraub ST, Li YL. 2021. Selective labeling and identification of the tumor cell proteome of pancreatic cancer in vivo. *J. Proteome Res.* 20:858-866.
- Bailey DJ, McDevitt MT, Westphall MS, Pagliarini DJ, Coon JJ. 2014. Intelligent data acquisition blends targeted and discovery methods. *J. Proteome Res.* 13:2152-2161.
- Bardsen K, Gjerstad MD, Partinen M, Kvivik I, Tjensvol AB, Ruoff P, Omdal R, Brede C. 2019. Considerably lower levels of hypocretin-1 in cerebrospinal fluid is revealed by a novel mass spectrometry method compared with standard radioimmunoassay. *Anal. Chem.* 91:9323-9329.
- Batth TS, Francavilla C, Olsen JV. 2014. Off-line high-pH reversed-phase fractionation for in-depth phosphoproteomics. *J. Proteome Res.* 13:6176-6186.
- Baxi AB, Lombard-Banek C, Moody SA, Nemes P. 2018. Proteomic characterization of the neural ectoderm fated cell clones in the *Xenopus laevis* embryo by high-resolution mass spectrometry. *ACS Chem. Neurosci.* 9:2064-2073.
- Baxi AB, Pade LR, Nemes P. 2021. Mass spectrometry-based proteomics for developmental neurobiology in the amphibian *Xenopus laevis*. In: Sokol SY, Editor *Amphibian Models of Development and Disease*. San Diego: Elsevier Academic Press Inc. p 205-231.
- Bechtel TJ, Li C, Kisty EA, Maurais AJ, Weerapana E. 2020. Profiling cysteine reactivity and oxidation in the endoplasmic reticulum. *ACS Chem. Biol.* 15:543-553.
- Behrens HL, Li LJ. 2010. Monitoring neuropeptides in vivo via microdialysis and mass spectrometry. In: Soloviev M, Editor *Peptidomics: Methods and Protocols*. Totowa: Humana Press Inc. p 57-73.
- Bensen RC, Standke SJ, Colby DH, Kothapalli NR, Le-McClain AT, Patten MA, Tripathi A, Heinlen JE, Yang ZB, Burgett AWG. 2021. Single-cell mass spectrometry quantification of anticancer drugs: proof of concept in cancer patients. *ACS Pharmacol. Transl. Sci.* 4:96-100.
- Biesemann C, Gronborg M, Luquet E, Wichert SP, Bernard V, Bungers SR, Cooper B, Varoqueaux F, Li LY, Byrne JA, Urlaub H, Jahn O, Brose N, Herzog E. 2014. Proteomic screening of glutamatergic mouse brain synaptosomes isolated by fluorescence-activated sorting. *EMBO J.* 33:157-170.
- Binz PA, Barkovich R, Beavis RC, Creasy D, Horn DM, Julian RK, Seymour SL, Taylor CF, Vandenbrouck Y. 2008. Guidelines for reporting the use of mass spectrometry informatics in proteomics. *Nat. Biotechnol.* 26:862.
- Bird RE, Lemmel SA, Yu X, Zhou QA. 2021. Bioorthogonal chemistry and its applications. *Bioconjugate Chem.* 32:2457-2479.

- Bongaerts J, Segers K, Van Oudenhove L, Van Wanseele Y, Van Hulle M, De Bundel D, Mangelings D, Smolders I, Vander Heyden Y, Van Eeckhaut A. 2020. A comparative study of UniSpray and electrospray sources for the ionization of neuropeptides in liquid chromatography-tandem mass spectrometry. *J. Chromatogr. A* 1628:14.
- Bonvin G, Veuthey JL, Rudaz S, Schappler J. 2012. Evaluation of a sheathless nanospray interface based on a porous tip sprayer for CE-ESI-MS coupling. *Electrophoresis* 33:552-562.
- Broeckhoven K, Desmet G. 2021. Advances and innovations in liquid chromatography stationary phase supports. *Anal. Chem.* 93:257-272.
- Brown HM, Pirro V, Cooks RG. 2018. From DESI to the MasSpec Pen: ambient ionization mass spectrometry for tissue analysis and intrasurgical cancer diagnosis. *Clin. Chem.* 64:628-630.
- Brunner AD, Thielert M, Vasilopoulou C, Ammar C, Coscia F, Mund A, Hoerning OB, Bache N, Apalategui A, Lubeck M, Richter S, Fischer DS, Raether O, Park MA, Meier F, Theis FJ, Mann M. 2022. Ultra-high sensitivity mass spectrometry quantifies single-cell proteome changes upon perturbation. *Mol. Syst. Biol.* 18:15.
- Budnik B, Levy E, Harmange G, Slavov N. 2018. SCoPE-MS: mass spectrometry of single mammalian cells quantifies proteome heterogeneity during cell differentiation. *Genome Biol.* 19:12.
- Burke M, Choksawangkarn W, Edwards N, Ostrand-Rosenberg S, Fenselau C. 2014. Exosomes from myeloid-derived suppressor cells carry biologically active proteins. *J. Proteome Res.* 13: 836-843.
- Burnum-Johnson KE, Zheng X, Dodds JN, Ash J, Fourches D, Nicora CD, Wendler JP, Metz TO, Waters KM, Jansson JK, Smith RD, Baker ES. 2019. Ion mobility spectrometry and the omics: distinguishing isomers, molecular classes and contaminant ions in complex samples. *Trends Anal. Chem.* 116: 292-299.
- Cadwell CR, Palasantza A, Jiang XL, Berens P, Deng QL, Yilmaz M, Reimer J, Shen S, Bethge M, Tolias KF, Sandberg R, Tolias AS. 2016. Electrophysiological, transcriptomic and morphologic profiling of single neurons using Patch-seq. *Nat. Biotech.* 34: 199-203.
- Cahill JF, Kertesz V, Weiskittel TM, Vavrek M, Freddo C, Van Berkel GJ. 2016. Online, absolute quantitation of propranolol from spatially distinct 20- and 40- μ m dissections of brain, liver, and kidney thin tissue sections by laser microdissection-liquid vortex capture-mass spectrometry. *Anal. Chem.* 88:6026-6034.
- Cahill JF, Riba J, Kertesz V. 2019. Rapid, untargeted chemical profiling of single cells in their native environment. *Anal. Chem.* 91:6118-6126.
- Cao QJW, Yu Q, Liu Y, Chen ZW, Li LJ. 2020. Signature-ion-triggered mass spectrometry approach enabled discovery of N- and O-linked glycosylated neuropeptides in the crustacean nervous system. *J. Proteome Res.* 19:634-643.
- Cao YQ, Zhang L, Zhang J, Guo YL. 2020. Single-cell on-probe derivatization-noncontact nanocarbon fiber ionization: unraveling cellular heterogeneity of fatty alcohol and sterol metabolites. *Anal. Chem.* 92:8378-8385.
- Castro DC, Xie YR, Rubakhin SS, Romanova EV, Sweedler JV. 2021. Image-guided MALDI mass spectrometry for high throughput single-organelle characterization. *Nat. Meth.* 18:1233-1238.
- Catherman AD, Skinner OS, Kelleher NL. 2014. Top-down proteomics: facts and perspectives. *Biochem. Biophys. Res. Commun.* 445:683-693.
- Cervenka J, Tyleckova J, Skalninkova HK, Kepkova KV, Poliakh I, Valekova I, Pfeiferova L, Kolar M, Vaskovicova M, Pankova T, Vodicka P. 2021. Proteomic characterization of human neural stem cells and their secretome during in vitro differentiation. *Front. Cell. Neurosci.* 14:20.
- Chappell J, Boward B, Dalton S. 2016. Expanding the utility of FUCCI reporters using FACS-based 'omics analysis. In: Turksen K, Editor *Embryonic Stem Cell Protocols*. Totowa: Humana Press Inc. p 101-110.
- Chatterji B, Pich A. 2013. MALDI imaging mass spectrometry and analysis of endogenous peptides. *Expert Rev. Proteomics* 10: 381-388.
- Chen L, Yang RR, Kwan T, Tang C, Watt S, Zhang YM, Bourque G, Ge B, Downes K, Frontini M, Ouwehand WH, Lin JW, Soranzo N, Pastinen T, Chen L. 2020. Paired rRNA-depleted and polyA-selected RNA sequencing data and supporting multi-omics data from human T cells. *Sci. Data* 7:376.
- Chen Q, Yan GQ, Gao MX, Zhang XM. 2015. Ultrasensitive proteome profiling for 100 living cells by direct cell injection, online digestion and nano-LC-MS/MS analysis. *Anal. Chem.* 87:6674-6680.
- Chen RB, Ma MM, Hui LM, Zhang J, Li LJ. 2009. Measurement of neuropeptides in crustacean hemolymph via MALDI mass spectrometry. *J. Am. Soc. Mass Spectrom.* 20:708-718.
- Chen X, Yin YD, Luo MD, Zhou ZW, Cai YP, Zhu ZJ. 2022. Trapped ion mobility spectrometry-mass spectrometry improves the coverage and accuracy of four-dimensional untargeted lipidomics. *Anal. Chim. Acta* 1210:9.
- Cheung TK, Lee CY, Bayer FP, McCoy A, Kuster B, Rose CM. 2021. Defining the carrier proteome limit for single-cell proteomics. *Nat. Meth.* 18:76-83.
- Choi SB, Munoz-Llanca P, Manzini MC, Nemes P. 2021. Data-dependent acquisition ladder for capillary electrophoresis mass spectrometry-based ultrasensitive (neuro)proteomics. *Anal. Chem.* 93:15964-15972.
- Choi SB, Polter AM, Nemes P. 2022. Patch-clamp proteomics of single neurons in tissue using electrophysiology and sub-cellular capillary electrophoresis mass spectrometry. *Anal. Chem.* 94:1637-1644.
- Choi SB, Zamarbide M, Manzini MC, Nemes P. 2017. Tapered-tip capillary electrophoresis nano-electrospray ionization mass spectrometry for ultrasensitive proteomics: the mouse cortex. *J. Am. Soc. Mass Spectrom.* 28:597-607.
- Coman C, Solari FA, Hentschel A, Sickmann A, Zahedi RP, Ahrends R. 2016. Simultaneous metabolite, protein, lipid extraction (SIMPLEX): a combinatorial multimolecular omics approach for systems biology. *Mol. Cell. Proteomics* 15: 1453-1466.
- Comi TJ, Makurath MA, Philip MC, Rubakhin SS, Sweedler JV. 2017. MALDI MS guided liquid microjunction extraction for capillary electrophoresis-electrospray ionization ms analysis of single pancreatic islet cells. *Anal. Chem.* 89:7765-7772.
- Cong YZ, Liang YR, Motamedchaboki K, Huguet R, Truong T, Zhao R, Shen YF, Lopez-Ferrer D, Zhu Y, Kelly RT. 2020. Improved single-cell proteome coverage using narrow-bore

- packed nanoLC columns and ultrasensitive mass spectrometry. *Anal. Chem.* 92:2665-2671.
- Ctorteccka C, Hartlmayr D, Seth A, Mendjan S, Tourniaire G, Mechtler K. 2022. An automated workflow for multiplexed single-cell proteomics sample preparation at unprecedented sensitivity. *bioRxiv* 2021:439828.
- Ctorteccka C, Krssakova G, Stejskal K, Penninger JM, Mendjan S, Mechtler K, Stadlmann J. 2022. Comparative proteome signatures of trace samples by multiplexed data-independent acquisition. *Mol. Cell. Proteomics* 21:11.
- Ctorteccka C, Mechtler K. 2021. The rise of single-cell proteomics. *Anal. Sci. Adv.* 2:84-94.
- Dagley LF, Infusini G, Larsen RH, Sandow JJ, Webb AI. 2019. Universal solid-phase protein preparation (USP³) for bottom-up and top-down proteomics. *J. Proteome Res.* 18: 2915-2924.
- Davey HM, Kell DB. 1996. Flow cytometry and cell sorting of heterogeneous microbial populations: the importance of single-cell analyses. *Microbiol. Rev.* 60:641-696.
- DeLaney K, Buchberger AR, Atkinson L, Grunder S, Mousley A, Li LJ. 2018. New techniques, applications and perspectives in neuropeptide research. *J. Exp. Biol.* 221:16.
- DeLaney K, Cao WF, Ma YD, Ma MM, Zhang YZ, Li LJ. 2020. PRESnovo: prescreening prior to de novo sequencing to improve accuracy and sensitivity of neuropeptide identification. *J. Am. Soc. Mass Spectrom.* 31:1358-1371.
- DeLaney K, Hu MZ, Hellenbrand T, Dickinson PS, Nusbaum MP, Li LJ. 2021. Mass spectrometry quantification, localization, and discovery of feeding-related neuropeptides in *Cancer borealis*. *ACS Chem. Neurosci.* 12:782-798.
- DeLaney K, Jia D, Iyer L, Yu Z, Choi SB, Marvar PJ, Nemes P. 2022. Microanalysis of brain angiotensin peptides using ultrasensitive capillary electrophoresis trapped ion mobility mass spectrometry. *Anal. Chem.* 94:9018-9025. <https://doi.org/10.1021/acs.analchem.2c01062>
- DeLaney K, Li LJ. 2019. Data independent acquisition mass spectrometry method for improved neuropeptidomic coverage in crustacean neural tissue extracts. *Anal. Chem.* 91:5150-5158.
- DeLaney K, Phetsanthad A, Li LJ. 2022. Advances in high-resolution MALDI mass spectrometry for neurobiology. *Mass Spectrom. Rev.* 41:194-214.
- DeLaney K, Sauer CS, Vu NQ, Li LJ. 2019. Recent advances and new perspectives in capillary electrophoresis-mass spectrometry for single cell "omics". *Molecules* 24:21.
- Demšar J, Curk T, Erjavec A, Gorup C, Hočevár T, Milutinović M, Možina M, Polajnar M, Toplak M, Starič A, Štajdohar M, Umek L, Žagar L, Žbontar J, Žitnik M, Zupan B. 2013. Orange: data mining toolbox in Python. *J. Mach. Learn. Res.* 14: 2349-2353.
- Desmet G, de Beeck JO, Van Raemdonck G, Van Mol K, Claerebout B, Van Landuyt N, Jacobs P. 2020. Separation efficiency kinetics of capillary flow micro-pillar array columns for liquid chromatography. *J. Chromatogr. A* 1626:11.
- Di Paolo A, Farias J, Garat J, Macklin A, Ignatchenko V, Kislinger T, Silveira JS. 2021. Rat sciatic nerve axoplasm proteome is enriched with ribosomal proteins during regeneration processes. *J. Proteome Res.* 20:2506-2520.
- Ding J, Blencowe M, Nghiem T, Ha SM, Chen YW, Li GY, Yang X. 2021. Mergeomics 2.0: a web server for multi-omics data integration to elucidate disease networks and predict therapeutics. *Nucleic Acids Res.* 49:W375-W387.
- Do TD, Comi TJ, Dunham SJB, Rubakhin SS, Sweedler JV. 2017. Single-cell profiling using ionic liquid matrix-enhanced secondary ion mass spectrometry for neuronal cell type differentiation. *Anal. Chem.* 89:3078-3086.
- Domenick TM, Vedam-Mai V, Yost RA. 2020. Design and implementation of a dual-probe microsampling apparatus for the direct analysis of adherent mammalian cells by ion mobility-mass spectrometry. *Anal. Chem.* 92:12055-12061.
- Drouin N, Ramautar R. 2021. Capillary electrophoresis-mass spectrometry for metabolomics: possibilities and perspectives. In: Simionato AVC, Editor *Separation Techniques Applied to Omics Sciences: From Principles to Relevant Applications*. Cham: Springer International Publishing Ag. p 159-178.
- Drouin N, van Mever M, Zhang W, Tobolkina E, Ferre S, Servais AC, Gou MJ, Nyssen L, Fillet M, Lageveen-Kammeijer GSM, Nouta J, Chetwynd AJ, Lynch I, Thorn JA, Meixner J, Lossner C, Taverna M, Liu S, Tran NT, Francois Y, Lechner A, Nehm R, Banni GAD, Nasreddine R, Colas C, Lindner HH, Faserl K, Neuss C, Nelke M, Lammerer S, Perrin C, Bich-Muracciole C, Barbas C, Gonzalvez AL, Guttman A, Szigeti M, Britz-McKibbin P, Kroezen Z, Shanmuganathan M, Nemes P, Portero EP, Hankemeier T, Codesido S, Gonzalez-Ruiz V, Rudaz S, Ramautar R. 2020. Capillary electrophoresis-mass spectrometry at trial by Metabo-ring: effective electrophoretic mobility for reproducible and robust compound annotation. *Anal. Chem.* 92: 14103-14112.
- Dupree EJ, Jayathirtha M, Yorkey H, Mihasan M, Petre BA, Dark CC. 2020. A critical review of bottom-up proteomics: the good, the bad, and the future of this field. *Proteomes* 8: 26.
- Eberlin LS, Ifa DR, Wu C, Cooks RG. 2010. Three-dimensional visualization of mouse brain by lipid analysis using ambient ionization mass spectrometry. *Angew. Chem. Int. Ed.* 49: 873-876.
- Espino JA, Jones LM. 2019. Illuminating biological interactions with in vivo protein footprinting. *Anal. Chem.* 91:6577-6584.
- Ezzoukhry Z, Henriët E, Cordelieres FP, Dupuy JW, Maitre M, Gay N, Di-Tommaso S, Mercier L, Goetz JG, Peter M, Bard F, Moreau V, Raymond AA, Saltel F. 2018. Combining laser capture microdissection and proteomics reveals an active translation machinery controlling invadosome formation. *Nat. Commun.* 9:11.
- Feng XJ, Liu BF, Li JJ, Liu X. 2015. Advances in coupling microfluidic chips to mass spectrometry. *Mass Spectrom. Rev.* 34:535-557.
- Feng XJ, Liu X, Luo QM, Liu BF. 2008. Mass spectrometry in systems biology: an overview. *Mass Spectrom. Rev.* 27:635-660.
- Fincher JA, Korte AR, Dyer JE, Yadavilli S, Morris NJ, Jones DR, Shanmugam VK, Pirlo RK, Vertes A. 2020. Mass spectrometry imaging of triglycerides in biological tissues by laser desorption ionization from silicon nanopost arrays. *J. Mass Spectrom.* 55:8.
- Fischer R, Kessler BM. 2015. Gel-aided sample preparation (GASP)—A simplified method for gel-assisted proteomic sample generation from protein extracts and intact cells. *Proteomics* 15:1224-1229.

- Fouque KJD, Garabedian A, Porter J, Baird M, Pang XQ, Williams TD, Li LJ, Shvartsburg A, Fernandez-Lima F. 2017. Fast and effective ion mobility-mass spectrometry separation of D-aAmino-acid-containing peptides. *Anal. Chem.* 89: 11787-11794.
- Fu TT, Oetjen J, Chapelle M, Verdu A, Szesny M, Chaumot A, Degli-Esposti D, Geffard O, Clement Y, Salvador A, Ayciriex S. 2020. In situ isobaric lipid mapping by MALDI-ion mobility separation-mass spectrometry imaging. *J. Mass Spectrom.* 55:6.
- Fujii T, Matsuda S, Tejedor ML, Esaki T, Sakane I, Mizuno H, Tsuyama N, Masujima T. 2015. Direct metabolomics for plant cells by live single-cell mass spectrometry. *Nat. Protoc.* 10: 1445-1456.
- Furtwängler B, Üresin N, Motamedchaboki K, Huguet R, Lopez-Ferrer D, Zabrouskov V, Porse BT, Schoof EM. 2022. Real-time search-assisted acquisition on a Tribrid mass spectrometer improves coverage in multiplexed single-cell proteomics. *Mol. Cell. Proteomics* 21:100219.
- Gao D, Liu HX, Jiang YY, Lin JM. 2013. Recent advances in microfluidics combined with mass spectrometry: technologies and applications. *Lab Chip* 13:3309-3322.
- Gatto L, Aebersold R, Cox J, Demichev V, Derks J, Emmott E, Franks AM, Ivanov AR, Kelly RT, Khoury L, Leduc A, MacCoss MJ, Nemes P, Perlman DH, Petelski AA, Rose CM, Schoof EM, Van Eyk J, Vanderaa C, Yates III JR, Slavov N. 2022. Initial recommendations for performing, benchmarking, and reporting single-cell proteomics experiments, *Nat. Methods*. arXiv. in print. <https://doi.org/10.48550/arXiv.2207.10815>
- Geiger T, Wehner A, Schaab C, Cox J, Mann M. 2012. Comparative proteomic analysis of eleven common cell lines reveals ubiquitous but varying expression of most proteins. *Mol. Cell. Proteomics* 11:1-11.
- Geladaki A, Britovsek NK, Breckels LM, Smith TS, Vennard OL, Mulvey CM, Crook OM, Gatto L, Lilley KS. 2019. Combining LOPIT with differential ultracentrifugation for high-resolution spatial proteomics. *Nat. Commun.* 10:15.
- Geller S, Lieberman H, Belanger AJ, Yew NS, Kloss A, Ivanov AR. 2022. Comparison of microflow and analytical flow liquid chromatography coupled to mass spectrometry global metabolomics methods using a urea cycle disorder mouse model. *J. Proteome Res.* 21:151-163.
- Gnann C, Cesnik AJ, Lundberg E. 2021. Illuminating non-genetic cellular heterogeneity with imaging-based spatial proteomics. *Trends Cancer* 7:278-282.
- Gomes FP, Yates, III. 2019. Recent trends of capillary electrophoresis-mass spectrometry in proteomics research. *Mass Spectrom. Rev.* 38:445-460.
- Gregus M, Kostas JC, Ray S, Abbatiello SE, Ivanov AR. 2020. Improved sensitivity of ultralow flow LC-MS-based proteomic profiling of limited samples using monolithic capillary columns and FAIMS technology. *Anal. Chem.* 92:14702-14712.
- Greving MP, Patti GJ, Siuzdak G. 2011. Nanostructure-initiator mass spectrometry metabolite analysis and imaging. *Anal. Chem.* 83:2-7.
- Gupta K, Vuckovic I, Zhang S, Xiong YN, Carlson BL, Jacobs J, Olson I, Petterson XM, Macura SI, Sarkaria J, Burns TC. 2020. Radiation-induced metabolic alterations associate with tumor aggressiveness and poor outcome in glioblastoma. *Front. Oncol.* 10:19.
- HaileMariam M, Eguez RV, Singh H, Bekele S, Ameni G, Pieper R, Yu YB. 2018. S-trap, an ultrafast sample-preparation approach for shotgun proteomics. *J. Proteome Res.* 17:2917-2924.
- Hale OJ, Sisley EK, Griffiths RL, Styles IB, Cooper HJ. 2020. Native LESA TWIMS-MSI: spatial, conformational, and mass analysis of proteins and protein complexes. *J. Am. Soc. Mass Spectrom.* 31:873-879.
- Hamilton JS, Aguilar R, Petros RA, Verbeck GF. 2017. DAPNe with micro-capillary separatory chemistry-coupled to MALDI-MS for the analysis of polar and non-polar lipid metabolism in one cell. *J. Am. Soc. Mass Spectrom.* 28:918-928.
- Hanson JN, Motala MJ, Heien ML, Gillette M, Sweedler J, Nuzzo RG. 2009. Textural guidance cues for controlling process outgrowth of mammalian neurons. *Lab Chip* 9:122-131.
- Hebbar S, Schulz WD, Sauer U, Schwudke D. 2014. Laser capture microdissection coupled with on-column extraction LC-MSⁿ enables lipidomics of fluorescently labeled Drosophila neurons. *Anal. Chem.* 86:5345-5352.
- Hebert AS, Richards AL, Bailey DJ, Ulbrich A, Coughlin EE, Westphall MS, Coon JJ. 2014. The one hour yeast proteome. *Mol. Cell. Proteomics* 13:339-347.
- Helbok R, Kofler M, Schieffeler AJ, Gaasch M, Rass V, Pfausler B, Beer R, Schmutzhard E. 2017. Clinical use of cerebral microdialysis in patients with aneurysmal subarachnoid hemorrhage-state of the art. *Front. Neurol.* 8:25.
- Hirayama A, Abe H, Yamaguchi N, Tabata S, Tomita M, Soga T. 2018. Development of a sheathless CE-ESI-MS interface. *Electrophoresis* 39:1382-1389.
- Hobson BD, Choi SJ, Mosharov EV, Soni RK, Sulzer D, Sims PA. 2022. Subcellular proteomics of dopamine neurons in the mouse brain. *eLife* 11:e70921.
- Hook V, Lietz CB, Podvin S, Cajka T, Fiehn O. 2018. Diversity of neuropeptide cell-cell signaling molecules generated by proteolytic processing revealed by neuropeptidomics mass spectrometry. *J. Am. Soc. Mass Spectrom.* 29:807-816.
- Hu MZ, Helfenbein K, Buchberger AR, DeLaney K, Liu Y, Li LJ. 2021. Exploring the sexual dimorphism of crustacean neuropeptide expression using *Callinectes sapidus* as a model organism. *J. Proteome Res.* 20:2739-2750.
- Hu R, Li Y, Yang YH, Liu, ML. 2023. Mass spectrometry-based strategies for single-cell metabolomics. *Mass Spectrom. Rev.* 42: 67-94.
- Huang LS, Fang ML, Cupp-Sutton KA, Wang Z, Smith K, Wu S. 2021. Spray-capillary-based capillary electrophoresis mass spectrometry for metabolite analysis in single cells. *Anal. Chem.* 93:4479-4487.
- Hughes CS, Foehr S, Garfield DA, Furlong EE, Steinmetz LM, Krijgsvelde J. 2014. Ultrasensitive proteome analysis using paramagnetic bead technology. *Mol. Syst. Biol.* 10:14.
- Iannaccone JM, Ren SF, Hatcher NG, Sweedler JV. 2009. Collecting peptide Release from the brain using porous polymer monolith-based solid phase extraction capillaries. *Anal. Chem.* 81:5433-5438.
- Jirayupat C, Nagashima K, Hosomi T, Takahashi T, Tanaka W, Samransuksamer B, Zhang GZ, Liu JY, Kanai M, Yanagida T. 2021. Image processing and machine learning for automated

- identification of chemo-/biomarkers in chromatography mass Spectrometry. *Anal. Chem.* 93:14708-14715.
- Joensson HN, Svahn HA. 2012. Droplet microfluidics-a tool for single-cell analysis. *Angew. Chem. Int. Ed.* 51:12176-12192.
- Jones LM, Sperry JB, Carroll JA, Gross ML. 2011. Fast photochemical oxidation of proteins for epitope mapping. *Anal. Chem.* 83:7657-7661.
- Kawai T, Ota N, Okada K, Imasato A, Owa Y, Morita M, Tada M, Tanaka Y. 2019. Ultrasensitive single-cell metabolomics by capillary electrophoresis-mass spectrometry with a thin-walled tapered emitter and large-volume dual sample pre-concentration. *Anal. Chem.* 91:10564-10572.
- Kelly RT. 2020. Single-cell proteomics: progress and prospects. *Mol. Cell. Proteomics* 19:1739-1748.
- Kenderdine T, Xia ZJ, Williams ER, Fabris D. 2018. Submicrometer nanospray emitters provide new insights into the mechanism of cation adduction to anionic oligonucleotides. *Anal. Chem.* 90:13541-13548.
- Kim H, Dan K, Shin H, Lee J, Wang JI, Han D. 2019. An efficient method for high-pH peptide fractionation based on C18 StageTips for in-depth proteome profiling. *Anal. Meth.* 11:4693-4698.
- Kim Y, Edwards N, Fenselau C. 2016. Extracellular vesicle proteomes reflect developmental phases of *Bacillus subtilis*. *Clin. Proteomics* 13:8.
- Korte AR, Morris NJ, Vertes A. 2019. High throughput complementary analysis and quantitation of metabolites by maldi and silicon nanopost array-laser desorption/ionization-mass spectrometry. *Anal. Chem.* 91:3951-3958.
- Kurczy ME, Piehowski PD, Van Bell CT, Heien ML, Winograd N, Ewing AG. 2010. Mass spectrometry imaging of mating Tetrahymena show that changes in cell morphology regulate lipid domain formation. *Proc. Natl Acad. Sci. U.S.A.* 107:2751-2756.
- Kwok R, Tobe SS. 2006. Hemolymph clotting in crustaceans: implications for neuropeptide extraction from invertebrate hemolymph. *Peptides* 27:590-596.
- Lai X, Wang L, Witzmann FA. 2013. Issues and applications in label-free quantitative mass spectrometry. *Int. J. Proteomics* 2013:756039.
- Lam SS, Martell JD, Kamer KJ, Deerinck TJ, Ellisman MH, Mootha VK, Ting AY. 2015. Directed evolution of APEX2 for electron microscopy and proximity labeling. *Nat. Methods* 12: 51-54.
- Landgraf P, Antileo ER, Schuman EM, Dieterich DC. 2015. BONCAT: metabolic labeling, click chemistry, and affinity purification of newly synthesized proteomes. *Methods Mol. Biol.* 1266:199-215.
- Lanekoff I, Burnum-Johnson K, Thomas M, Cha J, Dey SK, Yang PX, Conaway MCP, Laskin J. 2015. Three-dimensional imaging of lipids and metabolites in tissues by nanospray desorption electrospray ionization mass spectrometry. *Anal. Bioanal. Chem.* 407:2063-2071.
- Lanni EJ, Dunham SJB, Nemes P, Rubakhin SS, Sweedler JV. 2014. Biomolecular imaging with a C60-SIMS/MALDI dual ion source hybrid mass spectrometer: instrumentation, matrix enhancement, and single cell analysis. *J. Am. Soc. Mass Spectrom.* 25:1897-1907.
- Lanni EJ, Rubakhin SS, Sweedler JV. 2012. Mass spectrometry imaging and profiling of single cells. *J. Proteomics* 75: 5036-5051.
- Lapainis T, Rubakhin SS, Sweedler JV. 2009. Capillary electrophoresis with electrospray ionization mass spectrometric detection for single-cell metabolomics. *Anal. Chem.* 81: 5858-5864.
- Lee WD, Mukha D, Aizenshtein E, Shlomi T. 2019. Spatial-fluxomics provides a subcellular-compartmentalized view of reductive glutamine metabolism in cancer cells. *Nat. Commun.* 10:14.
- Lee YH, Tan HT, Chung MCM. 2010. Subcellular fractionation methods and strategies for proteomics. *Proteomics* 10: 3935-3956.
- Li B, Tang J, Yang QX, Li S, Cui XJ, Li YH, Chen YZ, Xue WW, Li XF, Zhu F. 2017. NOREVA: normalization and evaluation of MS-based metabolomics data. *Nucleic Acids Res.* 45:W162-W170.
- Li H, Smith BK, Mark L, Nemes P, Nazarian J, Vertes A. 2015. Ambient molecular imaging by laser ablation electrospray ionization mass spectrometry with ion mobility separation. *Int. J. Mass Spectrom.* 377:681-689.
- Li XT, Zhao SL, Hu HK, Liu YM. 2016. A microchip electrophoresis-mass spectrometric platform with double cell lysis nano-electrodes for automated single cell analysis. *J. Chromatogr. A* 1451:156-163.
- Liang YR, Acor H, McCown MA, Nwosu AJ, Boekweg H, Axtell NB, Truong T, Cong YZ, Payne SH, Kelly RT. 2021. Fully automated sample processing and analysis workflow for low-input proteome profiling. *Anal. Chem.* 93:1658-1666.
- Liao HW, Rubakhin SS, Philip MC, Sweedler JV. 2020. Enhanced single-cell metabolomics by capillary electrophoresis electrospray ionization-mass spectrometry with field amplified sample injection. *Anal. Chim. Acta* 1118:36-43.
- Lindenburg PW, Haselberg R, Rozing G, Ramautar R. 2015. Developments in interfacing designs for CE-MS: towards enabling tools for proteomics and metabolomics. *Chromatographia* 78:367-377.
- Lipovsek M, Bardy C, Cadwell CR, Hadley K, Kobak D, Tripathy SJ. 2021. Patch-seq: past, present, and future. *J. Neurosci.* 41: 937-946.
- Liu JX, Aerts JT, Rubakhin SS, Zhang XX, Sweedler JV. 2014. Analysis of endogenous nucleotides by single cell capillary electrophoresis-mass spectrometry. *Analyst* 139:5835-5842.
- Liu RM, Pan N, Zhu YL, Yang ZB. 2018. T-Probe: an integrated microscale device for online in situ single cell analysis and metabolic profiling using mass spectrometry. *Anal. Chem.* 90: 11078-11085.
- Liu YF, Chen XY, Zhang YQ, Liu J. 2019. Advancing single-cell proteomics and metabolomics with microfluidic technologies. *Analyst* 144:846-858.
- Liu YS, Beyer A, Aebersold R. 2016. On the dependency of cellular protein levels on mRNA abundance. *Cell* 165:535-550.
- Liu ZC, Portero EP, Jian YR, Zhao YJ, Onjiko RM, Zeng C, Nemes P. 2019. Trace, machine learning of signal images for trace-sensitive mass spectrometry: a case study from single-cell metabolomics. *Anal. Chem.* 91:5768-5776.
- Loh KH, Stawski PS, Draycott AS, Udeshi ND, Lehrman EK, Wilton DK, Svinkina T, Deerinck TJ, Ellisman MH, Stevens B, Carr SA, Ting AY. 2016. Proteomic analysis of unbounded cellular compartments: synaptic clefts. *Cell* 166:1295-1307.
- Lombard-Banek C, Choi SB, Nemes P. 2019. Single-cell proteomics in complex tissues using microprobe capillary electrophoresis

- mass spectrometry. In: Allbritton NL, Kovarik ML, Editors. *Enzyme Activity in Single Cells*. London: Academic Press Ltd-Elsevier Science Ltd. p 263-292.
- Lombard-Banek C, Li J, Portero EP, Onjiko RM, Singer CD, Plotnick DO, Al Shabeeb RQ, Nemes P. 2021. In vivo subcellular mass spectrometry enables proteo-metabolomic single-cell systems biology in a chordate embryo developing to a normally behaving tadpole (*X. laevis*). *Angew. Chem. Int. Ed.* 60:12852-12858.
- Lombard-Banek C, Moody SA, Manzin MC, Nemes P. 2019. Microsampling capillary electrophoresis mass spectrometry enables single-cell proteomics in complex tissues: developing cell clones in live *Xenopus laevis* and zebrafish embryos. *Anal. Chem.* 91:4797-4805.
- Lombard-Banek C, Moody SA, Nemes P. 2016. Single-cell mass spectrometry for discovery proteomics: quantifying translational cell heterogeneity in the 16-cell frog (*Xenopus*) embryo. *Angew. Chem. Int. Ed.* 128:2500-2504.
- Lombard-Banek C, Reddy S, Moody SA, Nemes P. 2016. Label-free quantification of proteins in single embryonic cells with neural fate in the cleavage-stage frog (*Xenopus laevis*) embryo using capillary electrophoresis electrospray ionization high-resolution mass spectrometry (CE-ESI-HRMS). *Mol. Cell. Proteomics* 15:2756-2768.
- Lombard-Banek C, Yu Z, Swiercz AP, Marvar PJ, Nemes P. 2019. A microanalytical capillary electrophoresis mass spectrometry assay for quantifying angiotensin peptides in the brain. *Anal. Bioanal. Chem.* 411:4661-4671.
- Lopes A, Magrinelli E, Telley L. 2020. Emerging roles of single-cell multi-omics in studying developmental temporal patterning. *Int. J. Mol. Sci.* 21:18.
- Lubeckjy RA, Sun LL. 2022. Laser capture microdissection-capillary zone electrophoresis-tandem mass spectrometry (LCM-CZE-MS/MS) for spatially resolved top-down proteomics: a pilot study of zebrafish brain. *Mol. Omics* 18:112-122.
- Maes K, Van Liefferinge J, Viaene J, Van Schoors J, Van Wanseele Y, Bechade G, Chambers EE, Morren H, Michotte Y, Heyden YV, Claereboudt J, Smolders I, Van Eeckhauta A. 2014. Improved sensitivity of the nano ultra-high performance liquid chromatography-tandem mass spectrometric analysis of low-concentrated neuropeptides by reducing aspecific adsorption and optimizing the injection solvent. *J. Chromatogr. A* 1360:217-228.
- Marginean I, Kelly RT, Page JS, Tang KQ, Smith RD. 2007. Electrospray characteristic curves: in pursuit of improved performance in the nanoflow regime. *Anal. Chem.* 79:8030-8036.
- Massonnet P, Heeren RMA. 2019. A concise tutorial review of TOF-SIMS-based molecular and cellular imaging *J. Anal. At. Spectrom.* 34:2217-2228.
- Mast DH, Checco JW, Sweedler JV. 2020. Differential post-translational amino acid isomerization found among neuropeptides in aplysia California. *ACS Chem. Biol.* 15:272-281.
- Mast DH, Liao HW, Romanova EV, Sweedler JV. 2021. Analysis of peptide stereochemistry in single cells by capillary electrophoresis-trapped ion mobility spectrometry mass spectrometry. *Anal. Chem.* 93:6205-6213.
- Masuda T, Sugiyama N, Tomita M, Ohtsuki S, Ishihama Y. 2020. Mass spectrometry-compatible subcellular fractionation for proteomics. *J. Proteome Res.* 19:75-84.
- Matula K, Rivello F, Huck WTS. 2020. Single-cell analysis using droplet microfluidics. *Adv. Biosyst.* 4:28.
- McCool EN, Lubeckjy RA, Shen XJ, Chen DY, Kou Q, Liu XW, Sun LL. 2018. Deep top-down proteomics using capillary zone electrophoresis tandem mass spectrometry: identification of 5700 proteoforms from the *Escherichia coli* proteome. *Anal. Chem.* 90:5529-5533.
- McMillen JC, Gutierrez DB, Judd AM, Spraggins JM, Caprioli RM. 2021. Enhancement of tryptic peptide signals from tissue sections using MALDI IMS postionization (MALDI-2). *J. Am. Soc. Mass Spectrom.* 32:2583-2591.
- Meier F, Geyer PE, Winter SV, Cox J, Mann M. 2018. BoxCar acquisition method enables single-shot proteomics at a depth of 10,000 proteins in 100 minutes. *Nat. Meth.* 15:440-448.
- Mellors JS, Jorabchi K, Smith LM, Ramsey JM. 2010. Integrated microfluidic device for automated single-cell analysis using electrophoretic separation and electrospray ionization mass spectrometry. *Anal. Chem.* 82:967-973.
- Meng YF, Cheng XL, Wang TT, Hang W, Li XP, Nie W, Liu R, Lin Z, Hang L, Yin ZB, Zhang BL, Yan XM. 2020. Micro-lensed fiber laser desorption mass spectrometry imaging reveals subcellular distribution of drugs within single cells. *Angew. Chem Int. Ed.* 59:17864-17871.
- Mihailova A, Malerod H, Wilson SR, Karaszewski B, Hauser R, Lundanes E, Greibrokk T. 2008. Improving the resolution of neuropeptides in rat brain with on-line HILIC-RP compared to on-line SCX-RP. *J. Sep. Sci.* 31:459-467.
- Miltenyi S, Muller W, Weichel W, Radbruch A. 1990. High-gradient magnetic cell separation with MACS. *Cytometry* 11:231-238.
- Mizuno H, Tsuyama N, Harada T, Masujima T. 2008. Live single-cell video-mass spectrometry for cellular and subcellular molecular detection and cell classification. *J. Mass Spectrom.* 43:1692-1700.
- Moini M. 2007. Simplifying CE-MS operation. 2. Interfacing low-flow separation techniques to mass spectrometry using a porous tip. *Anal. Chem.* 79:4241-4246.
- Moroz LL, Kohn AB. 2013. Single-neuron transcriptome and methylome sequencing for epigenomic analysis of aging. In: Tollefsbol TO, Editor *Biological Aging: Methods and Protocols*, 2nd Edition. Totowa: Humana Press Inc. p 323-352.
- Mukherjee P, Berns EJ, Patino CA, Mouly EH, Chang LQ, Nathamgari SSP, Kessler JA, Mrksich M, Espinosa HD. 2020. Temporal sampling of enzymes from live cells by localized electroporation and quantification of activity by samdi mass spectrometry. *Small* 16:10.
- Nemes P, Barton AA, Vertes A. 2009. Three-dimensional imaging of metabolites in tissues under ambient conditions by laser ablation electrospray ionization mass spectrometry. *Anal. Chem.* 81:6668-6675.
- Nemes P, Goyal S, Vertes A. 2008. Conformational and noncovalent complexation changes in proteins during electrospray ionization. *Anal. Chem.* 80:387-395.
- Nemes P, Marginean I, Vertes A. 2007. Spraying mode effect on droplet formation and ion chemistry in electrosprays. *Anal. Chem.* 79:3105-3116.
- Nemes P, Rubakhin SS, Aerts JT, Sweedler JV. 2013. Qualitative and quantitative metabolomic investigation of single neurons by capillary electrophoresis electrospray ionization mass spectrometry. *Nat. Protoc.* 8:783-799.

- Nemes P, Woods AS, Vertes A. 2010. Simultaneous imaging of small metabolites and lipids in rat brain tissues at atmospheric pressure by laser ablation electrospray ionization mass spectrometry. *Anal. Chem.* 82:982-988.
- Neumann EK, Ellis JF, Triplett AE, Rubakhin SS, Sweedler JV. 2019. Lipid analysis of 30,000 individual rodent cerebellar cells using high-resolution mass spectrometry. *Anal. Chem.* 91: 7871-7878.
- Neupert S, Fusca D, Kloppenburg P, Predel R. 2018. Analysis of single neurons by perforated patch clamp recordings and MALDI-TOF mass spectrometry. *ACS Chem. Neurosci.* 9: 2089-2096.
- Nickerson JL, Baghalabadi V, Rajendran S, Jakubec PJ, Said H, McMillen TS, Dang ZH, Doucette AA. 2021. Recent advances in top-down proteome sample processing ahead of MS analysis. *Mass Spectrom. Rev.* 1-39. <https://doi.org/10.1002/mas.21706>
- Onjiko RM, Moody SA, Nemes P. 2015. Single-cell mass spectrometry reveals small molecules that affect cell fates in the 16-cell embryo. *Proc. Natl Acad. Sci. U.S.A.* 112: 6545-6550.
- Onjiko RM, Morris SE, Moody SA, Nemes P. 2016. Single-cell mass spectrometry with multi-solvent extraction identifies metabolic differences between left and right blastomeres in the 8-cell frog (*Xenopus*) embryo. *Analyst* 141:3648-3656.
- Onjiko RM, Portero EP, Moody SA, Nemes P. 2017. In situ microprobe single-cell capillary electrophoresis mass spectrometry: metabolic reorganization in single differentiating cells in the live vertebrate (*Xenopus laevis*) embryo. *Anal. Chem.* 89:7069-7076.
- Onorato JM, Xu C, Chen XQ, Rose AV, Generaux C, Lentz K, Shipkova P, Arthur S, Hennen JK, Haskell R, Myers MC, Lawrence RM, Finlay HJ, Basso M, Bostwick J, Fernando G, Garcia R, Hellings S, Hsu MY, Zhang RG, Zhao L, Gargalovic P. 2019. Linking (Pyr)(1)apelin-13 pharmacokinetics to efficacy: stabilization and measurement of a high clearance peptide in rodents. *Anal. Biochem.* 568:41-50.
- Osborn DM, Weiss DJ, Lunte CE. 2000. On-line preconcentration methods for capillary electrophoresis. *Electrophoresis* 21: 2768-2779.
- OuYang CZ, Liang ZD, Li LJ. 2015. Mass spectrometric analysis of spatio temporal dynamics of crustacean neuropeptides. *BBA-Proteins Proteom.* 1854:798-811.
- Ovchinnikova K, Kovalev V, Stuart L, Alexandrov T. 2020. OffsampleAI: artificial intelligence approach to recognize off-sample mass spectrometry images. *BMC Bioinform.* 21:11.
- Pace CL, Simmons J, Kelly RT, Muddiman DC. 2021. Multimodal mass spectrometry imaging of rat brain using IR-MALDESI and nanoPOTS-LC-MS/MS. *J. Proteome Res.* 21:713-720.
- Paine MRL, Liu JB, Huang DN, Ellis SR, Trede D, Kobarg JH, Heeren RMA, Fernandez FM, MacDonald TJ. 2019. Three-dimensional mass spectrometry imaging identifies lipid markers of medulloblastoma metastasis. *Sci. Rep.* 9:10.
- Pang Z, Chong J, Zhou G, de Lima Morais DA, Chang L, Barrette M, Gauthier C, Jacques PÉ, Li S, Xia J. 2021. MetaboAnalyst 5.0: narrowing the gap between raw spectra and functional insights. *Nucleic Acids Res.* 49(W1):W388-W396.
- Passarelli MK, Pirkel A, Moellers R, Grinfeld D, Kollmer F, Havelund R, Newman CF, Marshall PS, Arlinghaus H, Alexander MR, West A, Horning S, Niehuis E, Makarov A, Dollery CT, Gilmore IS. 2017. The 3D OrbiSIMS-label-free metabolic imaging with subcellular lateral resolution and high mass-resolving power. *Nat. Methods* 14:1175-1183.
- Perez JD, Dieck ST, Alvarez-Castelao B, Tushev G, Chan ICW, Schuman EM. 2021. Subcellular sequencing of single neurons reveals the dendritic transcriptome of GABAergic interneurons. *eLife* 10:26.
- Peshkin L, Wuhr M, Pearl E, Haas W, Freeman RM, Gerhart JC, Klein AM, Horb M, Gygi SP, Kirschner MW. 2015. On the relationship of protein and mRNA dynamics in vertebrate embryonic development. *Dev. Cell* 35:383-394.
- Peuchen EH, Zhu GJ, Sun LL, Dovichi NJ. 2017. Evaluation of a commercial electro-kinetically pumped sheath-flow nanospray interface coupled to an automated capillary zone electrophoresis system. *Anal. Bioanal. Chem.* 409:1789-1795.
- Pfammatter S, Bonnell E, McManus FP, Prasad S, Bailey DJ, Belford M, Dunyach JJ, Thibault P. 2018. A novel differential ion mobility device expands the depth of proteome coverage and the sensitivity of multiplex proteomic measurements. *Mol. Cell. Proteomics* 17:2051-2067.
- Phetsanthad A, Vu NQ, Yu Q, Buchberger AR, Chen ZW, Keller C, Li LJ. 2021. Recent advances in mass spectrometry analysis of neuropeptides. *Mass Spectrom. Rev.* 1-45. <https://doi.org/10.1002/mas.21734>
- Piehowski PD, Zhu Y, Bramer LM, Stratton KG, Zhao R, Orton DJ, Moore RJ, Yuan J, Mitchell HD, Gao Y, Webb-Robertson B-JM, Dey SK, Kelly RT, Burnum-Johnson KE. 2020. Automated mass spectrometry imaging of over 2000 proteins from tissue sections at 100- μ m spatial resolution. *Nat. Commun.* 11:8.
- Piga I, Capitoli G, Clerici F, Mahajneh A, Brambilla V, Smith A, Leni D, L'Imperio V, Galimberti S, Pagni F, Magni F. 2021. Ex vivo thyroid fine needle aspirations as an alternative for MALDI-MSI proteomic investigation: intra-patient comparison. *Anal. Bioanal. Chem.* 413:1259-1266.
- Planell N, Lagani V, Sebastian-Leon P, van der Kloet F, Ewing E, Karathanasis N, Urdangarin A, Arozarena I, Jagodic M, Tsamardinos I, Tarazona S, Conesa A, Tegner J, Gomez-Cabrero D. 2021. STATegra: multi-omics data integration—A conceptual scheme with a bioinformatics pipeline. *Front. Genet.* 12:12.
- Portero EP, Nemes P. 2019. Dual cationic-anionic profiling of metabolites in a single identified cell in a live *Xenopus laevis* embryo by microprobe CE-ESI-MS. *Analyst* 144:892-900.
- Potocnik NO, Fisher GL, Prop A, Heeren RMA. 2017. Sequencing and identification of endogenous neuropeptides with matrix-enhanced secondary ion mass spectrometry tandem mass spectrometry. *Anal. Chem.* 89:8223-8227.
- Qin SY, Zhang YX, Tian Y, Xu FG, Zhang P. 2022. Subcellular metabolomics: isolation, measurement, and applications. *J. Pharm. Biomed. Anal.* 210:14.
- Qu YY, Dubiak KM, Peuchen EH, Champion MM, Zhang ZB, Hebert AS, Wright S, Coon JJ, Huber PW, Dovichi NJ. 2020. Quantitative capillary zone electrophoresis-mass spectrometry reveals the N-glycome developmental plan during vertebrate embryogenesis. *Mol. Omics* 16:210-220.
- Race AM, Palmer AD, Dexter A, Steven RT, Styles IB, Bunch J. 2016. SpectralAnalysis: software for the masses. *Anal. Chem.* 88:9451-9458.

- Rappez L, Stadler M, Triana S, Gathungu RM, Ovchinnikova K, Phapale P, Heikenwalder M, Alexandrov T. 2021. SpaceM reveals metabolic states of single cells. *Nat. Meth.* 18:799-805.
- Reubsæet L, Sweredoski MJ, Moradian A, Lomenick B, Eggleston-Rangel R, Garbis SD. 2020. Nano volume fractionation strategy for dilute-and-shoot injections in off-line loss-less proteomic workflows for extensive protein identifications of ultra-low sample amounts. *J. Chromatogr. A* 1609:10.
- Reyzer ML, Caprioli RM. 2013. Imaging mass spectrometry of proteins and peptides. In: Desiderio DM, Nibbering NMM, Editors. *Mass Spectrometry for Drug Discovery and Drug Development*. John Wiley & Sons, Inc.
- Romanova EV, Rubakhin SS, Sweedler JV. 2008. One-step sampling, extraction, and storage protocol for peptidomics using dihydroxybenzoic acid. *Anal. Chem.* 80:3379-3386.
- Rosenberger G, Bludau I, Schmitt U, Heusel M, Hunter CL, Liu YS, MacCoss MJ, MacLean BX, Nesvizhskii AI, Pedrioli PGA, Reiter L, Rost HL, Tate S, Ting YS, Collins BC, Aebersold R. 2017. Statistical control of peptide and protein error rates in large-scale targeted data-independent acquisition analyses. *Nat. Methods* 14:921-927.
- Roux KJ, Kim DI, Raida M, Burke B. 2012. A promiscuous biotin ligase fusion protein identifies proximal and interacting proteins in mammalian cells. *J. Cell Biol.* 196:801-810.
- Rovira-Clave X, Jiang SZ, Bai YH, Zhu BK, Barlow G, Bhate S, Coskun AF, Han GJ, Ho CMK, Hitzman C, Chen SY, Bava FA, Nolan GP. 2021. Subcellular localization of biomolecules and drug distribution by high-definition ion beam imaging. *Nat. Commun.* 12:18.
- Saadatpour A, Lai SJ, Guo GJ, Yuan GC. 2015. Single-cell analysis in cancer genomics. *Trends Genet.* 31:576-586.
- Saha-Shah A, Esmaeili M, Sidoli S, Hwang H, Yang J, Klein PS, Garcia BA. 2019. Single cell proteomics by data-independent acquisition to study embryonic asymmetry in *Xenopus laevis*. *Anal. Chem.* 91:8891-8899.
- Saha-Shah A, Green CM, Abraham DH, Baker LA. 2016. Segmented flow sampling with push-pull theta pipettes. *Analyst* 141:1958-1965.
- Samarah LZ, Khatrar R, Tran TH, Stopka SA, Brantner CA, Parlanti P, Velickovic D, Shaw JB, Agtuca BJ, Stacey G, Pasa-Tolic L, Tolic N, Anderton CR, Vertes A. 2020. Single-cell metabolic profiling: metabolite formulas from isotopic fine structures in heterogeneous plant cell populations. *Anal. Chem.* 92:7289-7298.
- Sans M, Feider CL, Eberlin LS. 2018. Advances in mass spectrometry imaging coupled to ion mobility spectrometry for enhanced imaging of biological tissues. *Curr. Opin. Chem. Biol.* 42:138-146.
- Sansone SA, Fan T, Goodacre R, Griffin JL, Hardy NW, Kaddurah-Daouk R, Kristal BS, Lindon J, Mendes P, Morrison N, Nikolau B, Robertson D, Sumner LW, Taylor C, van der Werf M, van Ommen B, Fiehn O. 2007. The metabolomics standards initiative. *Nat. Biotechnol.* 25:844-848.
- Satpathy S, Jaehnig EJ, Krug K, Kim BJ, Saltzman AB, Chan DW, Holloway KR, Anurag M, Huang C, Singh P, Gao A, Namai N, Dou YC, Wen B, Vasaikar SV, Mutch D, Watson MA, Ma C, Ademuyiwa FO, Rimawi MF, Schiff R, Hoog J, Jacobs S, Malovannaya A, Hyslop T, Clauser KR, Mani DR, Perou CM, Miles G, Zhang B, Gillette MA, Carr SA, Ellis MJ. 2020. Microscaled proteogenomic methods for precision oncology. *Nat. Commun.* 11:16.
- Sauer CS, Li LJ. 2021. Mass spectrometric profiling of neuropeptides in response to copper toxicity via isobaric tagging. *Chem. Res. Toxicol.* 34:1329-1336.
- Schirmer EC, Yates JR, Gerace L. 2003. MudPIT: a powerful proteomics tool for discovery. *Discov. Med.* 3:38-39.
- Schmerberg CM, Li LJ. 2013. Mass spectrometric detection of neuropeptides using affinity-enhanced microdialysis with antibody-coated magnetic nanoparticles. *Anal. Chem.* 85:915-922.
- Schmerberg CM, Liang ZD, Li LJ. 2015. Data-independent MS/MS quantification of neuropeptides for determination of putative feeding-related neurohormones in microdialysate. *ACS Chem. Neurosci.* 6:174-180.
- Schoof EM, Furtwangler B, Uresin N, Rapin N, Savickas S, Gentil C, Lechman E, Keller UAD, Dick JE, Porse BT. 2021. Quantitative single-cell proteomics as a tool to characterize cellular hierarchies. *Nat. Commun.* 12:15.
- Sergent-Tanguy S, Chagneau C, Neveu I, Naveilhan P. 2003. Fluorescent activated cell sorting (FACS): a rapid and reliable method to estimate the number of neurons in a mixed population. *J. Neurosci. Meth.* 129:73-79.
- Seydel C. 2021. Single-cell metabolomics hits its stride. *Nat. Methods* 18:1452-1456.
- Shang LR, Cheng Y, Zhao YJ. 2017. Emerging droplet microfluidics. *Chem. Rev.* 117:7964-8040.
- Shao X, Wang XT, Guan S, Lin HZ, Yan GQ, Gao MX, Deng CH, Zhang XM. 2018. Integrated proteome analysis device for fast single-cell protein profiling. *Anal. Chem.* 90:14003-14010.
- Shen SC, Li J, Huo SA, Ma M, Zhu XY, Rasam S, Duan XT, Qu M, Titus MA, Qu J. 2021. Parallel, high-quality proteomic and targeted metabolomic quantification using laser capture microdissected tissues. *Anal. Chem.* 93:8711-8718.
- Simone NL, Bonner RF, Gillespie JW, Emmert-Buck MR, Liotta LA. 1998. Laser-capture microdissection: opening the microscopic frontier to molecular analysis. *Trends Genet.* 14:272-276.
- Slavov N. 2021. Single-cell protein analysis by mass spectrometry. *Curr. Opin. Chem. Biol.* 60:1-9.
- Slavov N. 2022. Scaling up single-cell proteomics. *Mol. Cell. Proteomics* 21:7.
- Smith LM, Agar JN, Chamot-Rooke J, Danis PO, Ge Y, Loo JA, Pasa-Tolic L, Tsybin YO, Kelleher NL, Consortium Top-Down P. 2021. The human proteoform project: defining the human proteome. *Sci. Adv.* 7:8.
- Specht H, Emmott E, Petelski AA, Huffman RG, Perlman DH, Serra M, Kharchenko P, Koller A, Slavov N. 2021. Single-cell proteomic and transcriptomic analysis of macrophage heterogeneity using SCoPE2. *Genome Biol.* 22:27.
- Sperber H, Mathieu J, Wang YL, Ferreccio A, Hesson J, Xu ZJ, Fischer KA, Devi A, Detraux D, Gu HW, Battle SL, Showalter M, Valensisi C, Bielas JH, Ericson NG,

- Margaretha L, Robitaille AM, Margineantu D, Fiehn O, Hockenbery D, Blau CA, Raftery D, Margolin AA, Hawldns RD, Moon RT, Ware CB, Ruohola-Baker H. 2015. The metabolome regulates the epigenetic landscape during naive-to-primed human embryonic stem cell transition. *Nat. Cell Biol.* 17:1523-1535.
- Stopka SA, Samarah LZ, Shaw JB, Liyu AV, Velickovic D, Agtuca BJ, Kukolj C, Koppelaar DW, Stacey G, Pasa-Tolic L, Anderton CR, Vertes A. 2019. Ambient metabolic profiling and imaging of biological samples with ultrahigh molecular resolution using laser ablation electrospray ionization 21 Tesla FTICR mass spectrometry. *Anal. Chem.* 91:5028-5035.
- Suvannapruk W, Edney MK, Kim D-H, Scurr DJ, Ghaemmaghami AM, Alexander MR. 2022. Single-cell metabolic profiling of macrophages using 3D OrbiSIMS: correlations with phenotype. *Anal. Chem.* 94: 9389-9398.
- Szymanska E, Davies AN, Buydens LMC. 2016. Chemometrics for ion mobility spectrometry data: recent advances and future prospects. *Analyst* 141:5689-5708.
- Taylor CF, Binz PA, Aebersold R, Affolter M, Barkovich R, Deutsch EW, Horn DM, Huhmer A, Kussmann M, Lilley K, Macht M, Mann M, Mueller D, Neubert TA, Nickson J, Patterson SD, Raso R, Resing K, Seymour SL, Tsugita A, Xenarios I, Zeng R, Julian RK. 2008. Guidelines for reporting the use of mass spectrometry in proteomics. *Nat. Biotechnol.* 26:860-861.
- Taylor MJ, Lukowski JK, Anderton CR. 2021. Spatially resolved mass spectrometry at the single cell: recent innovations in proteomics and metabolomics. *J. Am. Soc. Mass Spectrom.* 32: 872-894.
- Taylor MJ, Mattson S, Liyu A, Stopka SA, Ibrahim YM, Vertes A, Anderton CR. 2021. Optical microscopy-guided laser ablation electrospray ionization ion mobility mass spectrometry: ambient single cell metabolomics with increased confidence in molecular identification. *Metabolites* 11:15.
- Tharkeshwar AK, Gevaert K, Annaert W. 2018. Organellar omics—A reviving strategy to untangle the biomolecular complexity of the cell. *Proteomics* 18:12.
- Tharkeshwar AK, Trekker J, Vermeire W, Pauwels J, Sannerud R, Priestman DA, te Vrugte D, Vints K, Baatsen P, Decuypere JP, Lu HQ, Martin S, Vangheluwe P, Swinnen JV, Lagae L, Impens F, Platt FM, Gevaert K, Annaert W. 2017. A novel approach to analyze lysosomal dysfunctions through subcellular proteomics and lipidomics: the case of NPC1 deficiency. *Sci. Rep.* 7:20.
- Tian Q, Stepaniants SB, Mao M, Weng L, Feetham MC, Doyle MJ, Yi EC, Dai HY, Thorsson V, Eng J, Goodlett D, Berger JP, Gunter B, Linseley PS, Stoughton RB, Aebersold R, Collins SJ, Hanlon WA, Hood LE. 2004. Integrated genomic and proteomic analyses of gene expression in mammalian cells. *Mol. Cell. Proteomics* 3:960-969.
- Tiway S, Levy R, Gutenbrunner P, Soto FS, Palaniappan KK, Deming L, Berndt M, Brant A, Cimermanic P, Cox J. 2019. High-quality MS/MS spectrum prediction for data-dependent and data-independent acquisition data analysis. *Nat. Meth.* 16: 519-525.
- Tsuyama N, Mizuno H, Tokunaga E, Masujima T. 2008. Live single-cell molecular analysis by video-mass spectrometry. *Anal. Sci.* 24:559-561.
- Turiak L, Toth G, Ozohanics O, Revesz A, Acs A, Vekey K, Zaia J, Draho L. 2018. Sensitive method for glycosaminoglycan analysis of tissue sections. *J. Chromatogr. A* 1544: 41-48.
- Tyanova S, Temu T, Sinitcyn P, Carlson A, Hein MY, Geiger T, Mann M, Cox J. 2016. The Perseus computational platform for comprehensive analysis of (prote)omics data. *Nat. Methods* 13: 731-740.
- Velasquez E, Martins-de-Souza D, Velasquez I, Carneiro GRA, Schmitt A, Falkai P, Domont GB, Nogueira FCS. 2019. Quantitative subcellular proteomics of the orbitofrontal cortex of schizophrenia patients. *J. Proteome Res.* 18: 4240-4253.
- Vermeire PJ, Van Schepdael A, Petersen NJ. 2020. Development of a novel sheathless CE-ESI-MS interface via a CO₂ laser ablated opening. *Talanta* 214:8.
- Vo TD, Palsson BO. 2007. Building the power house: recent advances in mitochondrial studies through proteomics and systems biology. *Am. J. Physiol.-Cell Physiol.* 292: C164-C177.
- Vocat C, Dunand M, Hubers SA, Bourdillon N, Millet GP, Brown NJ, Wuerzner G, Grouzmann E, Eugster PJ. 2020. Quantification of neuropeptide Y and four of its metabolites in human plasma by micro-UHPLC-MS/MS. *Anal. Chem.* 92:859-866.
- Vogel C, Marcotte EM. 2012. Insights into the regulation of protein abundance from proteomic and transcriptomic analyses. *Nat. Rev. Genet.* 13:227-232.
- von Eggeling F, Hoffmann F. 2020. Microdissection—An essential prerequisite for spatial cancer omics. *Proteomics* 20:10.
- Waas M, Kislinger T. 2020. Addressing cellular heterogeneity in cancer through precision proteomics. *J. Proteome Res.* 19: 3607-3619.
- Wang JH, Jiang XY, Sturm RM, Li LJ. 2009. Combining tissue extraction and off-line capillary electrophoresis matrix-assisted laser desorption/ionization Fourier transform mass spectrometry for neuropeptide analysis in individual neuronal organs using 2,5-dihydroxybenzoic acid as a multi-functional agent. *J. Chromatogr. A* 1216: 8283-8288.
- Wang S, Xiao CS, Li Y, Ling L, Chen XS, Guo XH. 2017. A surface pattern on MALDI steel plate for one-step in-situ self-desalting and enrichment of peptides/proteins. *J. Am. Soc. Mass Spectrom.* 28:428-433.
- Whiteman SA, Nuhse TS, Ashford DA, Sanders D, Maathuis FJM. 2008. A proteomic and phosphoproteomic analysis of *Oryza sativa* plasma membrane and vacuolar membrane. *Plant J.* 56: 146-156.
- Wijetunge CD, Saeed I, Boughton BA, Spraggins JM, Caprioli RM, Bacic A, Roessner U, Halgamuge SK. 2015. EXIMS: an improved data analysis pipeline based on a new peak-picking method for exploring imaging mass spectrometry data. *Bioinformatics* 31:3198-3206.
- Williams MD, Zhang X, Park JJ, Siems WF, Gang DR, Resar LMS, Reeves R, Hill HH. 2015. Characterizing

- metabolic changes in human colorectal cancer. *Anal. Bioanal. Chem.* 407:4581-4595.
- Williams SM, Liyu AV, Tsai CF, Moore RJ, Orton DJ, Chrisler WB, Gaffrey MJ, Liu T, Smith RD, Kelly RT, Pasa-Tolic L, Zhu Y. 2020. Automated coupling of nanodroplet sample preparation with liquid chromatography-mass spectrometry for high throughput single-cell proteomics. *Anal. Chem.* 92: 10588-10596.
- Wilson RE, Jaquins-Gerstl A, Weber SG. 2018. On-column dimethylation with capillary liquid chromatography tandem mass spectrometry for online determination of neuropeptides in rat brain microdialysate. *Anal. Chem.* 90:4561-4568.
- Wiseman JM, Ifa DR, Song QY, Cooks RG. 2006. Tissue imaging at atmospheric pressure using desorption electrospray ionization (DESI) mass spectrometry. *Angew. Chem. Int. Ed.* 45: 7188-7192.
- Wisniewski JR, Zougman A, Nagaraj N, Mann M. 2009. Universal sample preparation method for proteome analysis. *Nat. Meth.* 6:359-362.
- Woo HK, Northen TR, Yanes O, Siuzdak G. 2008. Nanostructure-initiator mass spectrometry: a protocol for preparing and applying NIMS surfaces for high-sensitivity mass analysis. *Nat. Protoc.* 3:1341-1349.
- Woo J, Clair GC, Williams SM, Feng S, Tsai C-F, Moore RJ, Chrisler WB, Smith RD, Kelly RT, Paša-Tolić L, Ansong C, Zhu Y. 2022. Three-dimensional feature matching improves coverage for single-cell proteomics based on ion mobility filtering. *Cell Syst.* 13:426-434.
- Woo J, Williams SM, Markillie LM, Feng S, Tsai CF, Aguilera-Vazquez V, Sontag RL, Moore RJ, Hu DH, Mehta HS, Cantlon-Bruce J, Liu T, Adkins JN, Smith RD, Clair GC, Pasa-Tolic L, Zhu Y. 2021. High throughput and high-efficiency sample preparation for single-cell proteomics using a nested nanowell chip. *Nat. Commun.* 12:1.
- Wu HK, Wheeler A, Zare RN. 2004. Chemical cytometry on a picoliter-scale integrated microfluidic chip. *Proc. Natl. Acad. Sci. U.S.A.* 101:12809-12813.
- Wu L, Liu F, Cai H. 2021. IOAT: an interactive tool for statistical analysis of omics data and clinical data. *BMC Bioinform.* 22:326.
- Xiang PL, Zhu Y, Yang Y, Zhao ZT, Williams SM, Moore RJ, Kelly RT, Smith RD, Liu SR. 2020. Picoflow liquid chromatography-mass spectrometry for ultrasensitive bottom-up proteomics using 2- μ m-i.d. open tubular columns. *Anal. Chem.* 92:4711-4715.
- Xu X, Liu K, Fan ZH. 2012. Microscale 2D separation systems for proteomic analysis. *Expert Rev. Proteomics* 9:135-147.
- Yanes O, Tautenhahn R, Patti GJ, Siuzdak G. 2011. Expanding coverage of the metabolome for global metabolite profiling. *Anal. Chem.* 83:2152-2161.
- Yang LW, George J, Wang J. 2020. Deep profiling of cellular heterogeneity by emerging single-cell proteomic technologies. *Proteomics* 20:12.
- Yang YY, Huang YY, Wu JH, Liu N, Deng JW, Luan TG. 2017. Single-cell analysis by ambient mass spectrometry. *Trends Anal. Chem.* 90:14-26.
- Yang ZC, Shen XJ, Chen DY, Sun LL. 2019. Improved nanoflow RPLC-CZE-MS/MS system with high peak capacity and sensitivity for nanogram bottom-up proteomics. *J. Proteome Res.* 18:4046-4054.
- Yang ZC, Sun LL. 2021. Recent technical progress in sample preparation and liquid-phase separation-mass spectrometry for proteomic analysis of mass-limited samples. *Anal. Meth.* 13:1214-1225.
- Yi L, Piehowski PD, Shi TJ, Smith RD, Qian WJ. 2017. Advances in microscale separations towards nanoproteomics applications. *J. Chromatogr. A* 1523:40-48.
- Yoon S, Park S, Kim MS, Lee CY. 2018. Concomitant desalting and concentration of neuropeptides on a donut-shaped surface pattern for MALDI mass spectrometry. *Chem. Commun.* 54: 5688-5691.
- Yu DH, Wang Z, Cupp-Sutton KA, Guo YT, Kou Q, Smith K, Liu XW, Wu S. 2021. Quantitative top-down proteomics in complex samples using protein-level tandem mass tag labeling. *J. Am. Soc. Mass Spectrom.* 32:1336-1344.
- Yuan ZY, Zhou QM, Cai LS, Pan L, Sun WL, Qumu SW, Yu S, Feng JX, Zhao HS, Zheng YC, Shi ML, Li S, Chen Y, Zhang XR, Zhang MQ. 2021. SEAM is a spatial single nuclear metabolomics method for dissecting tissue microenvironment. *Nat. Methods* 18:1223-1232.
- Zemenova J, Sykora D, Freislebenova A, Maletinska L. 2017. LC-MS/MS analysis of lipidized analogs of prolactin-releasing peptide utilizing a monolithic column and simple sample preparation. *Bioanalysis* 9:1319-1328.
- Zestos AG, Kennedy RT. 2017. Microdialysis coupled with LC-MS/MS for in vivo neurochemical monitoring. *Aaps J.* 19: 1284-1293.
- Zhang FF, Ge WG, Ruan G, Cai X, Guo TN. 2020. Data-independent acquisition mass spectrometry-based proteomics and software tools: a glimpse in 2020. *Proteomics* 20:12.
- Zhang JL, Rector J, Lin JQ, Young JH, Sans M, Katta N, Giese N, Yu WD, Nagi C, Suliburk J, Liu JS, Bensussan A, DeHoog RJ, Garza KY, Ludolph B, Sorace AG, Syed A, Zahedivash A, Milner TE, Eberlin LS. 2017. Nondestructive tissue analysis for ex vivo and in vivo cancer diagnosis using a handheld mass spectrometry system. *Sci. Transl. Med.* 9:11.
- Zhang L, Vertes A. 2018. Single-cell mass spectrometry approaches to explore cellular heterogeneity. *Angew. Chem. Int. Ed.* 57: 4466-4477.
- Zhang LW, Khattar N, Kemenes I, Kemenes G, Zrinyi Z, Pirger Z, Vertes A. 2018. Subcellular peptide localization in single identified neurons by capillary microsampling mass spectrometry. *Sci. Rep.* 8:10.
- Zhang XC, Wei ZW, Gong XY, Si XY, Zhao YY, Yang CD, Zhang SC, Zhang XR. 2016. Integrated droplet-based micro-extraction with ESI-MS for removal of matrix interference in single-cell analysis. *Sci. Rep.* 6:9.
- Zhang YY, Zheng TT, Wang L, Feng L, Wang M, Zhang ZC, Feng HH. 2021. From passive to active sorting in microfluidics: a review. *Rev. Adv. Mater. Sci.* 60:313-324.
- Zhang ZB, Dubiak KM, Huber PW, Dovichi NJ. 2020. Miniaturized filter-aided sample preparation (MICRO-FASP) method for high throughput, ultrasensitive proteomics sample preparation reveals proteome asymmetry in *Xenopus laevis* embryos. *Anal. Chem.* 92:5554-5560.

- Zhao SS, Zhong XF, Tie C, Chen DDY. 2012. Capillary electrophoresis-mass spectrometry for analysis of complex samples. *Proteomics* 12:2991-3012.
- Zhou MW, Uwugiaren N, Williams SM, Moore RJ, Zhao R, Goodlett D, Dapic I, Pasa-Tolic L, Zhu Y. 2020. Sensitive top-down proteomics analysis of a low number of mammalian cells using a nanodroplet sample processing platform. *Anal. Chem.* 92:7087-7095.
- Zhou XH, Xiao W, Su ZY, Cheng JH, Zheng CY, Zhang ZJ, Wang YQ, Wang L, Xu BH, Li SP, Yang XF, Hoi MPM. 2019. Hippocampal proteomic alteration in triple transgenic mouse model of Alzheimer's disease and implication of PINK 1 regulation in donepezil treatment. *J. Proteome Res.* 18: 1542-1552.
- Zhu GZ, Shao YL, Liu YX, Pei T, Li LJ, Zhang DT, Guo GS, Wang XY. 2021. Single-cell metabolite analysis by electrospray ionization mass spectrometry. *Trends Anal. Chem.* 143:18.
- Zhu HY, Li QQ, Liao TP, Yin X, Chen Q, Wang ZY, Dai MF, Yi L, Ge SY, Miao CJ, Zeng WP, Qu LL, Ju ZY, Huang GM, Cang CL, Xiong W. 2021. Metabolomic profiling of single enlarged lysosomes. *Nat. Meth.* 18:788-798.
- Zhu XP, Xu TY, Peng C, Wu SH. 2022. Advances in MALDI mass spectrometry imaging single cell and tissues. *Front. Chem.* 9:27.
- Zhu Y, Clair G, Chrisler WB, Shen Y, Zhao R, Shukla AK, Moore RJ, Misra RS, Pryhuber GS, Smith RD, Ansong C, Kelly RT. 2018. Proteomic analysis of single mammalian cells enabled by microfluidic nanodroplet sample preparation and ultrasensitive nanoLC-MS. *Angew. Chem. Int. Ed.* 57: 12370-12374.
- Zhu Y, Dou MW, Piehowski PD, Liang YR, Wang FJ, Chu RK, Chrisler WB, Smith JN, Schwarz KC, Shen YF, Shukla AK, Moore RJ, Smith RD, Qian WJ, Kelly RT. 2018. Spatially resolved proteome mapping of laser capture microdissected tissue with automated sample transfer to nanodroplets. *Mol. Cell. Proteomics* 17:1864-1874.
- Zhu Y, Piehowski PD, Zhao R, Chen J, Shen YF, Moore RJ, Shukla AK, Petyuk VA, Campbell-Thompson M, Mathews CE, Smith RD, Qian WJ, Kelly RT. 2018. Nanodroplet processing platform for deep and quantitative proteome profiling of 10-100 mammalian cells. *Nat. Commun.* 9:10.
- Zhu Y, Scheibinger M, Ellwanger DC, Krey JF, Choi D, Kelly RT, Heller S, Barr-Gillespie PG. 2019. Single-cell proteomics reveals changes in expression during hair-cell development. *eLife* 8:26.
- Zimmerman TA, Rubakhin SS, Sweedler JV. 2011. MALDI mass spectrometry imaging of neuronal cell cultures. *J. Am. Soc. Mass Spectrom.* 22:828-836.
- Zivraj KH, Tung YCL, Piper M, Gumy L, Fawcett JW, Yeo GSH, Holt CE. 2010. Subcellular profiling reveals distinct and developmentally regulated repertoire of growth cone mRNAs. *J. Neurosci.* 30:15464-15478.
- Zougman A, Pilch B, Podtelejnikov A, Kiehnopf M, Schnabel C, Kurnar C, Mann M. 2008. Integrated analysis of the cerebrospinal fluid peptidome and proteome. *J. Proteome Res.* 7:386-399.
- Zougman A, Wilson JP, Roberts LD, Banks RE. 2020. Detergent-free simultaneous sample preparation method for proteomics and metabolomics. *J. Proteome Res.* 19:2838-2844.

AUTHOR BIOGRAPHIES



Leena R. Pade is a PhD candidate in Chemistry under the mentorship of Dr. Peter Nemes at the University of Maryland, College Park. She is a fellow under the National Cancer Institute–University of Maryland Partnership for Integrative Cancer Research. She received her master's degree in Pharmaceutical Analysis at the National Institute of Pharmaceutical Education and Research, Hyderabad (India) in 2018. Her research focuses on developing microanalytical proteomic approaches to elucidate molecular mechanisms underlying the differentiation of embryonic cells and stem-like cell populations to neural and craniofacial tissues during vertebrate embryonic development.



Kaitlyn E. Stepler earned her PhD in Chemistry from Vanderbilt University in 2021 under the mentorship of Dr. Renā A. S. Robinson. Her dissertation research used mass spectrometry-based proteomics to understand racial disparities in Alzheimer's disease. Dr. Stepler is currently a postdoctoral associate at the University of Maryland, College Park in Dr. Peter Nemes' laboratory, where her research focuses on understanding various aspects of embryonic development using proteomics.



Erika P. Portero obtained her PhD in Chemistry from the University of Maryland, College Park under the mentorship of Dr. Peter Nemes. She developed single-cell mass spectrometry tools and methodologies to investigate metabolic reorganization during vertebrate embryogenesis. Dr. Portero is presently a postdoctoral fellow at the Dana-Farber Cancer Institute and the Harvard Medical School in Dr. Nika Danial's laboratory, where she investigates the metabolic characteristics and associated vulnerabilities in pediatric gliomas.



Kellen DeLaney obtained her PhD in Chemistry from the University of Wisconsin-Madison, WI under the mentorship of Dr. Lingjun Li, where she developed mass spectrometry techniques to study neuropeptides in crustacean model organisms. As a postdoctoral associate, she focused on using capillary electrophoresis with ion mobility mass spectrometry to study endogenous

peptides in the mammalian brain in Dr. Peter Nemes' laboratory at the University of Maryland, College Park.



Peter Nemes holds a PhD from the George Washington University (advisor: Dr. Akos Vertes), where he (co)-invented laser ablation electrospray ionization for ambient mass spectrometry and 2- and 3-dimensional molecular imaging. As a postdoctoral associate with Dr. Jonathan V. Sweedler at the University of Illinois, Urbana–Champaign, IL, he developed mass spectrometry instruments and methods for analytical neuroscience. Research in the Nemes Laboratory, currently at the University of Maryland, College Park, builds ultrasensitive mass spectrometry technologies to study proteo-metabolomic mechanisms

governing single-cell processes in biological models of (neuro)developmental biology and neuroscience. Results from the group led to the finding of previously unknown molecular differences between identified cells and neurons as well as the discovery of noncanonical effectors of cell fate decisions and mechanism controlling the patterning of the vertebrate body plan and organismal behavior.

How to cite this article: Pade LR, Stepler KE, Portero EP, DeLaney K, Nemes P. Biological mass spectrometry enables spatiotemporal 'omics: from tissues to cells to organelles. *Mass Spectrometry Reviews*, (2022);e21824.

<https://doi.org/10.1002/mas.21824>